

Datasheet for ABIN6229569 **SBEM CLIA Kit**



Overview

| Overview | |
|---|--|
| Quantity: | 96 tests |
| Target: | SBEM |
| Reactivity: | Human |
| Method Type: | Sandwich ELISA |
| Detection Range: | 0.312 ng/mL - 20 ng/mL |
| Minimum Detection Limit: | 0.312 ng/mL |
| Application: | ELISA |
| Product Details | |
| Purpose: | Human Small Breast Epithelial Mucin (SBEM) CLIA Kit is a Sandwich CLIA Kit for use with Serum, plasma, tissue homogenates, saliva and other biological fluids. |
| | ,, |
| Sample Type: | Plasma, Saliva, Serum, Tissue Homogenate |
| Sample Type: Analytical Method: | |
| | Plasma, Saliva, Serum, Tissue Homogenate |
| Analytical Method: | Plasma, Saliva, Serum, Tissue Homogenate Quantitative |
| Analytical Method: Detection Method: | Plasma, Saliva, Serum, Tissue Homogenate Quantitative Chemiluminescent |
| Analytical Method: Detection Method: Sensitivity: | Plasma, Saliva, Serum, Tissue Homogenate Quantitative Chemiluminescent |

Application Details

| Application Notes: | Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less |
|--------------------|---|
| | than 5 % within the expiration date under appropriate storage conditions. To minimize |
| | performance fluctuations, operation procedures and lab conditions should be strictly controlled. |
| | It is also strongly suggested that the whole assay is performed by the same user throughout. |
| | Recommended dilutions: Optimal dilutions/concentrations should be determined by the end |
| | user. |
| | Standard Form: Lyophilized |
| Plate: | Pre-coated |
| Restrictions: | For Research Use only |
| | |
| Handling | |
| Storage: | 4 °C/-20 °C |
| Storage Comment: | Upon receipt, store the kit according to the storage instruction in the kit's manual. |
| Expiry Date: | 6 months |