

Datasheet for ABIN6234120 **C11orf82 CLIA Kit**



[Go to Product page](#)

Overview

Quantity:	96 tests
Target:	C11orf82
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.312 ng/mL - 20 ng/mL
Minimum Detection Limit:	0.312 ng/mL
Application:	ELISA

Product Details

Purpose:	Human Nitric Oxide Inducible Gene Protein (NOXIN/DDIAS) CLIA Kit is a Sandwich CLIA Kit for use with Tissue homogenates, cell lysates and other biological fluids.
Sample Type:	Cell Lysate, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Chemiluminescent
Sensitivity:	< 0.114 ng/mL

Target Details

Target:	C11orf82
Alternative Name:	Nitric Oxide Inducible Gene Protein (C11orf82 Products)

Application Details

Application Notes: Stability: The stability of the kit is determined by the rate of activity loss. The loss rate is less than 5 % within the expiration date under appropriate storage conditions. To minimize performance fluctuations, operation procedures and lab conditions should be strictly controlled. It is also strongly suggested that the whole assay is performed by the same user throughout. Recommended dilutions: Optimal dilutions/concentrations should be determined by the end user.

Standard Form: Lyophilized

Plate: Pre-coated

Restrictions: For Research Use only

Handling

Storage: 4 °C/-20 °C

Storage Comment: Upon receipt, store the kit according to the storage instruction in the kit's manual.

Expiry Date: 6 months
