antibodies -online.com







Adrenomedullin ELISA Kit



Image



\sim				
	11/6	٦r١	/10	۱۸.

Quantity:	96 tests	
Target:	Adrenomedullin (ADM)	
Reactivity:	Mouse	
Method Type:	Competition ELISA	
Detection Range:	0.1 ng/mL - 1000 ng/mL	
Minimum Detection Limit:	0.1 ng/mL	
Application:	ELISA	
Product Details		
Purpose:	Human/Mouse/Rat Adrenomedullin EIA Kit optimized for serum, plasma and cell culture media. Competition-based ELISA on a 96-well strip plate.	
Sample Type:	Cell Culture Supernatant, Serum	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This EIA kit is designed to detect human, mouse, and rat Adrenomedullin	
Sensitivity:	0.9 ng/mL	
Characteristics:	 Strip plates and additional reagents allow for use in multiple experiments Quantitative protein detection Establishes normal range The best products for confirmation of antibody array data 	

Product Details

Components:

- · Pre-Coated 96-well Strip Microplate
- · Wash Buffer
- · Standard Peptide
- · Assay Diluent(s)
- · Biotinylated Peptide
- · HRP-Streptavidin
- · TMB One-Step Substrate
- · Stop Solution
- · Assay Diagram
- · Positive Control Sample
- · Capture Antibody
- · User Manual

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µL to 1 mL volumes
- Adjustable 1-25 mL pipettes for reagent preparation
- 100 mL and 1 liter graduated cylinders
- · Tubes to prepare standard and sample dilutions
- · Orbital shaker
- · Aluminum foil
- · Saran Wrap
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- SigmaPlot software (or other software that can perform four-parameter logistic regression models)


Target Details

Target:	Adrenomedullin (ADM)	
Alternative Name:	Adrenomedullin (ADM Products)	
Target Type:	Hormone	
Gene ID:	11535	
UniProt:	P97297	
Pathways:	Hormone Transport, Hormone Activity, C21-Steroid Hormone Metabolic Process, cAMP Metabolic Process, Myometrial Relaxation and Contraction, Regulation of G-Protein Coupled	
	Receptor Protein Signaling, Tube Formation	

Application Details

Application Notes:	Optimal working dilution should be determined by the investigator.	
Plate:	Pre-coated	
Protocol:	1. Prepare all reagents, samples and standards as instructed.	
	2. Add 100 µL detection antibody to each well.	
	3. Incubate 1.5 h at RT or O/N at 4 °C.	
	4. Add 100 μL standard or sample to each well.	
	5. Incubate 2.5 h at RT.	
	6. Add 100 μL prepared streptavidin solution.	
	7. Incubate 45 min at RT.	
	8. Add 100 μL TMB One-Step Substrate Reagent to each well.	
	9. Incubate 30 min at RT.	
	10. Add 50 μL Stop Solution to each well.	
	11. Read plate at 450 nm immediately.	
Reagent Preparation:	Recommended Dilution for serum and plasma samplesHuman: 2x / Mouse: 2x / Rat: 2x	
Restrictions:	For Research Use only	
Handling		
Storage:	-20 °C	
Storage Comment:	Standard, biotinylated peptide, and positive control should be stored at -20°C after arrival. Avoid	
	multiple freeze-thaws. The remaining kit components may be stored at 4°C. Opened microplate	
	wells and antibody (Item N) may be stored for up to 1 month at 2° to 8°C. Return unused wells	
	to the pouch containing desiccant pack and reseal along entire edge.	
Expiry Date:	6 months	

Adrenomedullin EIA



ELISA

Image 1. Standard Curve