.-online.com antibodies

Datasheet for ABIN624951 BMP4 ELISA Kit

1 Image

2 Publications



Overview

Quantity:	96 tests
Target:	BMP4
Reactivity:	Human
Method Type:	Sandwich ELISA
Application:	ELISA

Product Details

Purpose:	Human BMP-4 ELISA Kit for cell and tissue lysate samples.
Sample Type:	Cell Lysate, Tissue Lysate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	Cross Reactivity: This ELISA kit shows no cross-reactivity with any of the cytokines tested:
	Human Angiogenin, BDNF, BLC, BMP-6, BMP-7, ENA-78, FGF-4, IL-1 alpha, IL-1 beta, IL-2, IL-3,
	IL-4, IL-5, IL-7, IL-8, IL-9, IL-10, IL-11, IL-12 p70, IL-12 p40, IL-13, IL-15, I-309, IP-10, G-CSF, GM-
	CSF, IFN-gamma, Leptin, MCP-1, MCP-2, MCP-3, MDC, MIP-1 alpha, MIP-1 beta, MIP-1 delta,
	PARC, PDGF, RANTES, SCF, TARC, TGF-beta, TIMP-1, TIMP-2, TNF-alpha, TNF-beta, TPO, VEGF.
Sensitivity:	15 pg/mL
Characteristics:	Strip plates and additional reagents allow for use in multiple experiments
	Quantitative protein detection
	Establishes normal range
	 The best products for confirmation of antibody array data

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/6 | Product datasheet for ABIN624951 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Product Details

Components:	Pre-Coated 96-well Strip MicroplateWash Buffer
	Stop Solution
	 Assay Diluent(s)
	Lyophilized Standard
	Biotinylated Detection Antibody
	Streptavidin-Conjugated HRP
	TMB One-Step Substrate
Material not included:	Distilled or deionized water
	- Precision pipettes to deliver 2 μ L to 1 μ L volumes
	 Adjustable 1-25 µL pipettes for reagent preparation
	 100 μL and 1 liter graduated cylinders
	Tubes to prepare standard and sample dilutions
	Absorbent paper
	Microplate reader capable of measuring absorbance at 450nm
	Log-log graph paper or computer and software for ELISA data analysis
	Cell lysate buffer

Target Details

Target:	BMP4
Alternative Name:	BMP4 (BMP4 Products)
Background:	The Human BMP-4 ELISA (Enzyme-Linked Immunosorbent Assay) kit is an in vitro enzyme- linked immunosorbent assay for the quantitative measurement of human BMP-4 cell lysate and tissue lysate. This assay employs an antibody specific for human BMP-4 coated on a 96-well plate. Standards and samples are pipetted into the wells and BMP-4 present in a sample is bound to the wells by the immobilized antibody. The wells are washed and biotinylated anti- human BMP-4 antibody is added. After washing away unbound biotinylated antibody, HRP- conjugated streptavidin is pipetted to the wells. The wells are again washed, a TMB substrate solution is added to the wells and color develops in proportion to the amount of BMP-4 bound. The Stop Solution changes the color from blue to yellow, and the intensity of the color is
	measured at 450 nm. Reproducibility: Intra-Assay: CV<10% Inter-Assay: CV<12%.
Gene ID:	002
UniProt:	P12644
Pathways:	Steroid Hormone Mediated Signaling Pathway, Regulation of Muscle Cell Differentiation, Tube Formation, Skeletal Muscle Fiber Development

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/6 | Product datasheet for ABIN624951 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Application Details


Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 μ L of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 μ L of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	 Add 100 μL of prepared Streptavidin solution to each well. Incubate 45 min at RT.
	8. Add 100 μL of TMB One-Step Substrate Reagent to each well.
	9. Incubate 30 min at RT.
	10. Add 50 μ L of Stop Solution to each well.
	11. Read at 450 nm immediately.
Reagent Preparation:	1. Bring all reagents and samples to room temperature (18 - 25 °C) before use.
	2. Sample dilution: Tissue lysate and cell lysate sample should be diluted at least 5-folds with
	1x Sample Diluent Buffer.
	3. Sample Diluent Buffer (Item D) and Assay Diluent (Item E) should be diluted 5-folds with
	deionized or distilled water before use.
	4. Preparation of standard: Briefly spin the vial of Item C. Add 400 μL 1x Sample Diluent Buffer
	(Item D) into Item C vial to prepare a 50 ng/mL standard. Dissolve the powder thoroughly by a
	gentle mix. Add 80 μ L BMP-4 standard from the vial of Item C, into a tube with 586.7 μ L Sample
	Diluent Buffer to prepare a 6,000 pg/mL stock standard solution. Pipette 400 μ L 1x Sample
	Diluent Buffer into each tube. Use the stock standard solution to produce a dilution series . Mix
	each tube thoroughly before the next transfer. 1x Sample Diluent Buffer serves as the zero
	standard (0 pg/mL). 200 μL 200 μL 200 μL 200myl 200 μL 200 μL 80 μL standard + 586.7 μL
	6,000 2,000 666.7 222.2 74.07 24.69 8.23 0 pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL
	pg/mL
	5. If the Wash Concentrate (20x) (Item B) contains visible crystals, warm to room temperature
	and mix gently until dissolved. Dilute 20 ml of Wash Buffer Concentrate into deionized or
	distilled water to yield 400 ml of 1x Wash Buffer.
	6. Briefly spin the Detection Antibody vial (Item F) before use. Add 100 μL of 1x Assay Diuent
	into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (th
	concentrate can be stored at 4 °C for 5 days). The detection antibody concentrate should be
	diluted 65-foldss with 1x Assay Diuent and used in step 4 of Part VI Assay Procedure.
	7. Briefly spin the HRP-Streptavidin concentrate vial (Item G) before use. HRP-Streptavidin
	concentrate should be diluted 80-fold with 1x Assay Diuent. For example: Briefly spin the vial

	(Item G) and pipette up and down to mix gently . Add 125 μL of HRP-Streptavidin concentrate
	into a tube with 10 ml 1x Assay Diluent to prepare a final 80 fold diluted HRP-Streptavidin
	solution. Mix well.
	8. Cell lysate buffer (Item J) should be diluted 2-fold with deionized or distilled water (for cell
	lysate and tissue lysate).
Assay Procedure:	1. Bring all reagents and samples to room temperature (18 - 25 °C) before use. It is
	recommended that all standards and samples be run at least in duplicate.
	2. Add 100 µL of each standard (see Reagent Preparation step 2) and sample into appropriate
	wells. Cover well with plate holder and incubate for 2.5 hours at room temperature or over night
	at 4 °C with gentle shaking.
	3. Discard the solution and wash 4 times with 1x Wash Solution. Wash by filling each well with
	Wash Buffer (300 myl) using a multi-channel Pipette or autowasher. Complete removal of liquic
	at each step is essential to good performance. After the last wash, remove any remaining Wash
	Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
	4. Add 100 µL of 1x prepared biotinylated antibody (Reagent Preparation step 6) to each well.
	Incubate for 1 hour at room temperature with gentle shaking.
	5. Discard the solution. Repeat the wash as in step
	6. Add 100 µL of prepared Streptavidin solution (see Reagent Preparation step 7) to each well.
	Incubate for 45 minutes at room temperature with gentle shaking.
	7. Discard the solution. Repeat the wash as in step
	8. Add 100 µL of TMB One-Step Substrate Reagent (Item H) to each well. Incubate for 30
	minutes at room temperature in the dark with gentle shaking.
	9. Add 50 μL of Stop Solution (Item I) to each well. Read at 450 nm immediately.
Calculation of Results:	Calculate the mean absorbance for each set of duplicate standards, controls and samples, and
	subtract the average zero standard optical density. Plot the standard curve on log-log graph
	paper or using Sigma plot software, with standard concentration on the x-axis and absorbance
	on the y-axis. Draw the best-fit straight line through the standard points.
	Typical Data: These standard curves are for demonstration only. A standard curve must be run
	with each assay. Assay Diluent Human BMP-4 concentration (pg/mL) 1 10 100 1000 10000 0 [
	=4 50 n m 0.001 0.01 0.1 1 10
	Sensitivity: The minimum detectable dose of BMP-4 is typically less than 15 pg/mL.
	Recovery: Recovery was determined by spiking various levels of human BMP-4 into human
	tissue lysate and cell lysate. Mean recoveries are as follows: Sample Type Average % Recovery
	Range (%) Tissue lysate 89.47 78-102 Cell lysate 92.45 80-103
	Linearity: Sample Type Tissue Cell Lysate lysate 1:2 Average % of 84.31 85.75 Expected Range

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/6 | Product datasheet for ABIN624951 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Application Details	
	%) 77-103 77-102 1:4 Average % of 86.78 88.54 Expected Range (%) 79-102 80-104
	Reproducibility: Intra-Assay: CV<10 % Inter-Assay: CV<12 %
Assay Precision:	Intra-Assay: CV< 10 % Inter-Assay: CV< 12 %
Restrictions:	For Research Use only
Handling	
Handling Advice:	Avoid repeated freeze-thaw cycles.
Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated
	freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is
	recommended to store at -80°C.
Expiry Date:	6 months
Publications	
Product cited in:	Adali, Yorulmaz, Ozkanli, Ulasoglu, Bayraktar, Orhun, Colak, Tuncer: "Serum concentrations of
	insulin-like growth factor-binding protein 5 in Crohn's disease." in: World journal of
	gastroenterology, Vol. 19, Issue 47, pp. 9049-56, (2013) (PubMed).
	Colak, Senates, Ozturk, Yilmaz, Zemheri, Yilmaz Enc, Ulasoglu, Aksaray, Bozbeyoglu, Kiziltas,
	Kurdas, Tuncer et al.: "Serum concentrations of human insulin-like growth factor-1 and levels of
	insulin-like growth factor-binding protein-5 in patients with nonalcoholic fatty liver disease:
	association with liver" in: European journal of gastroenterology & hepatology, Vol. 24, Issue
	3, pp. 255-61, (2012) (PubMed).
	Schalla, Hearn, Taylor, Eavenson, Valdiserri, Essien: "CDC's Model Performance Evaluation
	Program: assessment of the quality of laboratory performance for HIV-1 antibody testing." in:
	Public health reports (Washington, D.C.: 1974), Vol. 105, Issue 2, pp. 167-71, (1990) (PubMed

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 5/6 | Product datasheet for ABIN624951 | 09/12/2023 | Copyright antibodies-online. All rights reserved. Images



Human BMP-4 concentration (pg/ml)

ELISA

Image 1.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 6/6 | Product datasheet for ABIN624951 | 09/12/2023 | Copyright antibodies-online. All rights reserved.