antibodies - online.com







IL17RA ELISA Kit





Publications



Overview

Quantity:	96 tests
Target:	IL17RA
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	10-3000 pg/mL
Minimum Detection Limit:	10 pg/mL
Application:	ELISA

Product Details

Purpose:	Human IL-17 RA ELISA Kit for cell culture supernatants, plasma, and serum samples.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA kit shows no cross-reactivity with the following cytokines tested: human Angiogenin,
	BDNF, BLC, ENA-78, FGF- 4, IL-1 alpha, IL-1 beta, IL-2, IL-3, IL-4, IL-5, IL-6, IL-7, IL-8, IL-9, IL-11, IL-
	12 p70, IL-12 p40, IL-13, IL-15, IL-17, I-309, IP-10, G-CSF, GM-CSF, IFN-gamma, Leptin (OB),
	MCP-1, MCP-2, MCP-3, MDC, MIP-1 alpha, MIP-1 beta, MIP-1 delta, PARC, PDGF, RANTES, SCF,
	TARC, TGF-beta, TIMP-1, TIMP-2, TNF-alpha, TNF-beta, TPO, VEGF.
Sensitivity:	10 pg/mL
Characteristics:	Strip plates and additional reagents allow for use in multiple experiments

Product Details

- · Quantitative protein detection
- · Establishes normal range
- · The best products for confirmation of antibody array data

Components:

- Pre-Coated 96-well Strip Microplate
- · Wash Buffer
- Stop Solution
- · Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

Material not included:

- · Distilled or deionized water
- Precision pipettes to deliver 2 μL to 1 μL volumes
- Adjustable 1-25 µL pipettes for reagent preparation
- 100 µL and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	IL17RA
Alternative Name:	IL-17R (IL17RA Products)
Background:	The Human IL-17R (Interleukin 17 receptor) ELISA (Enzyme-Linked Immunosorbent Assay) kit
	is an in vitro enzyme-linked immunosorbent assay for the quantitative measurement of human
	IL-17R in serum, plasma, cell culture supernatants and urine. This assay employs an antibody
	specific for human IL-17R coated on a 96-well plate. Standards and samples are pipetted into
	the wells and IL-17R present in a sample is bound to the wells by the immobilized antibody. The
	wells are washed and biotinylated anti-human IL-17R antibody is added. After washing away
	unbound biotinylated antibody, HRP-conjugated streptavidin is pipetted to the wells. The wells
	are again washed, a TMB substrate solution is added to the wells and color develops in
	proportion to the amount of IL-17R bound. The Stop Solution changes the color from blue to
	yellow, and the intensity of the color is measured at 450 nm. Reproducibility: Intra-Assay:
	CV<10% Inter-Assay: CV<12%.
Gene ID:	23765

Target Details

UniProt: Q96F46

Pathways: SARS-CoV-2 Protein Interactome

Application Details

Application Notes:	Recommended Dilution for serum and plasma samples2 fold
Sample Volume:	100 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 µL of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4 °C.
	4. Add 100 μL of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 μL of prepared Streptavidin solution to each well.
	7. Incubate 45 min at RT.
	8. Add 100 µL of TMB One-Step Substrate Reagent to each well.
	9. Incubate 30 min at RT.
	10. Add 50 µL of Stop Solution to each well.
	11. Read at 450 nm immediately.

Reagent Preparation:

1. Bring all reagents and samples to room temperature (18 - 25°C) before use. 2. Sample dilution: Assay Diluent C (Item L) is used for dilution of serum/plasma samples, 1x Assay Diluent B (Item E) can be used for dilution of cell culture supernates/urine. 3. Assay Diluent B should be diluted 5-fold with deionized or distilled water before use. 4. Preparation of standard: Briefly spin the vial of Item C and then add 400 µl Assay Diluent C (for serum/plasma samples) or 1x Assay Diluent B (for cell culture supernates/urine) into Item C vial to prepare a 50 ng/ml standard. Dissolve the powder thoroughly by a gentle mix. Add 30 µl IL-17R standard (50 ng/ml) from the vial of Item C, into a tube with 470 µl Assay Diluent C or 1x Assay Diluent B to prepare a 3,000 pg/ml standard solution. Pipette 400 µl Assay Diluent C or 1x Assay Diluent B into each tube. Use the 3,000 pg/ml standard solution to produce a dilution series (shown below). Mix each tube thoroughly before the next transfer. Gently vortex to mix. Assay Diluent C or 1x Assay Diluent B serves as the zero standard (0 pg/ml). 5. If the Wash Concentrate (20x) (Item B) contains visible crystals, warm to room temperature and mix gently until dissolved. Dilute 20 ml of Wash Buffer Concentrate into deionized or distilled water to yield 400 ml of 1x Wash Buffer. 6. Briefly spin the Detection Antibody vial (Item F) before use. Add 100 μ l of 1x Assay Diluent B into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4°C for 5 days). The detection antibody concentrate should be

diluted 80-fold with 1x Assay Diluent B and used in step 4 of Part VI Assay Procedure. 7. Briefly spin the HRP-Streptavidin concentrate vial (Item G) and pipette up and down to mix gently before use. HRP-Streptavidin concentrate should be diluted 10,000-fold with 1x Assay Diluent B. For example: Briefly spin the vial (Item G) and pipette up and down to mix gently . Add 2 μ l of HRP-Streptavidin concentrate into a tube with 198 μ l 1x Assay Diluent B to prepare a 100-fold diluted HRP- Streptavidin solution (do not store the diluted solution for next day use). Mix through and then pipette 100 μ l of prepared 100-fold diluted solution into a tube with 10 ml 1x Assay Diluent B to prepare a final 10,000 fold diluted HRP-Streptavidin solution.

Assay Procedure:

1. Bring all reagents and samples to room temperature ($18 - 25^{\circ}$ C) before use. It is recommended that all standards and samples be run at least in duplicate. 2. Add $100 \, \mu$ l of each standard (see Reagent Preparation step 2) and sample into appropriate wells. Cover well and incubate for 2.5 hours at room temperature or over night at 4°C with gentle shaking. 3. Discard the solution and wash 4 times with 1x Wash Solution. Wash by filling each well with Wash Buffer (300 μ l) using a multi-channel Pipette or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels. 4. Add $100 \, \mu$ l of 1x prepared biotinylated antibody (Reagent Preparation step 6) to each well. Incubate for 1 hour at room temperature with gentle shaking. 5. Discard the solution. Repeat the wash as in step 3. 6. Add $100 \, \mu$ l of prepared Streptavidin solution (see Reagent Preparation step 7) to each well. Incubate for 45 minutes at room temperature with gentle shaking. 7. Discard the solution. Repeat the wash as in step 3. 8. Add $100 \, \mu$ l of TMB One-Step Substrate Reagent (Item H) to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking. 9. Add $100 \, \mu$ l of Stop Solution (Item I) to each well. Read at 450 nm immediately.

Calculation of Results:

Calculate the mean absorbance for each set of duplicate standards, controls and samples, and subtract the average zero standard optical density. Plot the standard curve on log-log graph paper or using Sigma plot software, with standard concentration on the x-axis and absorbance on the y-axis. Draw the best-fit straight line through the standard points.

Restrictions:

For Research Use only

Handling

Storage: -20 °C

Storage Comment:

The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.

Expiry Date:

6 months

Publications

Product cited in:


Martin, Nguyen, Molinolo, Gutkind: "Accumulation of dephosphorylated 4EBP after mTOR inhibition with rapamycin is sufficient to disrupt paracrine transformation by the KSHV vGPCR oncogene." in: **Oncogene**, Vol. 33, Issue 18, pp. 2405-12, (2014) (PubMed).


Said, Frierson, Sanchez-Carbayo, Brekken, Theodorescu: "Loss of SPARC in bladder cancer enhances carcinogenesis and progression." in: **The Journal of clinical investigation**, Vol. 123, Issue 2, pp. 751-66, (2013) (PubMed).

Degoul, Borel, Jacquemot, Besse, Communal, Mishellany, Papon, Penault-Llorca, Donnarieix, Doly, Maigne, Miot-Noirault, Cayre, Cluzel, Moins, Chezal, Bonnet: "In vivo efficacy of melanoma internal radionuclide therapy with a 131I-labelled melanin-targeting heteroarylcarboxamide molecule." in: **International journal of cancer. Journal international du cancer**, Vol. 133, Issue 5, pp. 1042-53, (2013) (PubMed).

Cohen, Gitay-Goren, Neufeld, Levi: "High levels of biologically active vascular endothelial growth factor (VEGF) are produced by the baculovirus expression system." in: **Growth factors (Chur, Switzerland)**, Vol. 7, Issue 2, pp. 131-8, (1992) (PubMed).

Ferrara, Houck, Jakeman, Winer, Leung: "The vascular endothelial growth factor family of polypeptides." in: **Journal of cellular biochemistry**, Vol. 47, Issue 3, pp. 211-8, (1992) (PubMed).





ELISA

Image 1.