Datasheet for ABIN625033

CXCL10 ELISA Kit

1 Image | 32 Publications

Go to Product page

Overview

Quantity:	96 tests
Target:	CXCL10
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	8-6000 pg/mL
Minimum Detection Limit:	8 pg/mL
Application:	ELISA

Product Details

Purpose:	Human IP-10 (CXCL10) ELISA Kit for cell culture supernatants, plasma, and serum samples.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA kit shows no cross-reactivity with any of the cytokines tested: Human BDNF, BLC, ENA-78, FGF-4, IL-1 alpha, IL-1 beta, IL-2, IL-3, IL-4, IL-5, IL-7, IL-8, IL-9, IL-10, IL-11, IL-12 p70, IL-12 p40, IL-13, IL-15, I-309, G-CSF, GM-CSF, IFN-gamma, Leptin (OB), MCP-1, MCP- 2, MCP-3, MDC, MIP-1 alpha, MIP-1 beta, MIP-1 delta, PARC, PDGF, RANTES, SCF, TARC, TGF-beta, TIMP-1, TIMP-2, TNF-alpha, TNF-beta, TPO, VEGF.
Sensitivity:	< 8 pg/mL
Characteristics:	• Strip plates and additional reagents allow for use in multiple experiments

Product Details

	• Quantitative protein detection • Establishes normal range • The best products for confirmation of antibody array data
Components:	• Pre-Coated 96-well Strip Microplate • Wash Buffer • Stop Solution • Assay Diluent(s) • Lyophilized Standard • Biotinylated Detection Antibody • Streptavidin-Conjugated HRP • TMB One-Step Substrate
Material not included:	• Distilled or deionized water • Precision pipettes to deliver 2 µL to 1 µL volumes • Adjustable 1-25 µL pipettes for reagent preparation • 100 µL and 1 liter graduated cylinders • Tubes to prepare standard and sample dilutions • Absorbent paper • Microplate reader capable of measuring absorbance at 450nm • Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	CXCL10
Alternative Name:	IP-10 (CXCL10 Products)
Background:	IP-10 (Interferon-inducible protein-10) is also called gamma-IP-10 or INP-10. It has a length of 98 amino acids and belongs to the family of chemotactic cytokines. IP-10 has been detected in keratinocytes, lymphocytes, monocytes, and endothelial cells in immunologically mediated processes. It has been suggested that IP-10 may play an important role in hypersensitivity reactions of the delayed type. Increased levels of IP-10 are found in psoriatic plaques characterized by the infiltration of neutrophils. The Human IP-10 ELISA (Enzyme-Linked Immunosorbent Assay) kit is an in vitro enzyme-linked immunosorbent assay for the quantitative measurement of human IP-10 in serum, plasma, cell culture supernatants and urine. This assay employs an antibody specific for human IP-10 coated on a 96-well plate. Standards and samples are pipetted into the wells and IP-10 present in a sample is bound to the wells by the immobilized antibody. The wells are washed and biotinylated anti-human IP-10 antibody is added. After washing away unbound biotinylated antibody, HRP-conjugated streptavidin is pipetted to the wells. The wells are again washed, a TMB substrate solution is

Target Details

	added to the wells and color develops in proportion to the amount of IP-10 bound. The Stop Solution changes the color from blue to yellow, and the intensity of the color is measured at 450 nm. Reproducibility: Intra-Assay: CV<10% Inter-Assay: CV<12%.
Gene ID:	3627
UniProt:	P02778

Application Details

Application Notes:	Recommended Dilution for serum and plasma samples2 fold
Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual. 2. Add 100 µL of standard or sample to each well. 3. Incubate 2.5 h at RT or O/N at 4 °C. 4. Add 100 µL of prepared biotin antibody to each well. 5. Incubate 1 h at RT. 6. Add 100 µL of prepared Streptavidin solution to each well. 7. Incubate 45 min at RT. 8. Add 100 µL of TMB One-Step Substrate Reagent to each well. 9. Incubate 30 min at RT. 10. Add 50 µL of Stop Solution to each well. 11. Read at 450 nm immediately.
Reagent Preparation:	1. Bring all reagents and samples to room temperature (18-25 °C) before use. 2. Sample dilution: If your samples need to be diluted, Assay Diluent A (Item D) should be used for dilution of serum/plasma samples. 1x Assay Diluent B (Item E) should be used for dilution of culture supernatants and urine. Suggested dilution for normal serum/plasma: 2 fold*. *Please note that levels of the target protein may vary between different specimens. Optimal dilution factors for each sample must be determined by the investigator. 3. Assay Diluent B should be diluted 5-fold with deionized or distilled water. 4. Preparation of standard: Briefly spin the vial of Item C. Add 400 µL Assay Diluent A (for serum/plasma samples) or 1x Assay Diluent B (for cell culture medium and urine) into Item C vial to prepare a 50 ng/mL standard. Dissolve the powder thoroughly by a gentle mix. Add 80 µL IP-10 standard from the vial of Item C, into a tube with 586.7 µL Assay Diluent A or 1x Assay Diluent B to prepare a 6000 pg/mL stock standard solution. Pipette 400 µL Assay Diluent A or 1x Assay Diluent B into each tube. Use the stock standard solution to produce a dilution series . Mix each tube thoroughly before the next transfer. Assay Diluent A or 1x Assay Diluent B serves

	as the zero standard (0 pg/mL). 200 µL 80 µL standard +586.7 µL 200myl 200 µL 200 µL 200 µL 200 µL 6000 2000 666.7 222.2 74.07 24.69 8.23 0 pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL 5. If the Wash Concentrate (20x) (Item B) contains visible crystals, warm to room temperature and mix gently until dissolved. Dilute 20 ml of Wash Buffer Concentrate into deionized or distilled water to yield 400 ml of 1x Wash Buffer. 6. Briefly spin the Detection Antibody vial (Item F) before use. Add 100 µL of 1x Assay Diluent B into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4 °C for 5 days). The detection antibody concentrate should be diluted 100-fold with 1x Assay Diluent B and used in step 4 of Part VI Assay Procedure. 7. Briefly spin the HRP-Streptavidin concentrate vial (Item G) and pipette up and down to mix gently before use. HRP-Streptavidin concentrate should be diluted 500-fold with 1x Assay Diluent B. For example: Briefly spin the vial (Item G) and pipette up and down to mix gently . Add 30 µL of HRP-Streptavidin concentrate into a tube with 15 ml 1x Assay Diluent B to prepare a final 500 fold diluted HRP-Streptavidin solution (don't store the diluted solution for next day use). Mix well.
Assay Procedure:	1. Bring all reagents and samples to room temperature (18 - 25 °C) before use. It is recommended that all standards and samples be run at least in duplicate. 2. Add 100 µL of each standard (see Reagent Preparation step 2) and sample into appropriate wells. Cover well and incubate for 2.5 hours at room temperature or over night at 4 °C with gentle shaking. 3. Discard the solution and wash 4 times with 1x Wash Solution. Wash by filling each well with Wash Buffer (300 myl) using a multi-channel Pipette or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels. 4. Add 100 µL of 1x prepared biotinylated antibody (Reagent Preparation step 6) to each well. Incubate for 1 hour at room temperature with gentle shaking. 5. Discard the solution. Repeat the wash as in step 6. Add 100 µL of prepared Streptavidin solution (see Reagent Preparation step 7) to each well. Incubate for 45 minutes at room temperature with gentle shaking. 7. Discard the solution. Repeat the wash as in step 8. Add 100 µL of TMB One-Step Substrate Reagent (Item H) to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking. 9. Add 50 µL of Stop Solution (Item I) to each well. Read at 450 nm immediately.
Calculation of Results:	Calculate the mean absorbance for each set of duplicate standards, controls and samples, and

Application Details

	subtract the average zero standard optical density. Plot the standard curve on log-log graph paper or using Sigma plot software, with standard concentration on the x-axis and absorbance on the y-axis. Draw the best-fit straight line through the standard points. Typical Data: These standard curves are for demonstration only. A standard curve must be run with each assay. Assay Diluent A IP-10 concentration (pg/mL) 1 10 100 1000 10000 O D =4 50 n m 0.01 0.1 1 10 Assay Diluent B IP-10 concentration (pg/mL) 1 10 100 1000 10000 O D =4 50 n m 0.01 0.1 1 10 Sensitivity: The minimum detectable dose of IP-10 is typically less than 8 pg/mL. Recovery: Recovery was determined by spiking various levels of human IP-10 into human serum, plasma and cell culture media. Mean recoveries are as follows: Sample Type Average % Recovery Range (%) Serum 94.49 82-104 Plasma 96.13 83-103 Cell culture media 101.34 84-105 Linearity: Sample Type Serum Plasma Cell culture media 1:2 Average % of Expected 92 94 98 Range (%) 82-103 84-104 86-105 1:4 Average % of Expected 97 96 95 Range (%) 83-105 84-104 82-103 Reproducibility: Intra-Assay: CV<10 % Inter-Assay: CV<12 %
Assay Precision:	Intra-Assay: CV< 10 % Inter-Assay: CV< 12 %
Restrictions:	For Research Use only

Handling

Handling Advice:	Avoid repeated freeze-thaw cycles.
Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.
Expiry Date:	6 months

Publications

Product cited in:	Collignon, Perles-Barbacaru, Robert, Silvy, Martinez, Crenon, Germain, Garcia, Viola, Lombardo, Mas, Béraud: "A pancreatic tumor-specific biomarker characterized in humans and mice as an immunogenic onco-glycoprotein is efficient in dendritic cell vaccination." in: **Oncotarget**, Vol. 6, Issue 27, pp. 23462-79, (2015) (PubMed).

Publications

There are more publications referencing this product on: Product page

Images



ELISA

Image 1.