## .-online.com antibodies

# Datasheet for ABIN625411 IL-6 Receptor ELISA Kit

1 Image

1 Publication



#### Overview

Quantity:	96 tests
Target:	IL-6 Receptor (IL6R)
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	6-1500 pg/mL
Minimum Detection Limit:	6 pg/mL
Application:	ELISA

## Product Details

Purpose:	Mouse IL-6 R ELISA Kit for cell culture supernatants, plasma, and serum samples.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA kit shows no cross-reactivity with the following cytokines tested: Mouse CD30, L
	CD30, T CD40, CRG-2, CTACK, CXCL16, Eotaxin , Eotaxin-2, Fas Ligand, Fractalkine, GCSF, GM-
	CSF, IGFBP-3, IGFBP-5, IGFBP-6, IL-1 alpha, IL-1 beta, IL-2, IL-3 Rb, IL-4, IL-5, IL-9, IL-10, IL-12
	p70, IL-13, IL-17, KC, Leptin R, LEPTIN(OB), LIX, L-Selectin, Lymphotactin, MCP-1, MCP-5, M-
	CSF, MIG, MIP-1 alpha, MIP-1 gamma, MIP-2, MIP-3 beta, MIP-3 alpha, PF-4, P-Selectin,
	RANTES, SCF, SDF-1 alpha, TARC, TCA-3, TECK, TIMP-1, TNF-alpha, TNF RI, TNF RII, TPO,
	VCAM-1, VEGF
Sensitivity:	< 6 pg/mL

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/6 | Product datasheet for ABIN625411 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

#### Product Details

Characteristics:	<ul> <li>Strip plates and additional reagents allow for use in multiple experiments</li> <li>Quantitative protein detection</li> <li>Establishes normal range</li> <li>The best products for confirmation of antibody array data</li> </ul>
Components:	Pre-Coated 96-well Strip Microplate
	• Wash Buffer
	Stop Solution
	Assay Diluent(s)
	Lyophilized Standard
	Biotinylated Detection Antibody
	Streptavidin-Conjugated HRP
	TMB One-Step Substrate
Material not included:	Distilled or deionized water
	<ul> <li>Precision pipettes to deliver 2 µL to 1 µL volumes</li> </ul>
	<ul> <li>Adjustable 1-25 µL pipettes for reagent preparation</li> </ul>
	<ul> <li>100 μL and 1 liter graduated cylinders</li> </ul>
	<ul> <li>Tubes to prepare standard and sample dilutions</li> </ul>
	Absorbent paper
	Microplate reader capable of measuring absorbance at 450nm
	<ul> <li>Log-log graph paper or computer and software for ELISA data analysis</li> </ul>

#### Target Details

Target:	IL-6 Receptor (IL6R)
Alternative Name:	IL-6 R (IL6R Products)
Gene ID:	16194
UniProt:	P22272
Pathways:	JAK-STAT Signaling, Autophagy, Growth Factor Binding, Cancer Immune Checkpoints

### Application Details

Application Notes:	Recommended Dilution for serum and plasma samples20 - 200 fold
Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual. 2. Add 100 $\mu L$ of standard or sample to each well.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/6 | Product datasheet for ABIN625411 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

	<ol> <li>Incubate 2.5 h at RT or O/N at 4 °C.</li> <li>Add 100 μL of prepared biotin antibody to each well.</li> <li>Incubate 1 h at RT.</li> <li>Add 100 μL of prepared Streptavidin solution to each well.</li> <li>Incubate 45 min at RT.</li> <li>Add 100 μL of TMB One-Step Substrate Reagent to each well.</li> <li>Incubate 30 min at RT.</li> <li>Add 50 μL of Stop Solution to each well.</li> <li>Read at 450 nm immediately.</li> </ol>
Reagent Preparation:	<ol> <li>Bring all reagents and samples to room temperature (18 - 25 °C) before use.</li> <li>Sample dilution: If your samples need to be diluted, Assay Diluent A (Item D) should be used for dilution of serum/plasma samples. 1x Assay Diluent B (Item E) should be used for dilution of culture supernatants. Suggested dilution for normal serum/plasma: 20-200 fold*. * Please note that levels of the target protein may vary between different specimens. Optimal dilution factors for each sample must be determined by the investigator.</li> <li>Assay Diluent B should be diluted 5-fold with deionized or distilled water before use.</li> <li>Preparation of standard: Briefly spin the vial of Item C and then add 400 µL Assay Diluent A (for serum/plasma samples) or 1x Assay Diluent B (for cell culture supernates) into Item C vial to prepare a 50 ng/mL standard. Dissolve the powder thoroughly by a gentle mix. Add 15 µL IL-6R standard (50 ng/mL) from the vial of Item C, into a tube with 485 µL Assay Diluent A or 1x Assay Diluent B to prepare a 1,500 pg/mL standard solution. Pipette 300 µL Assay Diluent A or 1x Assay Diluent B into each tube. Use the 1,500 pg/mL standard solution to produce a dilution series . Mix each tube thoroughly before the next transfer. Assay Diluent A or 1x Assay Diluent B serves as the zero standard (0 pg/mL). 200 µL 200 µL 200 µL 200 µL 200 µL 15 µL standard + 485 µL 1,500 600 240 96 38.4 15.36 6.14 0 pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL</li> <li>5. If the Wash Concentrate (20x) (Item B) contains visible crystals, warm to room temperature and mix gently until dissolved. Dilute 20 ml of Wash Buffer.</li> <li>6. Briefly spin the Detection Antibody vial (Item F) before use. Add 100 µL of 1x Assay Diluent B into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4 °C for 5 days). The detection antibody concentrate should be diluted 80-fold with 1x Assay Diluent B and used in step 4 of Part VI Assay</li></ol>
	Diluent B. For example: Briefly spin the vial (Item G) and pipette up and down to mix gently . Add


	30 µL of HRP-Streptavidin concentrate into a tube with 12 ml 1x Assay Diluent B to prepare a
	400-fold diluted HRP- Streptavidin solution (don't store the diluted solution for next day use).
	Mix well.
Assay Procedure:	1. Bring all reagents and samples to room temperature (18 - 25 °C) before use. It is
	recommended that all standards and samples be run at least in duplicate.
	2. Add 100 $\mu$ L of each standard (see Reagent Preparation step 2) and sample into appropriate
	wells. Cover well and incubate for 2.5 hours at room temperature or over night at 4 °C with
	gentle shaking.
	3. Discard the solution and wash 4 times with 1x Wash Solution. Wash by filling each well with
	Wash Buffer (300 myl) using a multi-channel Pipette or autowasher. Complete removal of liqui
	at each step is essential to good performance. After the last wash, remove any remaining Was
	Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
	4. Add 100 µL of 1x prepared biotinylated antibody (Reagent Preparation step 6) to each well.
	Incubate for 1 hour at room temperature with gentle shaking.
	5. Discard the solution. Repeat the wash as in step
	6. Add 100 µL of prepared Streptavidin solution (see Reagent Preparation step 7) to each well.
	Incubate for 45 minutes at room temperature with gentle shaking.
	7. Discard the solution. Repeat the wash as in step
	8. Add 100 $\mu$ L of TMB One-Step Substrate Reagent (Item H) to each well. Incubate for 30
	minutes at room temperature in the dark with gentle shaking.
	9. Add 50 $\mu L$ of Stop Solution (Item I) to each well. Read at 450 nm immediately.
Calculation of Results:	Calculate the mean absorbance for each set of duplicate standards, controls and samples, and
	subtract the average zero standard optical density. Plot the standard curve on log-log graph
	paper or using Sigma plot software, with standard concentration on the x-axis and absorbance
	on the y-axis. Draw the best-fit straight line through the standard points.
	Typical Data: These standard curves are for demonstration only. A standard curve must be rur
	with each assay. Assay Diluent A Mouse IL-6R concentration (pg/mL) 1 10 100 1000 10000 0
	=4 50 n m 0.01 0.1 1 10 Assay Diluent B Mouse IL-6R concentration (pg/mL) 1 10 100 1000
	10000 O D =4 50 n m 0.01 0.1 1 10
	Sensitivity: The minimum detectable dose of IL-6R is typically less than 6 pg/mL.
	Recovery: Recovery was determined by spiking various levels mouse IL-6R into mouse serum,
	plasma and cell culture media. Mean recoveries are as follows: Sample Type Average %
	Recovery Range ( %) Serum 88.93 82-97 Plasma 89.63 85-95 Cell culture media 103.2 95-111
	Linearity: Sample Type Serum Plasma Cell Culture Media 1:2 Average % of Expected 102.4
	103.8 105.3 Range ( %) 112-133 96-112 97-113 1:4 Average % of Expected 94.13 103.3 112.6

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/6 | Product datasheet for ABIN625411 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

	Range ( %) 86-102 95-111 105-121
	Reproducibility: Intra-Assay: CV<10 % Inter-Assay: CV<12 %
Assay Precision:	Intra-Assay: CV< 10 % Inter-Assay: CV< 12 %
Restrictions:	For Research Use only
Handling	
Handling Advice:	Avoid repeated freeze-thaw cycles.
Storage:	-20 °C
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeate freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.
Expiry Date:	6 months
Publications	
Product cited in:	Zheng, Pan, Wang, Liu, Shi, Ding: "HMGB1 Turns Renal Tubular Epithelial Cells into
	Inflammatory Promoters by Interacting with TLR4 During Sepsis." in: Journal of interferon &
	cytokine research : the official journal of the International Society for Interferon and
	Cytokine Research, Vol. 36, Issue 1, pp. 9-19, (2016) (PubMed).
	Chang, Jia, Sun, Xu, Wu, Zhang, Wei: "APRIL promotes proliferation, secretion and invasion of
	fibroblast-like synoviocyte from rats with adjuvant induced arthritis." in: Molecular immunology
	Vol. 64, Issue 1, pp. 90-8, (2015) (PubMed).
	Jiang, Zhang, Tian, Wang, Zhao, Wang, Li, Liu, Li, Zhang, Guan: "The monoacylglycerol lipase
	inhibitor JZL184 decreases inflammatory response in skeletal muscle contusion in rats." in: <b>European journal of pharmacology</b> , Vol. 761, pp. 1-10, (2015) (PubMed).
	Zhang, Jiang, Tian, Wang, Zhao, Wang, Li, Liu, Li, Zhang, Guan: "CB2R orchestrates fibrogenesis through regulation of inflammatory response during the repair of skeletal muscle contusion." ir <b>International journal of clinical and experimental pathology</b> , Vol. 8, Issue 4, pp. 3491-502, ( 2015) (PubMed).
	Willett, Thote, Lin, Moran, Raji, Sridaran, Stevens, Guldberg: "Intra-articular injection of

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 5/6 | Product datasheet for ABIN625411 | 09/12/2023 | Copyright antibodies-online. All rights reserved. micronized dehydrated human amnion/chorion membrane attenuates osteoarthritis development." in: **Arthritis research & therapy**, Vol. 16, Issue 1, pp. R47, (2014) (PubMed).

#### Images



## ELISA

Image 1.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 6/6 | Product datasheet for ABIN625411 | 09/12/2023 | Copyright antibodies-online. All rights reserved.