

Datasheet for ABIN625425  
**Sonic Hedgehog ELISA Kit**[Go to Product page](#)

## 1 Image

## Overview

Quantity:	96 tests
Target:	Sonic Hedgehog (SHH)
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	5-1000 pg/mL
Minimum Detection Limit:	5 pg/mL
Application:	ELISA

## Product Details

Purpose:	Mouse Sonic Hedgehog N-Terminal (Shh-N) ELISA Kit for cell culture supernatants, EDTA and/or Citrate treated plasma, and serum samples. Heparin is not recommended as an anticoagulant for use in this assay.
Sample Type:	Plasma, Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA kit shows no cross-reactivity with the following cytokines tested: Mouse CD30, L CD30, T CD40, CRG-2, CTACK, CXCL16, Eotaxin , Eotaxin-2, Fas Ligand, Fractalkine, GCSF, GM-CSF, IFN- gamma, IGFBP-3, IGFBP-5, IGFBP-6, IL-1 alpha, IL-1 beta, IL-2, IL-4, IL-9, IL-10, IL-13, KC, Leptin R, LEPTIN(OB), LIX, L-Selectin, Lymphotactin, MCP-1, MCP-5, M-CSF, MIG, MIP-1 alpha, MIP-1 gamma, MIP-2, MIP-3 beta, MIP-3 alpha, PF-4, P-Selectin, RANTES, SCF, SDF-1 alpha, TARC, TCA- 3, TECK, TIMP-1, TNF RI, TNF RII, TPO, VCAM-1, VEGF.

## Product Details

Cross-Reactivity (Details):	This ELISA kit shows no cross-reactivity with the following cytokines tested: Mouse CD30, L CD30, T CD40, CRG-2, CTACK, CXCL16, Eotaxin , Eotaxin-2, Fas Ligand, Fractalkine, GCSF, GM-CFS, IFN- gamma, IGFBP-3, IGFBP-5, IGFBP-6, IL-1 alpha, IL-1beta, IL-2, IL-4, IL-9, IL-10, IL-13, KC, Leptin R, LEPTIN(OB), LIX, L-Selectin, Lymphotactin, MCP-1, MCP-5, M-CSF, MIG, MIP-1alpha, MIP-1gamma, MIP-2, MIP-3beta, MIP-3alpha, PF-4, P-Selectin, RANTES, SCF, SDF-1alpha, TARC, TCA- 3, TECK, TIMP-1, TNF RI, TNF RII, TPO, VCAM-1, VEGF.
Sensitivity:	< 5 pg/mL
Characteristics:	<ul style="list-style-type: none"><li>• Strip plates and additional reagents allow for use in multiple experiments</li><li>• Quantitative protein detection</li><li>• Establishes normal range</li><li>• The best products for confirmation of antibody array data</li></ul>
Components:	<ul style="list-style-type: none"><li>• Pre-Coated 96-well Strip Microplate</li><li>• Wash Buffer</li><li>• Stop Solution</li><li>• Assay Diluent(s)</li><li>• Lyophilized Standard</li><li>• Biotinylated Detection Antibody</li><li>• Streptavidin-Conjugated HRP</li><li>• TMB One-Step Substrate</li></ul>
Material not included:	<ul style="list-style-type: none"><li>• Distilled or deionized water</li><li>• Precision pipettes to deliver 2 µL to 1 µL volumes</li><li>• Adjustable 1-25 µL pipettes for reagent preparation</li><li>• 100 µL and 1 liter graduated cylinders</li><li>• Tubes to prepare standard and sample dilutions</li><li>• Absorbent paper</li><li>• Microplate reader capable of measuring absorbance at 450nm</li><li>• Log-log graph paper or computer and software for ELISA data analysis</li></ul>

## Target Details

Target:	Sonic Hedgehog (SHH)
Alternative Name:	Sonic Hedgehog ( <a href="#">SHH Products</a> )
Background:	Sonic hedgehog protein (SHH) (HHG-1)
Gene ID:	20423
UniProt:	<a href="#">Q62226</a>

## Target Details

Pathways: [Hedgehog Signaling](#), [Dopaminergic Neurogenesis](#), [Regulation of Muscle Cell Differentiation](#), [Tube Formation](#), [Skeletal Muscle Fiber Development](#)

## Application Details

Application Notes:	Recommended Dilution for serum and plasma samples2 fold
Sample Volume:	100 µL
Plate:	Pre-coated
Protocol:	<ol style="list-style-type: none"><li>1. Prepare all reagents, samples and standards as instructed in the manual.</li><li>2. Add 100 µL of standard or sample to each well.</li><li>3. Incubate 2.5 h at RT or O/N at 4 °C.</li><li>4. Add 100 µL of prepared biotin antibody to each well.</li><li>5. Incubate 1 h at RT.</li><li>6. Add 100 µL of prepared Streptavidin solution to each well.</li><li>7. Incubate 45 min at RT.</li><li>8. Add 100 µL of TMB One-Step Substrate Reagent to each well.</li><li>9. Incubate 30 min at RT.</li><li>10. Add 50 µL of Stop Solution to each well.</li><li>11. Read at 450 nm immediately.</li></ol>
Reagent Preparation:	<ol style="list-style-type: none"><li>1. Bring all reagents and samples to room temperature (18 - 25 °C) before use.</li><li>2. Sample dilution: If your samples need to be diluted, Assay Diluent A (Item D) should be used for dilution of serum/plasma samples. 1x Assay Diluent B (Item E) should be used for dilution of culture supernatants and urine. Suggested dilution for normal serum/plasma: 2 fold*. * Please note that levels of the target protein may vary between different specimens. Optimal dilution factors for each sample must be determined by the investigator.</li><li>3. Assay Diluent B should be diluted 5-fold with deionized or distilled water before use.</li><li>4. Preparation of standard: Briefly spin the vial of Item C and then add 400 µL Assay Diluent A (for serum/plasma samples) or 1x Assay Diluent B (for cell culture supernates) into Item C vial to prepare a 50 ng/mL standard. Dissolve the powder thoroughly by a gentle mix. Add 20 µL Shh-N standard (50 ng/mL) from the vial of Item C, into a tube with 980 µL Assay Diluent A or 1x Assay Diluent B to prepare a 1,000 pg/mL standard solution. Pipette 300 µL Assay Diluent A or 1x Assay Diluent B into each tube. Use the 1,000 pg/mL standard solution to produce a dilution series . Mix each tube thoroughly before the next transfer. Assay Diluent A or 1x Assay Diluent B serves as the zero standard (0 pg/mL). 20 µL standard + 980 µL 200 µL 200 µL 200 µL 200 µL 200 µL 200µl 1,000 400 160 64 25.6 10.24 4.10 0 pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL pg/mL</li></ol>

5. If the Wash Concentrate (20x) (Item B) contains visible crystals, warm to room temperature and mix gently until dissolved. Dilute 20 ml of Wash Buffer Concentrate into deionized or distilled water to yield 400 ml of 1x Wash Buffer.
6. Briefly spin the Detection Antibody vial (Item F) before use. Add 100  $\mu$ L of 1x Assay Diluent B into the vial to prepare a detection antibody concentrate. Pipette up and down to mix gently (the concentrate can be stored at 4 °C for 5 days). The detection antibody concentrate should be diluted 80-fold with 1x Assay Diluent B and used in step 4 of Part VI Assay Procedure.
7. Briefly spin the HRP-Streptavidin concentrate vial (Item G) and pipette up and down to mix gently before use. HRP-Streptavidin concentrate should be diluted 300-fold with 1x Assay Diluent B. For example: Briefly spin the vial (Item G) and pipette up and down to mix gently . Add 40  $\mu$ L of HRP-Streptavidin concentrate into a tube with 12 ml 1x Assay Diluent B to prepare a 300-fold diluted HRP- Streptavidin solution (don't store the diluted solution for next day use). Mix well.

---

### Assay Procedure:

1. Bring all reagents and samples to room temperature (18 - 25 °C) before use. It is recommended that all standards and samples be run at least in duplicate.
2. Add 100  $\mu$ L of each standard (see Reagent Preparation step 2) and sample into appropriate wells. Cover well and incubate for 2.5 hours at room temperature or over night at 4 °C with gentle shaking.
3. Discard the solution and wash 4 times with 1x Wash Solution. Wash by filling each well with Wash Buffer (300  $\mu$ l) using a multi-channel Pipette or autowasher. Complete removal of liquid at each step is essential to good performance. After the last wash, remove any remaining Wash Buffer by aspirating or decanting. Invert the plate and blot it against clean paper towels.
4. Add 100  $\mu$ L of 1x prepared biotinylated antibody (Reagent Preparation step 6) to each well. Incubate for 1 hour at room temperature with gentle shaking.
5. Discard the solution. Repeat the wash as in step
6. Add 100  $\mu$ L of prepared Streptavidin solution (see Reagent Preparation step 7) to each well. Incubate for 45 minutes at room temperature with gentle shaking.
7. Discard the solution. Repeat the wash as in step
8. Add 100  $\mu$ L of TMB One-Step Substrate Reagent (Item H) to each well. Incubate for 30 minutes at room temperature in the dark with gentle shaking.
9. Add 50  $\mu$ L of Stop Solution (Item I) to each well. Read at 450 nm immediately.

---

### Calculation of Results:

Calculate the mean absorbance for each set of duplicate standards, controls and samples, and subtract the average zero standard optical density. Plot the standard curve on log-log graph paper or using Sigma plot software, with standard concentration on the x-axis and absorbance on the y-axis. Draw the best-fit straight line through the standard points.

---

## Application Details

Typical Data: These standard curves are for demonstration only. A standard curve must be run with each assay. Assay Diluent A Mouse Shh-N concentration (pg/mL) O D =4 50 n m 0.01 0.1 1 10 1 10 100 1,000 10,000 Assay Diluent B Mouse Shh-N concentration (pg/mL) O D =4 50 n m 0.01 0.1 1 10 1 10 100 1,000 10,000

Sensitivity: The minimum detectable dose of Shh-N is typically less than 5 pg/mL.

Recovery: Recovery was determined by spiking various levels mouse Shh-N into mouse serum, plasma (EDTA and citrate) and cell culture media. Mean recoveries are as follows: Sample Type Average % Recovery Range ( %) Serum 78.88 71-87 Plasma 77.03 71-79 Cell culture media 82.11 78-86

Linearity: Sample Type Serum Plasma Cell Culture Media 1:2 Average % of Expected 99.00 90.69 84.65 Range ( %) 91-107 83-103 76-92 1:4 Average % of Expected 92.99 85.01 79.86 Range ( %) 85-101 68-83 71-87

Reproducibility: Intra-Assay: CV<10 % Inter-Assay: CV<12 %

Assay Precision:	Intra-Assay: CV< 10 % Inter-Assay: CV< 12 %
------------------	---

Restrictions:	For Research Use only
---------------	-----------------------

## Handling

Handling Advice:	Avoid repeated freeze-thaw cycles.
------------------	------------------------------------

Storage:	-20 °C
----------	--------

Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.
------------------	--

Expiry Date:	6 months
--------------	----------

ELISA

Image 1.

