antibodies .- online.com





Datasheet for ABIN6384386

CD59 IQ-ELISA Kit



Overview

Quantity:	96 tests
Target:	CD59
Reactivity:	Human
Method Type:	Sandwich ELISA
Application:	ELISA
Product Details	
Purpose:	Human Immunoquantitative (PCR-Based) CD59 ELISA Kit for cell culture supernatants, plasma,
	and serum samples.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Semi-Quantitative
Detection Method:	qPCR
Sensitivity:	4 pg/mL
Characteristics:	IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and
	samples are pipetted into the wells, the target protein in the standards and samples binds to
	the immobilized antibody. The wells are washed and the detection affinity reagent is added to
	the wells where it binds to any captured antigen. The wells are washed, and primers and PCR
	master mix are added to each well. The plate is placed into a real time PCR instrument for
	cycling and measurement of DNA amplification. The cycle number where amplification is
	detected is proportional to the amount of affinity detection reagent that bound to captured
	antigen in each well.

Product Details

Components:

- CD59 Microplate (Item A): 96 well PCR plate coated with anti-Human CD59
- Wash Buffer I Concentrate (20x) (Item B): 25 ml of 20x concentrated solution
- · Standards (Item C): 2 vials of recombinant Human CD59
- Assay Diluent A (Item D): 30 ml diluent buffer, 0.09% sodium azide as preservative. For Standard/Sample (serum/plasma) diluent
- Assay Diluent B (Item E): 15 ml of 5x concentrated buffer. For Standard/Sample (cell culture medium/urine) diluent
- Detection Affinity Reagent for CD59 (Item F): 2 vials of a 4x concentrated solution of anti-Human CD59 affinity reagent
- IQELISA Detection Reagent (Item G): 1.4ml of a 10x concentrated stock
- Primer Solution (Item I): 1.7 ml vial
- PCR Master Mix (Item J): 1.2 ml vial
- PCR Preparation buffer (Item K): 1ml vial of 10x concentrated buffer
- · Final Wash Buffer (Item L): 10ml vial of 10x concentrated buffer

Material not included:

- · Real-time PCR instrument, Bio-Rad recommended
- Precision pipettes to deliver 2 µL to 1 mL volumes
- · Adjustable 1-25 mL pipettes for reagent preparation
- · 100 mL and 1 liter graduated cylinders
- · Absorbent paper
- · Distilled or deionized water
- · Log-log graph paper or computer and software for data analysis
- · Tubes to prepare standard or sample dilutions
- Heating block or water bath capable of 80°C

Target Details

Target:	CD59
Alternative Name:	CD59 (CD59 Products)
Gene ID:	966
UniProt:	P13987
Pathways:	Complement System

Application Details

Application Notes:

The Immuno-Quantitative ELISA (IQELISA) kits are an innovative assay platform that combines the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. This results in an assay that is simultaneously familiar and cutting edge and enables the use of only 1/10th the sample volume while also providing 10x more sensitivity than a traditional ELISA.

Application Details

Comment:	The Immuno-Quantitative ELISA (IQELISA™) kits are an innovative assay platform that
	combines the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. Also
	called immuno-PCR, this detection platform results in an assay that is simultaneously familiar
	and cutting edge. Compared to traditional ELISA, IQELISA™ enables the use of only 1/10th the
	sample volume while also providing 10x more sensitivity.
Sample Volume:	25 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed
	2. Add 25 μ L standard or sample to each well. Incubate for 2.5 hours at room temperature or overnight at 4 $^{\circ}$ C
	3. Add 25 µL detection affinity reagent to each well. Incubate 1 hour at room temperature
	4. Add 25µL of IQELISA Detection Reagent to each well. Incubate 1 hour
	5. Add 15µL Primer solution 10µL of PCR master mix to each well
	6. Run real-time PCR
Restrictions:	For Research Use only
Handling	
Storage:	4 °C,-20 °C,-80 °C
Storage Comment:	May be stored for up to 6 months at 2° to 8°C from the date of shipment. Standard
	(recombinant protein) should be stored at -20°C or -80°C (recommended at -80°C) after
	reconstitution. Opened PCR plate or reagents may be stored for up to 1 month at 2° to 8°C.
	Note: the kit can be used within one year if the whole kit is stored at -20°C. Avoid repeated
	freeze-thaw cycles.
Expiry Date:	6 months