.-online.com antibodies

Datasheet for ABIN6385105 CCL19 IQ-ELISA Kit



Overview

Quantity:	96 tests
Target:	CCL19
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Application:	ELISA
Product Details	
Purpose:	Mouse Immunoquantitative (PCR-Based) MIP-3 beta ELISA Kit for cell culture supernatants,
	plasma, and serum samples.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Semi-Quantitative
Detection Method:	qPCR
Sensitivity:	0.5 pg/mL
Characteristics:	IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and
	samples are pipetted into the wells, the target protein in the standards and samples binds to
	the immobilized antibody. The wells are washed and the detection affinity reagent is added to
	the wells where it binds to any captured antigen. The wells are washed, and primers and PCR
	master mix are added to each well. The plate is placed into a real time PCR instrument for
	cycling and measurement of DNA amplification. The cycle number where amplification is
	detected is proportional to the amount of affinity detection reagent that bound to captured
	antigen in each well.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN6385105 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

Product Details

Components:	 MIP-3 beta Microplate (Item A): 96 well PCR plate coated with anti-Mouse MIP-3 beta Wash Buffer I Concentrate (20x) (Item B): 25 ml of 20x concentrated solution Standards (Item C): 2 vials of recombinant Mouse MIP-3 beta Assay Diluent A (Item D): 30 ml diluent buffer, 0.09% sodium azide as preservative. For Standard/Sample (serum/plasma) diluent Assay Diluent B (Item E): 15 ml of 5x concentrated buffer. For Standard/Sample (cell culture medium/urine) diluent Detection Affinity Reagent for MIP-3 beta (Item F): 2 vials of a 4x concentrated solution of anti-Mouse MIP-3 beta affinity reagent IQELISA Detection Reagent (Item G): 1.4ml of a 10x concentrated stock Primer Solution (Item I): 1.7 ml vial PCR Master Mix (Item J): 1.2 ml vial
	PCR Preparation buffer (Item K): 1ml vial of 10x concentrated buffer Final Week Puffer (Item L): 10ml vial of 10x concentrated buffer
	Final Wash Buffer (Item L): 10ml vial of 10x concentrated buffer
Material not included:	Real-time PCR instrument, Bio-Rad recommended
	 Precision pipettes to deliver 2 µL to 1 mL volumes
	 Adjustable 1-25 mL pipettes for reagent preparation
	 100 mL and 1 liter graduated cylinders
	Absorbent paper
	Distilled or deionized water
	 Log-log graph paper or computer and software for data analysis
	Tubes to prepare standard or sample dilutions
	 Heating block or water bath capable of 80°C

Target Details

Target:	CCL19
Alternative Name:	MIP-3 beta (CCL19 Products)
UniProt:	070460
Pathways:	Positive Regulation of Immune Effector Process
Application Details	
Application Notes:	The Immuno-Quantitative ELISA (IQELISA) kits are an innovative assay platform that combines the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. This results in an assay that is simultaneously familiar and cutting edge and enables the use of only 1/10th the sample volume while also providing 10x more sensitivity than a traditional ELISA.
Comment:	The Immuno-Quantitative ELISA (IQELISA™) kits are an innovative assay platform that combines the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. Also

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN6385105 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

	called immuno-PCR, this detection platform results in an assay that is simultaneously familiar and cutting edge. Compared to traditional ELISA, IQELISA™ enables the use of only 1/10th the sample volume while also providing 10x more sensitivity.
Sample Volume:	25 µL
Plate:	Pre-coated
Protocol:	 Prepare all reagents, samples and standards as instructed Add 25 μL standard or sample to each well. Incubate for 2.5 hours at room temperature or overnight at 4°C
	3. Add 25 µL detection affinity reagent to each well. Incubate 1 hour at room temperature
	 Add 25µL of IQELISA Detection Reagent to each well. Incubate 1 hour Add 15µL Primer solution 10µL of PCR master mix to each well
	6. Run real-time PCR
Restrictions:	For Research Use only
Handling	
Storage:	4 °C,-20 °C,-80 °C
Storage Comment:	May be stored for up to 6 months at 2° to 8°C from the date of shipment. Standard
	(recombinant protein) should be stored at -20°C or -80°C (recommended at -80°C) after
	reconstitution. Opened PCR plate or reagents may be stored for up to 1 month at 2° to 8°C.
	Note: the kit can be used within one year if the whole kit is stored at -20°C. Avoid repeated
	freeze-thaw cycles.
Expiry Date:	6 months