.-online.com antibodies

Datasheet for ABIN6385358 CDKN1B IQ-ELISA Kit



Overview

Quantity: 96 tests Target: CDKN1B Reactivity: Mouse Method Type: Sandwich ELISA Application: ELISA Product Details Product Details Purpose: Mouse Immunoquantitative (PCR-Based) p27/Kip1 ELISA Kit for cell culture supernatar plasma, and serum samples. Sample Type: Cell Culture Supernatant, Plasma, Serum Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISA semploy specific capture antibodies coated on a 96-well PCR plate. Standards samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture antigen in each well.	
Reactivity: Mouse Method Type: Sandwich ELISA Application: ELISA Product Details ELISA Purpose: Mouse Immunoquantitative (PCR-Based) p27/Kip1 ELISA Kit for cell culture supernatar plasma, and serum samples. Sample Type: Cell Culture Supernatant, Plasma, Serum Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture detecting the tamount of affinity detection reagent that bou	
Method Type:Sandwich ELISAApplication:ELISAProduct DetailsMouse Immunoquantitative (PCR-Based) p27/Kip1 ELISA Kit for cell culture supernatar plasma, and serum samples.Sample Type:Cell Culture Supernatant, Plasma, SerumAnalytical Method:Serni-QuantitativeDetection Method:qPCRCharacteristics:IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture	
Application: ELISA Product Details ELISA Purpose: Mouse Immunoquantitative (PCR-Based) p27/Kip1 ELISA Kit for cell culture supernatar plasma, and serum samples. Sample Type: Cell Culture Supernatant, Plasma, Serum Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standarde samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture detection reagent that bound to capture detected is proportional to the amount of affinity detection reagent that bound to capture detected is proportional to the amount of affinity detection reagent that bound to capture detected is proportional to the amount of affinity detection reagent that bound to capture detected is proportional to the amount of affinity detection reagent that bound to capture detected is proportional to the amount of affinity detection reagent that bound to capture detected is proportional to the amount of affinity detection reagent that bound to capture detected is proportional to the amount of affinity detection reagent that bound to capture detected is proportional to the amount of affinity detection reagent that bound to capture detected is proportional to the amount of affinity detection reagent that bound to capture detected is proportional to the amount of affinity detect	
Product DetailsPurpose:Mouse Immunoquantitative (PCR-Based) p27/Kip1 ELISA Kit for cell culture supernatar plasma, and serum samples.Sample Type:Cell Culture Supernatant, Plasma, SerumAnalytical Method:Semi-QuantitativeDetection Method:qPCRCharacteristics:IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture	
Purpose:Mouse Immunoquantitative (PCR-Based) p27/Kip1 ELISA Kit for cell culture supernatar plasma, and serum samples.Sample Type:Cell Culture Supernatant, Plasma, SerumAnalytical Method:Semi-QuantitativeDetection Method:qPCRCharacteristics:IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture	
plasma, and serum samples.Sample Type:Cell Culture Supernatant, Plasma, SerumAnalytical Method:Semi-QuantitativeDetection Method:qPCRCharacteristics:IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture	
Sample Type:Cell Culture Supernatant, Plasma, SerumAnalytical Method:Semi-QuantitativeDetection Method:qPCRCharacteristics:IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture	ants,
Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture	
Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers and master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture detection affinity detection reagent that bound to capture detection affinity detection reagent that bound to capture detected is proportional to the amount of affinity detection reagent that bound to capture detection reagent that bound to capture detection affinity detection reagent that bound to capture detection detection detected is proportional to the amount of affinity detection reagent that bound to capture detection detected is proportional to the amount of affinity detection reagent that bound to capture detection detected is proportional to the amount of affinity detection detection detection detected is proportional to the amount	
Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to capture	
samples are pipetted into the wells, the target protein in the standards and samples bin the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers an master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to captu	
the immobilized antibody. The wells are washed and the detection affinity reagent is ad the wells where it binds to any captured antigen. The wells are washed, and primers and master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to captu	ds and
the wells where it binds to any captured antigen. The wells are washed, and primers and master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to captu	oinds to
master mix are added to each well. The plate is placed into a real time PCR instrument cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to captu	added to
cycling and measurement of DNA amplification. The cycle number where amplification detected is proportional to the amount of affinity detection reagent that bound to captu	and PCR
detected is proportional to the amount of affinity detection reagent that bound to captu	nt for
	on is
	otured
Components: • p27/Kip1 Microplate (Item A): 96 well PCR plate coated with anti-Mouse p27/Kip1	

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/3 | Product datasheet for ABIN6385358 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

	 Wash Buffer I Concentrate (20x) (Item B): 25 ml of 20x concentrated solution Standards (Item C): 2 vials of recombinant Mouse p27/Kip1 Assay Diluent A (Item D): 30 ml diluent buffer, 0.09% sodium azide as preservative. For Standard/Sample (serum/plasma) diluent Assay Diluent B (Item E): 15 ml of 5x concentrated buffer. For Standard/Sample (cell culture medium/urine) diluent Detection Affinity Reagent for p27/Kip1 (Item F): 2 vials of a 4x concentrated solution of anti- Mouse p27/Kip1 affinity reagent IQELISA Detection Reagent (Item G): 1.4ml of a 10x concentrated stock Primer Solution (Item I): 1.7 ml vial PCR Master Mix (Item J): 1.2 ml vial
	 PCR Preparation buffer (Item K): 1ml vial of 10x concentrated buffer
	 Final Wash Buffer (Item L): 10ml vial of 10x concentrated buffer
Material not included:	Real-time PCR instrument, Bio-Rad recommended
	 Precision pipettes to deliver 2 µL to 1 mL volumes
	Adjustable 1-25 mL pipettes for reagent preparation
	100 mL and 1 liter graduated cylinders
	Absorbent paper
	Distilled or deionized water
	Log-log graph paper or computer and software for data analysis
	Tubes to prepare standard or sample dilutions

• Heating block or water bath capable of 80°C

Target Details

Target:	CDKN1B
Alternative Name:	p27/Kip1 (CDKN1B Products)
Gene ID:	821314
UniProt:	Q43735
Pathways:	Cell Division Cycle, Fc-epsilon Receptor Signaling Pathway, EGFR Signaling Pathway, Neurotrophin Signaling Pathway, Positive Regulation of Peptide Hormone Secretion, Negative Regulation of Hormone Secretion, Sensory Perception of Sound, Mitotic G1-G1/S Phases, DNA Replication, Positive Regulation of Endopeptidase Activity, Synthesis of DNA, Autophagy

Application Details

Application Notes:	The Immuno-Quantitative ELISA (IQELISA) kits are an innovative assay platform that combines
	the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. This results in

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/3 | Product datasheet for ABIN6385358 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

Application Details	
	an assay that is simultaneously familiar and cutting edge and enables the use of only 1/10th the sample volume while also providing 10x more sensitivity than a traditional ELISA.
Comment:	The Immuno-Quantitative ELISA (IQELISA [™]) kits are an innovative assay platform that combines the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. Also called immuno-PCR, this detection platform results in an assay that is simultaneously familiar and cutting edge. Compared to traditional ELISA, IQELISA [™] enables the use of only 1/10th the sample volume while also providing 10x more sensitivity.
Sample Volume:	25 µL
Plate:	Pre-coated
Protocol:	 Prepare all reagents, samples and standards as instructed Add 25 μL standard or sample to each well. Incubate for 2.5 hours at room temperature or overnight at 4°C Add 25 μL detection affinity reagent to each well. Incubate 1 hour at room temperature Add 25μL of IQELISA Detection Reagent to each well. Incubate 1 hour Add 15μL Primer solution 10μL of PCR master mix to each well Run real-time PCR
Restrictions:	For Research Use only
Handling	
Storage:	4 °C,-20 °C,-80 °C
Storage Comment:	May be stored for up to 6 months at 2° to 8°C from the date of shipment. Standard (recombinant protein) should be stored at -20°C or -80°C (recommended at -80°C) after reconstitution. Opened PCR plate or reagents may be stored for up to 1 month at 2° to 8°C. Note: the kit can be used within one year if the whole kit is stored at -20°C. Avoid repeated freeze-thaw cycles.
Expiry Date:	6 months