antibodies -online.com





Datasheet for ABIN6385541

DDX11 IQ-ELISA Kit

Go to Product page

()	11/0	K\ /	iew
	\cup	'I V/I	$I \cap VV$

Overview		
Quantity:	96 tests	
Target:	DDX11	
Reactivity:	Human	
Method Type:	Sandwich ELISA	
Application:	ELISA	
Product Details		
Purpose:	Human Immunoquantitative (PCR-Based) CHL-1 ELISA Kit for cell culture supernatants,	
	plasma, and serum samples.	
Sample Type:	Cell Culture Supernatant, Plasma, Serum	
Analytical Method:	Semi-Quantitative	
Detection Method:	qPCR	
Characteristics:	IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to	
	the immobilized antibody. The wells are washed and the detection affinity reagent is added to	
	the wells where it binds to any captured antigen. The wells are washed, and primers and PCR	
	master mix are added to each well. The plate is placed into a real time PCR instrument for	
	cycling and measurement of DNA amplification. The cycle number where amplification is	
	detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	
Components:	CHL-1 Microplate (Item A): 96 well PCR plate coated with anti-Human CHL-1	

- Wash Buffer I Concentrate (20x) (Item B): 25 ml of 20x concentrated solution
- Standards (Item C): 2 vials of recombinant Human CHL-1
- Assay Diluent A (Item D): 30 ml diluent buffer, 0.09% sodium azide as preservative. For Standard/Sample (serum/plasma) diluent
- Assay Diluent B (Item E): 15 ml of 5x concentrated buffer. For Standard/Sample (cell culture medium/urine) diluent
- Detection Affinity Reagent for CHL-1 (Item F): 2 vials of a 4x concentrated solution of anti-Human CHL-1 affinity reagent
- IQELISA Detection Reagent (Item G): 1.4ml of a 10x concentrated stock
- Primer Solution (Item I): 1.7 ml vial
- PCR Master Mix (Item J): 1.2 ml vial
- PCR Preparation buffer (Item K): 1ml vial of 10x concentrated buffer
- · Final Wash Buffer (Item L): 10ml vial of 10x concentrated buffer

Material not included:

- · Real-time PCR instrument, Bio-Rad recommended
- Precision pipettes to deliver 2 µL to 1 mL volumes
- · Adjustable 1-25 mL pipettes for reagent preparation
- · 100 mL and 1 liter graduated cylinders
- · Absorbent paper
- · Distilled or deionized water
- · Log-log graph paper or computer and software for data analysis
- · Tubes to prepare standard or sample dilutions
- Heating block or water bath capable of 80°C

Target Details

Target:	DDX11
Alternative Name:	CHL-1 (DDX11 Products)
Gene ID:	10752
UniProt:	000533
Pathways:	ER-Nucleus Signaling

The Immuno-Quantitative ELISA (IQELISA) kits are an innovative assay platform that combines
the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. This results in
an assay that is simultaneously familiar and cutting edge and enables the use of only 1/10th
the sample volume while also providing 10x more sensitivity than a traditional ELISA.
The Immuno-Quantitative ELISA (IQELISA™) kits are an innovative assay platform that

Application Details

	combines the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. Also	
	called immuno-PCR, this detection platform results in an assay that is simultaneously familiar	
	and cutting edge. Compared to traditional ELISA, IQELISA™ enables the use of only 1/10th the	
	sample volume while also providing 10x more sensitivity.	
Sample Volume:	25 μL	
Plate:	Pre-coated	
Protocol:	1. Prepare all reagents, samples and standards as instructed	
	2. Add 25 μL standard or sample to each well. Incubate for 2.5 hours at room temperature or overnight at 4 $^{\circ} C$	
	3. Add 25 µL detection affinity reagent to each well. Incubate 1 hour at room temperature	
	4. Add 25µL of IQELISA Detection Reagent to each well. Incubate 1 hour	
	5. Add 15μL Primer solution 10μL of PCR master mix to each well	
	6. Run real-time PCR	
Restrictions:	For Research Use only	
Handling		
Storage:	4 °C,-20 °C,-80 °C	
Storage Comment:	May be stored for up to 6 months at 2° to 8°C from the date of shipment. Standard	
	(recombinant protein) should be stored at -20°C or -80°C (recommended at -80°C) after	
	reconstitution. Opened PCR plate or reagents may be stored for up to 1 month at 2° to 8°C.	
	Note: the kit can be used within one year if the whole kit is stored at -20°C. Avoid repeated	
	freeze-thaw cycles.	
- · · · · ·		
Expiry Date:	6 months	