antibodies -online.com







Transferrin Receptor IQ-ELISA Kit



Overview

| 96 tests |
|---|
| Transferrin Receptor (TFRC) |
| Human |
| Sandwich ELISA |
| ELISA |
| |
| Human Immunoquantitative (PCR-Based) TfR (Transferrin R/CD71) ELISA Kit for cell culture |
| supernatants, plasma, and serum samples. |
| Cell Culture Supernatant, Plasma, Serum |
| Semi-Quantitative |
| qPCR |
| 0.081 ng/mL |
| IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and |
| samples are pipetted into the wells, the target protein in the standards and samples binds to |
| the immobilized antibody. The wells are washed and the detection affinity reagent is added to |
| the wells where it binds to any captured antigen. The wells are washed, and primers and PCR |
| master mix are added to each well. The plate is placed into a real time PCR instrument for |
| cycling and measurement of DNA amplification. The cycle number where amplification is |
| detected is proportional to the amount of affinity detection reagent that bound to captured |
| |

antigen in each well.

Product Details

Components:

- TfR (Transferrin R/CD71) Microplate (Item A): 96 well PCR plate coated with anti-Human TfR (Transferrin R/CD71)
- Wash Buffer I Concentrate (20x) (Item B): 25 ml of 20x concentrated solution
- Standards (Item C): 2 vials of recombinant Human TfR (Transferrin R/CD71)
- Assay Diluent A (Item D): 30 ml diluent buffer, 0.09% sodium azide as preservative. For Standard/Sample (serum/plasma) diluent
- Assay Diluent B (Item E): 15 ml of 5x concentrated buffer. For Standard/Sample (cell culture medium/urine) diluent
- Detection Affinity Reagent for TfR (Transferrin R/CD71) (Item F): 2 vials of a 4x concentrated solution of anti-Human TfR (Transferrin R/CD71) affinity reagent
- IQELISA Detection Reagent (Item G): 1.4ml of a 10x concentrated stock
- Primer Solution (Item I): 1.7 ml vial
- PCR Master Mix (Item J): 1.2 ml vial
- PCR Preparation buffer (Item K): 1ml vial of 10x concentrated buffer
- Final Wash Buffer (Item L): 10ml vial of 10x concentrated buffer

Material not included:

- · Real-time PCR instrument, Bio-Rad recommended
- Precision pipettes to deliver 2 µL to 1 mL volumes
- · Adjustable 1-25 mL pipettes for reagent preparation
- · 100 mL and 1 liter graduated cylinders
- Absorbent paper
- · Distilled or deionized water
- · Log-log graph paper or computer and software for data analysis
- · Tubes to prepare standard or sample dilutions
- Heating block or water bath capable of 80°C

Target Details

| Target: | Transferrin Receptor (TFRC) |
|-------------------|--|
| Alternative Name: | TfR (Transferrin R/CD71) (TFRC Products) |
| Gene ID: | 7037 |
| UniProt: | P02786 |
| Pathways: | Transition Metal Ion Homeostasis |

Application Details

Application Notes:

The Immuno-Quantitative ELISA (IQELISA) kits are an innovative assay platform that combines the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. This results in an assay that is simultaneously familiar and cutting edge and enables the use of only 1/10th the sample volume while also providing 10x more sensitivity than a traditional ELISA.

Application Details

| Comment: | The Immuno-Quantitative ELISA (IQELISA™) kits are an innovative assay platform that |
|------------------|--|
| | combines the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. Also |
| | called immuno-PCR, this detection platform results in an assay that is simultaneously familiar |
| | and cutting edge. Compared to traditional ELISA, IQELISA™ enables the use of only 1/10th the |
| | sample volume while also providing 10x more sensitivity. |
| Sample Volume: | 25 μL |
| Plate: | Pre-coated |
| Protocol: | 1. Prepare all reagents, samples and standards as instructed |
| | 2. Add 25 μ L standard or sample to each well. Incubate for 2.5 hours at room temperature or overnight at 4 $^{\circ}$ C |
| | 3. Add 25 µL detection affinity reagent to each well. Incubate 1 hour at room temperature |
| | 4. Add 25µL of IQELISA Detection Reagent to each well. Incubate 1 hour |
| | 5. Add 15μL Primer solution 10μL of PCR master mix to each well |
| | 6. Run real-time PCR |
| Restrictions: | For Research Use only |
| Handling | |
| Storage: | 4 °C,-20 °C,-80 °C |
| Storage Comment: | May be stored for up to 6 months at 2° to 8°C from the date of shipment. Standard |
| | (recombinant protein) should be stored at -20°C or -80°C (recommended at -80°C) after |
| | reconstitution. Opened PCR plate or reagents may be stored for up to 1 month at 2° to 8°C. |
| | Note: the kit can be used within one year if the whole kit is stored at -20°C. Avoid repeated |
| | freeze-thaw cycles. |
| Expiry Date: | 6 months |
| | |