.-online.com antibodies

Datasheet for ABIN6574088 Corticosterone ELISA Kit

3 Images

25 Publications



Overview

Quantity:	96 tests
Target:	Corticosterone (CORT)
Reactivity:	Various Species
Method Type:	Competition ELISA
Detection Range:	1.23 ng/mL - 100 ng/mL
Minimum Detection Limit:	1.23 ng/mL
Application:	ELISA
Product Details	
Purpose:	The kit is a competitive inhibition enzyme immunoassay technique for the in vitro quantitative
	measurement of corticosterone in serum, plasma, urine.
	We offer validation data (WB) for the kit components. So you can be sure to order a reliable
	ELISA kit product composed of high quality reagents.
Sample Type:	Plasma, Serum, Urine
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Corticosterone (Cort)
Cross-Reactivity (Details):	No significant cross-reactivity or interference between Corticosterone (Cort) and analogues was observed.
Sensitivity:	0.51 ng/mL

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/6 | Product datasheet for ABIN6574088 | 11/30/2023 | Copyright antibodies-online. All rights reserved.

Product Details

Components:

- Pre-coated, ready to use 96-well strip plate, flat buttom
- Plate sealer for 96 wells
- Reference Standard
- Standard Diluent
- Detection Reagent A
- Detection Reagent B
- Assay Diluent A
- Assay Diluent B
- Reagent Diluent (if Detection Reagent is lyophilized)
- TMB Substrate
- Stop Solution
- Wash Buffer (30 x concentrate)
- Instruction manual

Target Details

Target:	Corticosterone (CORT)
Abstract:	CORT Products
Target Type:	Hormone

Application Details

Comment:	Information on standard material:
	The standard might be recombinant protein or natural protein, that will depend on the specific
	kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin
	300 in the standard as preservative.
	Information on reagents:
	The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash
	solution is TBS. The standard diluent contains 0.02 $\%$ sodium azide, assay diluent A and assay
	diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.
	Information on antibodies:
	The provided antibodies and their host vary in different kits.
Sample Volume:	50 μL
Assay Time:	2 h
Plate:	Pre-coated

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/6 | Product datasheet for ABIN6574088 | 11/30/2023 | Copyright antibodies-online. All rights reserved.

Application Details

Protocol:	1. Prepare all reagents, samples and standards,
	2. Add 50µL standard or sample to each well.
	Then add 50µL prepared Detection Reagent A immediately.
	Shake and mix. Incubate 1 hour at 37 °C,
	3. Aspirate and wash 3 times,
	4. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,
	5. Aspirate and wash 5 times,
	6. Add 90μL Substrate Solution. Incubate 10-20 minutes at 37 °C, 7. Add 50μL Stop Solution. Read at 450 nm immediately.
Reagent Preparation:	1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit
	will not be used up in one time, please only take out strips and reagents for present
	experiment, and leave the remaining strips and reagents in required condition.
	2. Standard - Reconstitute the Standard with 0.5 mL of Standard Diluent, kept for 10 minutes at
	room temperature, shake gently(not to foam). The concentration of the standard in the stock
	solution is 100 ng/mL. Please prepare 5 tubes containing 0.6 mL Standard Diluent and
	produce a triple dilution series according to the picture shown below. Mix each tube
	thoroughly before the next transfer. Set up 5 points of diluted standard such as 100 ng/mL,
	33.33 ng/mL, 11.11 ng/mL, 3.70 ng/mL, 1.23 ng/mL, and the last EP tubes with Standard
	Diluent is the blank as 0 ng/mL.
	3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection
	Reagent A with 150µL of Reagent Diluent, kept for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before
	use. Dilute them to the working concentration 100-fold with Assay Diluent A and B,
	respectively.
	4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized
	or distilled water to prepare 600 mL of Wash Solution (1x).
	5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not
	dump the residual solution into the vial again.
	Note:
	1. Making serial dilution in the wells directly is not permitted.
	 Prepare standard within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.
	3. Detection Reagent A and B are sticky solutions, therefore, slowly pipette them to reduce the volume errors.
	4. Please carefully reconstitute Standards or working Detection Reagent A and B according to
	the instruction, and avoid foaming and mix gently until the crystals are completely dissolved.
	To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors
	are calibrated. It is recommended to suck more than 10μ L for one pipetting.
	5. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only
	once.
	6. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature
	and mix gently until the crystals are completely dissolved.

	7. Contaminated water or container for reagent preparation will influence the detection result.
Sample Preparation:	 It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 - 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates. If the sample type is not specified in the instructions, a preliminary test is necessary to determinecompatibility with the kit. If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a possibility of causing a deviation due to the introduced chemical substance. The recommended dilution factor is for reference only. Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2).
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level of
	target were tested 20 times on one plate, respectively.
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level of
	target were tested on 3 different plates, 8 replicates in each plate.
	CV(%) = SD/meanX100
	Intra-Assay: CV < 10%
	Inter-Assay: CV < 12%
Restrictions:	For Research Use only
Handling	
Precaution of Use:	The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and
	clothing protection when using this material.
Storage:	4 °C/-20 °C
Storage Comment:	 For unopened kit: All reagents should be stored according to the labels on the vials. The Standard, Detection Reagent A, Detection Reagent B, and 96-well Strip Plate should be stored at -20 °C upon receipt, while the other reagents should be stored at 4 °C. For opened kits: the remaining reagents must be stored according to the above storage conditions. In addition, please return the unused wells to the foil pouch containing the desiccant and seal the foil pouch with the zipper.
Expiny Data:	6 months
Expiry Date:	o montens

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/6 | Product datasheet for ABIN6574088 | 11/30/2023 | Copyright antibodies-online. All rights reserved. Product cited in:

Dong, Han, Wang, Qin, Yang, Cao, Chen: "Role of serotonin on the intestinal mucosal immune response to stress-induced diarrhea in weaning mice." in: **BMC gastroenterology**, Vol. 17, Issue 1, pp. 82, (2018) (PubMed).

Lu, Liu, Wang, Sun, Yang, Xing, Sun, Wang, Zhao: "The Role of Psychological Stress on Heart Autophagy in Mice With Heart Failure." in: **Psychosomatic medicine**, Vol. 79, Issue 9, pp. 1036-1044, (2018) (PubMed).


Dong, Wang, Qin, Cao, Chen: "Role of serotonin in the intestinal mucosal epithelium barrier in weaning mice undergoing stress-induced diarrhea." in: **Journal of molecular histology**, Vol. 49, Issue 1, pp. 85-97, (2018) (PubMed).

Firoozmand, Sanches, Damaceno-Rodrigues, Perez, Aragão, Rosa, Marcondes, Casarini, Caldini, Cunha: "Blockade of AT1 type receptors for angiotensin II prevents cardiac microvascular fibrosis induced by chronic stress in Sprague-Dawley rats." in: **Stress (Amsterdam, Netherlands)**, pp. 1-10, (2018) (PubMed).

Ferreira, Armstrong-Jr, Kudo, de Jesus Correia, Dos Reis, Sannomiya, Breithaupt-Faloppa, Moreira: "Differential Effects of Brain Death on Rat Microcirculation and Intestinal Inflammation: Female Versus Male." in: **Inflammation**, Vol. 41, Issue 4, pp. 1488-1497, (2018) (PubMed).

There are more publications referencing this product on: Product page


Images



SDS-PAGE

Image 1. SDS-PAGE of Protein Standard from the Kit (BSA-Cort).

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 5/6 | Product datasheet for ABIN6574088 | 11/30/2023 | Copyright antibodies-online. All rights reserved.



ELISA

Image 2. Rabbit Capture antibody from the kit in ELISA with

Positive Control: Human, rat, mouse serum.

ELISA

Image 3. Typical standard curve

