# .-online.com antibodies

## Datasheet for ABIN6574128 IFNA ELISA Kit

4 Images

2 Publications



#### Overview

Quantity:	96 tests
Target:	IFNA
Reactivity:	Rat
Method Type:	Sandwich ELISA
Detection Range:	15.6 pg/mL - 1000 pg/mL
Minimum Detection Limit:	15.6 pg/mL
Application:	ELISA
Product Details	
Purpose:	The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of IFNa in rat
	serum, plasma, tissue homogenates, cell lysates, cell culture supernates.
	We offer <b>validation data</b> (WB) <b>for the kit components</b> . So you can be sure to order a reliable ELISA kit product composed of high quality reagents.
Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Interferon Alpha (IFNa)
Cross-Reactivity (Details):	No significant cross-reactivity or interference between Interferon Alpha (IFNa) and analogues was observed.
Sensitivity:	6.2 pg/mL

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/6 | Product datasheet for ABIN6574128 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

## Product Details

#### Components:

- Pre-coated, ready to use 96-well strip plate, flat buttom
- Plate sealer for 96 wells
- Reference Standard
- Standard Diluent
- Detection Reagent A
- Detection Reagent B
- Assay Diluent A
- Assay Diluent B
- Reagent Diluent (if Detection Reagent is lyophilized)
- TMB Substrate
- Stop Solution
- Wash Buffer (30 x concentrate)
- Instruction manual

## Target Details

Target:	IFNA
Abstract:	IFNA Products
UniProt:	P05011
Pathways:	JAK-STAT Signaling, TLR Signaling, Hepatitis C, Inflammasome

## Application Details

Comment:	Information on standard material:
	The standard might be recombinant protein or natural protein, that will depend on the specific
	kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin
	300 in the standard as preservative.
	Information on reagents:
	The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash
	solution is TBS. The standard diluent contains 0.02 % sodium azide, assay diluent A and assay
	diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.
	Information on antibodies:
	The provided antibodies and their host vary in different kits.
Sample Volume:	100 µL
Assay Time:	3 h

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/6 | Product datasheet for ABIN6574128 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

## Application Details

Protocol:       1. Prepare all reagents, samples and standards,         2. Add 100µL standard or sample to each well. Incubate 1 hours at 37 °C,         3. Aspirate and add 100µL prepared Detection Reagent A. Incubate 1 hours at 37 °C,         4. Aspirate and wash 3 times,         5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,         8. Add 50µL Stop Solution. Need at 450mm immediately.         Reagent Preparation:       1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents in required condition.         2. Standard - Reconstitute the Standard with 10 mL of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard is such as such as such highest standard to produce a double dilution sense. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1.000pg/mL, 500pg/mL, 250pg/mL, 25.5pg/mL, 13.2pg/mL, 15.6pg/mL, and the last microcentring tube. Thoroughly before the next transfer. Set up 7 points of diluted standard such as 1.000pg/mL, 300pg/mL, 250pg/mL, 25.5pg/mL, 3.12pg/mL, 16.6pg/mL, and the last microcentring tube. Thoroughly before the next transfer. Set up 7 points of diluted standard such as 1.000pg/mL, and the last microcentring to centrifuge the taskok bolation of and betection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Dilemt A and B, respectively.         3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the bated or distiled water to prepare 600 mL of W	Plate:	Pre-coated				
<ul> <li>3. Aspirate and add 100µL prepared Detection Reagent A. Incubate 1 hour at 37 °C,</li> <li>4. Aspirate and weah 3 times,</li> <li>5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,</li> <li>6. Aspirate and weah 5 times,</li> <li>7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,</li> <li>8. Add 50µL Stop Solution. Read at 450mm immediately.</li> </ul> Reagent Preparation: <ol> <li>1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents in required condition.</li> <li>2. Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 2.000pg/mL, Firstly dilute the stock solution to 1.000pg/mL and the diluted standard such as 1.000pg/mL, 500pg/mL, 125pg/mL, 125pg/mL, 31.2pg/mL, 13.5pg/mL, and the flast thoroughly before the next transfer. Set up 7 points of diluted standard such as 1.000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the flast microcentriting tube with Standard Diluent is the blank as 0.gg/mL.</li> <li>3. Detection Reagent A and Detection Reagent 10 fly ophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, Keep for 10 minutes at room temperature, shake gently (not to foam). Birefly spin or centrifuge the stock Detection A and Detection Reagent P in Mustes at S000 with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1s).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before</li></ol>	Protocol:	1. Prepare all reagents, samples and standards,				
<ul> <li>4. Aspirate and wash 3 times,</li> <li>5. Add 100,L prepared Detection Reagent B. Incubate 30 minutes at 37 °C,</li> <li>6. Aspirate and wash 5 times,</li> <li>7. Add 30,L Substrate Solution. Incubate 10-20 minutes at 37 °C,</li> <li>8. Add 50µL Stop Solution. Read at 450nm immediately.</li> </ul> Reagent Preparation: <ol> <li>1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents in required condition.</li> <li>2. Standard - Reconstitute the Standard with 1.0 mL of Standard Dileuch, teep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution to 1,000pg/mL and the diluted standard serves as the highest standard (1,000pg/mL). Then prepare 7 tubes containing</li> <li>0.5 mL Standard Dileunt and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15 6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150,L of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the solution with Assay Diluent A and B, respective).</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 3</li></ol>						
5       Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,       6. Aspirate and wash 5 times,         7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,       8. Add 50µL Stop Solution. Read at 450nm immediately.         Reagent Preparation:       1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents for present experiment, and leave the temaining strips and reagents for present experiment, and leave the temaining strips and reagents for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 2.0000pt/mL. Firstly dilute the stock solution to 1.0000pt/mL and the diluted standard serves as the highest standard (1.0000pt/mL). Then prepare 7 tubes containing 0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1.0000pt/mL, 500pt/mL, 250pt/mL, 7.25pt/mL, 7.25pt/mL, 7.12pt/mL, 16.6pt/mL, and the last microcentrifuge tube with Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at and Detection Before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.         4. Wash Solution - Dilute 20 mL of Wash Solution (1x).       5. TMB substrate - Aspirate + Aspirate + Aspirate		3. Aspirate and add 100 $\mu$ L prepared Detection Reagent A. Incubate 1 hour at 37 °C,				
<ul> <li>6. Aspirate and wash 5 times.</li> <li>7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,</li> <li>8. Add 50µL Stop Solution. Read at 450nm immediately.</li> </ul> Reagent Preparation: <ol> <li>1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents in required condition.</li> <li>2. Standard - Reconstitute the Standard with 10. Ph. of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 2.000g/mL. Firstly dilute the stock solution to 1.000g/mL and the diluted standard serves as the highest standard (1.0000g/mL). Then prepare 7 tubes containing 0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1.0000g/mL, S00g/mL, 250g/mL, 1250g/mL, 25.0g/mL, 31.2g/mL, 15.6g/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 05g/mL.</li> <li>8. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at noom temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection Reagent A with a Standard Diluent, keep for 10 minutes at noom temperature, shake gently (not to foam). Briefly spin or centrifuge the slote of an AD etection Before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively. <ol> <li>Wash Solution - Dilute 20 mL of Wash Solution (tw).</li> <li>TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the rea</li></ol></li></ol>		4. Aspirate and wash 3 times,				
<ul> <li>2. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C, 8. Add 50µL Stop Solution. Read at 450nm immediately.</li> <li>Reagent Preparation:         <ol> <li>Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents in required condition.</li> <li>Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution to 1.000pg/mL and the diluted standard serves as the highest standard (1.000pg/mL). Then prepare 7 tubes containing 0.5 mL. Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1.000pg/mL, S00gg/mL, 250gg/mL, 1255g/mL, 61.5 gpg/mL, 13.0 gg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 00gg/mL.</li> <li>Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection Before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>Wash Solution - Dilute 20 mL of Wash Solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li></ol></li></ul>		5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,				
<ul> <li>8. Add 50µL Stop Solution. Read at 450nm immediately.</li> <li>Reagent Preparation:         <ul> <li>1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents in required condition.</li> <li>2. Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 2,000pg/mL. Firstly dilute the stock solution to 1,000pg/mL and the diluted standard serves as the highest standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 250pg/mL, 250pg/mL, 250pg/mL, 32.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL, 36.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection Before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> </ul> </li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:         <ul> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reag</li></ul></li></ul>		6. Aspirate and wash 5 times,				
<ol> <li>Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents in required condition.</li> <li>Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 2,000pg/mL. Firstly dilute the stock solution to 1,000pg/mL and the diluted standard serves as the highest standard (10,000g/mL). Then prepare 7 tubes containing 0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 125pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 00pg/mL.</li> <li>Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection Before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilied water to prepare 600 mL of Wash Solution (11x).</li> <li>TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not durp the residual solution into the vial again.</li> <li>Note:         <ul> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and a</li></ul></li></ol>		7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,				
<ul> <li>will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents in required condition.</li> <li>Standard -Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 2,000pg/mL. Firstly dilute the stock solution to 1,000pg/mL and the diluted standard serves as the highest standard (1,000pg/mL). Then prepare 7 tubes containing 0.5 mL. Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix genty until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipet</li></ul>		8. Add 50µL Stop Solution. Read at 450nm immediately.				
<ul> <li>experiment, and leave the remaining strips and reagents in required condition.</li> <li>2. Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 2,000pg/mL. Firstly dilute the stock solution to 1,000pg/mL and the diluted standard serves as the highest standard (1,000pg/mL). Then prepare 7 tubes containing 0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection Before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> </ul>	Reagent Preparation:	1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit				
<ol> <li>Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 2.000pg/mL. Firstly dilute the stock solution to 1.000pg/mL and the diluted standard serves as the highest standard (1.000pg/mL). Then prepare 7 tubes containing 0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1.000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 125pg/mL, 31.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as Opg/mL.</li> <li>Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>Wash Solution - Dilute 20 mL of Wash Solution (1x).</li> <li>TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> </ol>		will not be used up in one time, please only take out strips and reagents for present				
<ul> <li>room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 2,000pg/mL. Firstly dilute the stock solution to 1,000pg/mL and the diluted standard serves as the highest standard (1,000pg/mL). Then prepare 7 tubes containing 0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 1250pg/mL, 1250g/mL, 25.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipetors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and B according to the instruction, and avoid foaming and mix gently untit be crystals are completely dissolv</li></ul>		experiment, and leave the remaining strips and reagents in required condition.				
<ul> <li>solution is 2,000pg/mL. Firstly dilute the stock solution to 1,000pg/mL and the diluted standard serves as the highest standard (1,000pg/mL). Then prepare 7 tubes containing 0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 250pg/mL, 250g/mL, 252gg/mL, 21.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>		2. Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at				
<ul> <li>standard serves as the highest standard (1,000pg/mL). Then prepare 7 tubes containing 0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>		room temperature, shake gently (not to foam). The concentration of the standard in the stock				
<ul> <li>0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and B according to once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>		solution is 2,000pg/mL. Firstly dilute the stock solution to 1,000pg/mL and the diluted				
<ul> <li>each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstitute Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>		standard serves as the highest standard (1,000pg/mL). Then prepare 7 tubes containing				
<ul> <li>each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>		0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix				
<ul> <li>1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 31.2pg/mL, 15.5pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstitute Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>						
<ul> <li>last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>						
<ul> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>						
<ul> <li>Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note: <ol> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ol> </li> </ul>						
<ul> <li>gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note:</li> <li>1. Making serial dilution in the wells directly is not permitted.</li> <li>2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>						
<ul> <li>use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note: <ol> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> </ol> </li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>						
<ul> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note: <ol> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> </ol> </li> </ul>		use. Dilute them to the working concentration 100-fold with Assay Diluent A and B,				
<ul> <li>or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note: <ol> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> </ol> </li> <li>If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>						
<ul> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> <li>Note: <ol> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ol> </li> </ul>						
<ul> <li>dump the residual solution into the vial again.</li> <li>Note: <ol> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ol> </li> </ul>						
<ol> <li>Note:</li> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ol>						
<ol> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ol>						
<ol> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> <li>Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ol>						
<ul> <li>37 °C directly.</li> <li>3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>						
<ul> <li>the instruction, and avoid foaming and mix gently until the crystals are completely dissolved.</li> <li>To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>						
<ul> <li>To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>		3. Please carefully reconstitute Standards or working Detection Reagent A and B according to				
<ul> <li>are calibrated. It is recommended to suck more than 10µL for one pipetting.</li> <li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li> <li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li> </ul>		the instruction, and avoid foaming and mix gently until the crystals are completely dissolved.				
<ul><li>4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.</li><li>5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature</li></ul>		To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors				
once. 5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature		are calibrated. It is recommended to suck more than $10\mu L$ for one pipetting.				
5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature						
		and mix gently until the crystals are completely dissolved.				

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 3/6 | Product datasheet for ABIN6574128 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

	6. Contaminated water or container for reagent preparation will influence the detection result.					
Sample Preparation:	<ul> <li>It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 - 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates.</li> <li>If the sample type is not specified in the instructions, a preliminary test is necessary to determinecompatibility with the kit.</li> <li>If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a possibility of causing a deviation due to the introduced chemical substance. The recommended dilution factor is for reference only.</li> <li>Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2).</li> </ul>					
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level of					
	target were tested 20 times on one plate, respectively.					
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level of					
	target were tested on 3 different plates, 8 replicates in each plate.					
	CV(%) = SD/meanX100					
	Intra-Assay: CV < 10%					
	Inter-Assay: CV < 12%					
Restrictions:	For Research Use only					
Handling						
Precaution of Use:	The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and					
	clothing protection when using this material.					
Storage:	4 °C/-20 °C					
Storage Comment:	<ol> <li>For unopened kit: All reagents should be stored according to the labels on the vials. The Standard, Detection Reagent A, Detection Reagent B, and 96-well Strip Plate should be stored at -20 °C upon receipt, while the other reagents should be stored at 4 °C.</li> <li>For opened kits: the remaining reagents must be stored according to the above storage conditions. In addition, please return the unused wells to the foil pouch containing the desiccant and seal the foil pouch with the zipper.</li> </ol>					
с. : р.:						
Expiry Date:	6 months					

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/6 | Product datasheet for ABIN6574128 | 09/10/2023 | Copyright antibodies-online. All rights reserved. Product cited in:


Li, Li, Wang, Zhang, Song, Zhang, Gao, Liao, He, You, Tan, Luo, Li, Tang, Weng, Yi, Peng, Liu, Tan, Bode, Cao: "IDH2 is a novel diagnostic and prognostic serum biomarker for non-small-cell lung cancer." in: **Molecular oncology**, Vol. 12, Issue 5, pp. 602-610, (2019) (PubMed).

Ren, Yang, Wang, Ma: "The effects of PGC-1α on the proliferation and energy metabolism of malignant endometrial cancer cells." in: **OncoTargets and therapy**, Vol. 8, pp. 769-74, (2015) ( PubMed).

Sun, Chen, Tan, Zhang, Yao, Zhou, Li, Gao, Liu, Tan, Zhou, He, Shao, Li, Qiu, Sun, Yu, Wang, Zhao, Shi, He: "Isocitrate dehydrogenase 1 is a novel plasma biomarker for the diagnosis of non-small cell lung cancer." in: **Clinical cancer research : an official journal of the American Association for Cancer Research**, Vol. 19, Issue 18, pp. 5136-45, (2013) (PubMed).

Tan, Jiang, Sun, Chen, Lv, Shao, Li, Qiu, Gao, Li, Tan, Zhou, Wang, Ding, Wang, Sun, Hang, Shi, Feng, He, He: "Identification of isocitrate dehydrogenase 1 as a potential diagnostic and prognostic biomarker for non-small cell lung cancer by proteomic analysis." in: **Molecular & cellular proteomics : MCP**, Vol. 11, Issue 2, pp. M111.008821, (2012) (PubMed).


#### Images




#### SDS-PAGE

**Image 1.** SDS-PAGE of Protein Standard from the Kit (Highly purified E. coli-expressed recombinant rat IFNa).

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 5/6 | Product datasheet for ABIN6574128 | 09/10/2023 | Copyright antibodies-online. All rights reserved.





#### ELISA

Image 2. Typical standard curve

Western Blotting								
Image 3.	WB	of	Protein	Standard:	different control			
antibodies	agai	inst	Highly	purified	E. coli-expressed			
recombinant rat IFNa.								

Please check the product details page for more images. Overall 4 images are available for ABIN6574128.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 6/6 | Product datasheet for ABIN6574128 | 09/10/2023 | Copyright antibodies-online. All rights reserved.