# .-online.com antibodies

## Datasheet for ABIN6574143 VEGFA ELISA Kit

5 Images

26 Publications



#### Overview

Quantity:	96 tests		
Target:	VEGFA		
Reactivity:	Human		
Method Type:	Sandwich ELISA		
Detection Range:	15.6 pg/mL - 1000 pg/mL		
Minimum Detection Limit:	15.6 pg/mL		
Application:	ELISA		
Product Details			
Purpose:	The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of VEGFA in		
	human serum, plasma, tissue homogenates, cell lysates, cell culture supernates.		
	We offer validation data (WB) for the kit components. So you can be sure to order a reliable		
	ELISA kit product composed of high quality reagents.		
Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate		
Analytical Method:	Quantitative		
Detection Method:	Colorimetric		
Specificity:	This assay has high sensitivity and excellent specificity for detection of Vascular Endothelial		
	Growth Factor A (VEGFA)		
Cross-Reactivity (Details):	No significant cross-reactivity or interference between Vascular Endothelial Growth Factor A		
	(VEGFA) and analogues was observed.		

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/6 | Product datasheet for ABIN6574143 | 01/16/2024 | Copyright antibodies-online. All rights reserved.

#### Product Details

Sensitivity:	6.2 pg/mL		
Components:	Pre-coated, ready to use 96-well strip plate, flat buttom		
	Plate sealer for 96 wells		
	Reference Standard		
	Standard Diluent		
	Detection Reagent A		
	Detection Reagent B		
	Assay Diluent A		
	Assay Diluent B		
	Reagent Diluent (if Detection Reagent is lyophilized)		
	TMB Substrate		
	Stop Solution		
	Wash Buffer (30 x concentrate)		
	Instruction manual		

## Target Details

Target:	VEGFA		
Abstract:	VEGFA Products		
UniProt:	P15692		
Pathways:	RTK Signaling, Glycosaminoglycan Metabolic Process, Regulation of Cell Size, Tube Formation, Signaling Events mediated by VEGFR1 and VEGFR2, Platelet-derived growth Factor Receptor		
	Signaling, VEGFR1 Specific Signals, VEGF Signaling		

## Application Details

Comment:	Information on standard material:		
	The standard might be recombinant protein or natural protein, that will depend on the specific		
	kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin		
	300 in the standard as preservative.		
	Information on reagents:		
	The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash		
	solution is TBS. The standard diluent contains 0.02 % sodium azide, assay diluent A and assay		
	diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.		

Information on antibodies:

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/6 | Product datasheet for ABIN6574143 | 01/16/2024 | Copyright antibodies-online. All rights reserved.

## Application Details


	The provided antibodies and their host vary in different kits.			
Sample Volume:	100 μL			
Assay Time:	3 h			
Plate:	Pre-coated			
Protocol:	1. Prepare all reagents, samples and standards, 2. Add 100µL standard or sample to each well. Incubate 1 hours at 37 °C,			
	<ol> <li>Aspirate and add 100µL prepared Detection Reagent A. Incubate 1 hour at 37 °C,</li> <li>Aspirate and wash 3 times,</li> </ol>			
	5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C, 6. Aspirate and wash 5 times,			
	7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,			
	8. Add 50µL Stop Solution. Read at 450nm immediately.			
Reagent Preparation:	1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit will not be used up in one time, please only take out strips and reagents for present experiment, and leave the remaining strips and reagents in required condition.			
	<ol> <li>Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes a room temperature, shake gently (not to foam). The concentration of the standard in the stoc solution is 1,000pg/mL. Prepare 7 tubes containing 0.5 mL Standard Diluent and produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.</li> </ol>			
	<ul> <li>3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.</li> <li>4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).</li> <li>5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.</li> </ul>			
	Note:			
	<ol> <li>Making serial dilution in the wells directly is not permitted.</li> <li>Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.</li> </ol>			
	3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10μL for one pipetting.			

	<ol> <li>The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used onl once.</li> </ol>			
	<ol> <li>If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature and mix gently until the crystals are completely dissolved.</li> </ol>			
	6. Contaminated water or container for reagent preparation will influence the detection result.			
Sample Preparation:	<ul> <li>It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 - 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates.</li> <li>If the sample type is not specified in the instructions, a preliminary test is necessary to determinecompatibility with the kit.</li> <li>If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a possibility of causing a deviation due to the introduced chemical substance. The recommended dilution factor is for reference only.</li> <li>Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2).</li> </ul>			
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level of			
	target were tested 20 times on one plate, respectively.			
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level of			
	target were tested on 3 different plates, 8 replicates in each plate.			
	CV(%) = SD/meanX100			
	Intra-Assay: CV < 10%			
	Inter-Assay: CV < 12%			
Restrictions:	For Research Use only			
Handling				
Precaution of Use:	The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and			
	clothing protection when using this material.			
Storage:	4 °C/-20 °C			
Storage Comment:	<ol> <li>For unopened kit: All reagents should be stored according to the labels on the vials Standard, Detection Reagent A, Detection Reagent B, and 96-well Strip Plate should at -20 °C upon receipt, while the other reagents should be stored at 4 °C.</li> <li>For opened kits: the remaining reagents must be stored according to the above stored</li> </ol>			

conditions. In addition, please return the unused wells to the foil pouch containing the


Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/6 | Product datasheet for ABIN6574143 | 01/16/2024 | Copyright antibodies-online. All rights reserved. desiccant and seal the foil pouch with the zipper.

onths Sun, Wu, Tan, Sun, Liu, Si, Xu, Huang, Zhou, Yang, Wang: "IL-17A promotes CR2-dependent angiogenesis in a mouse model of liver cancer." in: <b>Molecular medicine</b> <b>prts</b> , Vol. 20, Issue 2, pp. 1065-1074, (2019) (PubMed).
CR2-dependent angiogenesis in a mouse model of liver cancer." in: Molecular medicine
CR2-dependent angiogenesis in a mouse model of liver cancer." in: Molecular medicine
orts, Vol. 20, Issue 2, pp. 1065-1074, (2019) (PubMed).
Liu, Mu, Huang, Zhang, Zhao, Shu, Zhang, Hao, Li, Xue, Zhang, Zhang: "A DPP-IV-resistant
agon-like peptide-2 dimer with enhanced activity against radiation-induced intestinal injury.
ournal of controlled release : official journal of the Controlled Release Society, Vol. 260,
32-45, (2018) (PubMed).
, Martin, Tian, Cui, Wang, Wu, Gu: "Evaluation of Detection Methods and Values of
ulating Vascular Endothelial Growth Factor in Lung Cancer." in: Journal of Cancer, Vol. 9,
ie 7, pp. 1287-1300, (2018) (PubMed).
ng, Wang, Cheng, Li, Wang, Zhang, Qin: "Repulsive Guidance Molecule a Inhibits
iogenesis by Downregulating VEGF and Phosphorylated Focal Adhesion Kinase In Vitro." in:
ntiers in neurology, Vol. 8, pp. 504, (2017) (PubMed).
, Huang, Xu, Sun, Lv, Wang: "Advanced glycation end products promote VEGF expression
thus choroidal neovascularization via Cyr61-PI3K/AKT signaling pathway." in: Scientific
orts, Vol. 7, Issue 1, pp. 14925, (2017) (PubMed).
re are more publications referencing this product on: Product page
כ



#### ELISA

Image 1. Typical standard curve



#### Western Blotting

**Image 2.** Mouse Capture antibody from the kit in WB with Positive Control: Human lung tissue lysates.

Western Blotting				
Image 3.	WB of	Protein	Standard: different control	
antibodies	against	Highly	purified E. coli-expressed	
recombinant human VEGFA.				

Please check the product details page for more images. Overall 5 images are available for ABIN6574143.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 6/6 | Product datasheet for ABIN6574143 | 01/16/2024 | Copyright antibodies-online. All rights reserved.