antibodies - online.com







Adenylosuccinate Lyase ELISA Kit



Overview

Quantity:	96 tests		
Target:	Adenylosuccinate Lyase (ADSL)		
Reactivity:	Human		
Method Type:	Sandwich ELISA		
Application:	ELISA		

Product Details				
Purpose:	Human ADSL ELISA Kit.			
Sample Type:	Cell Culture Supernatant, Cell Samples, Plasma, Serum, Tissue Lysate			
Analytical Method:	Quantitative			
Detection Method:	Colorimetric			
Specificity:	This ELISA antibody pair recognizes Human ADSL.			
Characteristics:	 Strip plates and additional reagents allow for use in multiple experiments Quantitative protein detection Establishes normal range The best products for confirmation of antibody array data 			
Components:	 Pre-Coated 96-well Strip Microplate Wash Buffer Stop Solution Assay Diluent(s) Lyophilized Standard Biotinylated Detection Antibody 			

Product Details

- · Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- · Distilled or deionized water
- Precision pipettes to deliver 2 µl to 1 µl volumes
- Adjustable 1-25 µl pipettes for reagent preparation
- 100 µl and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:	Adenylosuccinate Lyase (ADSL)	
Alternative Name:	ADSL (ADSL Products)	
UniProt:	P30566	
Pathways:	Ribonucleoside Biosynthetic Process	

Application Details

Application Notes:	Optimal working dilution should be determined by the investigator.		
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.		
	2. Add 100 µl of standard or sample to each well.		
	3. Incubate 2.5 h at RT or O/N at 4°C.		
	4. Add 100 µl of prepared biotin antibody to each well.		
	5. Incubate 1 h at RT.		
	6. Add 100 µl of prepared Streptavidin solution to each well.		
	7. Incubate 45 min at RT.		
	8. Add 100 µl of TMB One-Step Substrate Reagent to each well.		
	9. Incubate 30 min at RT.		
	10. Add 50 µl of Stop Solution to each well.		
	11. Read at 450 nm immediately.		
Restrictions:	For Research Use only		
Handling			
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated		
	freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is		

			11.	
_	\sim	\sim	111	nc
_	14	111		111

recommended to store at -80°C.

Expiry Date: 6 months