# antibodies .- online.com





## Datasheet for ABIN6730402

### **CD73 ELISA Kit**



#### Overview

Quantity:	96 tests
Target:	CD73 (NT5E)
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	7 pg/mL - 1500 pg/mL
Minimum Detection Limit:	7 pg/mL
Application:	ELISA

#### **Product Details**

Mouse 5'-Nucleotidase/CD73 ELISA Kit for Serum, Plasma, and Cell Culture Supernatants.
Cell Culture Supernatant, Plasma, Serum
Quantitative
Colorimetric
This ELISA antibody pair recognizes Mouse 5'-Nucleotidase.
<ul> <li>Strip plates and additional reagents allow for use in multiple experiments</li> <li>Quantitative protein detection</li> <li>Establishes normal range</li> <li>The best products for confirmation of antibody array data</li> </ul>
<ul><li>Pre-Coated 96-well Strip Microplate</li><li>Wash Buffer</li><li>Stop Solution</li></ul>

#### **Product Details**

- Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

#### Material not included:

- · Distilled or deionized water
- Precision pipettes to deliver 2 μl to 1 μl volumes
- Adjustable 1-25 µl pipettes for reagent preparation
- 100 µl and 1 liter graduated cylinders
- · Tubes to prepare standard and sample dilutions
- · Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

#### **Target Details**

Target:	CD73 (NT5E)
Alternative Name:	5'-Nucleotidase (CD73 (NT5E Products)
UniProt:	Q61503
Pathways:	Synaptic Membrane, Ribonucleoside Biosynthetic Process
Application Details	

Application Notes:	Optimal working dilution should be determined by the investigator.
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 µl of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4°C.
	4. Add 100 μl of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 µl of prepared Streptavidin solution to each well.
	7. Incubate 45 min at RT.
	8. Add 100 µl of TMB One-Step Substrate Reagent to each well.
	9. Incubate 30 min at RT.
	10. Add 50 µl of Stop Solution to each well.
	11. Read at 450 nm immediately.
Reagent Preparation:	Recommended Dilution for serum and plasma samples100 fold
Restrictions:	For Research Use only

# Handling

Expiry Date:	6 months
	recommended to store at -80°C.
	freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is
Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated