

[Go to Product page](#)

## Datasheet for ABIN6730531 **CRELD2 ELISA Kit**

### Overview

Quantity:	96 tests
Target:	CRELD2
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	0.045 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.045 ng/mL
Application:	ELISA

### Product Details

Purpose:	Mouse CRELD2 ELISA Kit for Serum, Plasma, and Cell Culture Supernatants.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair recognizes Mouse CRELD2.
Characteristics:	<ul style="list-style-type: none"><li>• Strip plates and additional reagents allow for use in multiple experiments</li><li>• Quantitative protein detection</li><li>• Establishes normal range</li><li>• The best products for confirmation of antibody array data</li></ul>
Components:	<ul style="list-style-type: none"><li>• Pre-Coated 96-well Strip Microplate</li><li>• Wash Buffer</li><li>• Stop Solution</li></ul>

## Product Details

---

- Assay Diluent(s)
- Lyophilized Standard
- Biotinylated Detection Antibody
- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

---

### Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µl to 1 µl volumes
- Adjustable 1-25 µl pipettes for reagent preparation
- 100 µl and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

## Target Details

---

Target:	CRELD2
Alternative Name:	CRELD2 ( <a href="#">CRELD2 Products</a> )
UniProt:	<a href="#">Q9CYA0</a>

## Application Details

---

Application Notes:	Optimal working dilution should be determined by the investigator.
Protocol:	<ol style="list-style-type: none"><li>1. Prepare all reagents, samples and standards as instructed in the manual.</li><li>2. Add 100 µl of standard or sample to each well.</li><li>3. Incubate 2.5 h at RT or O/N at 4°C.</li><li>4. Add 100 µl of prepared biotin antibody to each well.</li><li>5. Incubate 1 h at RT.</li><li>6. Add 100 µl of prepared Streptavidin solution to each well.</li><li>7. Incubate 45 min at RT.</li><li>8. Add 100 µl of TMB One-Step Substrate Reagent to each well.</li><li>9. Incubate 30 min at RT.</li><li>10. Add 50 µl of Stop Solution to each well.</li><li>11. Read at 450 nm immediately.</li></ol>
Reagent Preparation:	Recommended Dilution for serum and plasma samples50 fold
Restrictions:	For Research Use only

## Handling

---

Storage Comment: The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.

---

Expiry Date: 6 months