

[Go to Product page](#)

Datasheet for ABIN6730772

**ROBO1 ELISA Kit**

## Overview

Quantity:	96 tests
-----------	----------

Target:	ROBO1
---------	-------

Reactivity:	Human
-------------	-------

Method Type:	Sandwich ELISA
--------------	----------------

Application:	ELISA
--------------	-------

## Product Details

Purpose:	Human ROBO1 ELISA Kit.
----------	------------------------

Sample Type:	Cell Culture Supernatant, Cell Samples, Plasma, Serum, Tissue Lysate
--------------	--

Analytical Method:	Quantitative
--------------------	--------------

Detection Method:	Colorimetric
-------------------	--------------

Specificity:	This ELISA antibody pair recognizes Human ROBO1.
--------------	--

Characteristics:	<ul style="list-style-type: none"><li>• Strip plates and additional reagents allow for use in multiple experiments</li><li>• Quantitative protein detection</li><li>• Establishes normal range</li><li>• The best products for confirmation of antibody array data</li></ul>
------------------	--

Components:	<ul style="list-style-type: none"><li>• Pre-Coated 96-well Strip Microplate</li><li>• Wash Buffer</li><li>• Stop Solution</li><li>• Assay Diluent(s)</li><li>• Lyophilized Standard</li><li>• Biotinylated Detection Antibody</li></ul>
-------------	---

## Product Details

---

- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

### Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µl to 1 µl volumes
- Adjustable 1-25 µl pipettes for reagent preparation
- 100 µl and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

## Target Details

---

Target: ROB01

Alternative Name: ROB01 ([ROB01 Products](#))

UniProt: [Q9Y6N7](#)

Pathways: [Positive Regulation of Endopeptidase Activity](#)

## Application Details

---

Application Notes: Optimal working dilution should be determined by the investigator.

Protocol:

1. Prepare all reagents, samples and standards as instructed in the manual.
2. Add 100 µl of standard or sample to each well.
3. Incubate 2.5 h at RT or O/N at 4°C.
4. Add 100 µl of prepared biotin antibody to each well.
5. Incubate 1 h at RT.
6. Add 100 µl of prepared Streptavidin solution to each well.
7. Incubate 45 min at RT.
8. Add 100 µl of TMB One-Step Substrate Reagent to each well.
9. Incubate 30 min at RT.
10. Add 50 µl of Stop Solution to each well.
11. Read at 450 nm immediately.

Restrictions: For Research Use only

## Handling

---

Storage Comment: The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is

## Handling

---

recommended to store at -80°C.

---

Expiry Date: 6 months