

[Go to Product page](#)

Datasheet for ABIN6730779 **Siglec E ELISA Kit**

Overview

Quantity:	96 tests
Target:	Siglec E (Siglece)
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Application:	ELISA

Product Details

Purpose:	Mouse Siglec-E ELISA Kit.
Sample Type:	Cell Culture Supernatant, Cell Samples, Plasma, Serum, Tissue Lysate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair recognizes Mouse Siglec-E.
Characteristics:	<ul style="list-style-type: none">• Strip plates and additional reagents allow for use in multiple experiments• Quantitative protein detection• Establishes normal range• The best products for confirmation of antibody array data
Components:	<ul style="list-style-type: none">• Pre-Coated 96-well Strip Microplate• Wash Buffer• Stop Solution• Assay Diluent(s)• Lyophilized Standard• Biotinylated Detection Antibody

Product Details

- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

Material not included:

- Distilled or deionized water
- Precision pipettes to deliver 2 µl to 1 µl volumes
- Adjustable 1-25 µl pipettes for reagent preparation
- 100 µl and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- Absorbent paper
- Microplate reader capable of measuring absorbance at 450nm
- Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target: Siglec E (Siglece)

Alternative Name: Siglec-E ([Siglece Products](#))

UniProt: [Q91Y57](#)

Application Details

Application Notes: Optimal working dilution should be determined by the investigator.

Protocol:

1. Prepare all reagents, samples and standards as instructed in the manual.
2. Add 100 µl of standard or sample to each well.
3. Incubate 2.5 h at RT or O/N at 4°C.
4. Add 100 µl of prepared biotin antibody to each well.
5. Incubate 1 h at RT.
6. Add 100 µl of prepared Streptavidin solution to each well.
7. Incubate 45 min at RT.
8. Add 100 µl of TMB One-Step Substrate Reagent to each well.
9. Incubate 30 min at RT.
10. Add 50 µl of Stop Solution to each well.
11. Read at 450 nm immediately.

Restrictions: For Research Use only

Handling

Storage Comment: The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is recommended to store at -80°C.

Handling

Expiry Date: 6 months