antibodies -online.com





Datasheet for ABIN6730838

TPSB2 ELISA Kit



Overview

Overview	
Quantity:	96 tests
Target:	TPSB2
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	0.84 ng/mL - 200 ng/mL
Minimum Detection Limit:	0.84 ng/mL
Application:	ELISA
Product Details	

Product Details	
Purpose:	Mouse Mast Cell Protease-6/Mcpt6 ELISA Kit for Serum, Plasma, and Cell Culture Supernatants.
Sample Type:	Cell Culture Supernatant, Plasma, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This ELISA antibody pair recognizes Mouse Mast Cell Protease-6.
Characteristics:	 Strip plates and additional reagents allow for use in multiple experiments Quantitative protein detection Establishes normal range The best products for confirmation of antibody array data
Components:	 Pre-Coated 96-well Strip Microplate Wash Buffer

Product Details

- · Stop Solution
- Assay Diluent(s)
- · Lyophilized Standard
- Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

Material not included:

- · Distilled or deionized water
- Precision pipettes to deliver 2 µl to 1 µl volumes
- Adjustable 1-25 µl pipettes for reagent preparation
- 100 µl and 1 liter graduated cylinders
- Tubes to prepare standard and sample dilutions
- · Absorbent paper

TPSB2

- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:

Alternative Name:	Mast Cell Protease-6 (Mcpt6 (TPSB2 Products)
UniProt:	P21845
Application Details	
Application Notes:	Optimal working dilution should be determined by the investigator.
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.
	2. Add 100 µl of standard or sample to each well.
	3. Incubate 2.5 h at RT or O/N at 4°C.
	4. Add 100 μl of prepared biotin antibody to each well.
	5. Incubate 1 h at RT.
	6. Add 100 µl of prepared Streptavidin solution to each well.
	7. Incubate 45 min at RT.
	8. Add 100 µl of TMB One-Step Substrate Reagent to each well.
	9. Incubate 30 min at RT.
	10. Add 50 µl of Stop Solution to each well.
	11. Read at 450 nm immediately.
Reagent Preparation:	Recommended Dilution for serum and plasma samples2 fold
Restrictions:	For Research Use only

Handling

Storage Comment:	The entire kit may be stored at -20°C for up to 1 year from the date of shipment. Avoid repeated freeze-thaw cycles. The kit may be stored at 4°C for up to 6 months. For extended storage, it is
	recommended to store at -80°C.
Expiry Date:	6 months