.-online.com antibodies

Datasheet for ABIN6730911 LRG1 ELISA Kit

5 Images

2 Publications



Overview

Quantity:	96 tests			
Target:	LRG1			
Reactivity:	Human			
Method Type:	Sandwich ELISA			
Detection Range:	3.12 ng/mL - 200 ng/mL			
Minimum Detection Limit:	3.12 ng/mL			
Application:	ELISA			
Product Details				
Purpose:	The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of LRG1 in			
	human serum, plasma, tissue homogenates, cell lysates, cell culture supernates.			
	We offer validation data (WB) for the kit components . So you can be sure to order a reliable			
	ELISA kit product composed of high quality reagents.			
Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate			
Analytical Method:	Quantitative			
Detection Method:	Colorimetric			
Specificity:	This assay has high sensitivity and excellent specificity for detection of Leucine Rich Alpha-			
	Glycoprotein 1 (LRG1)			
Cross-Reactivity (Details):	No significant cross-reactivity or interference between Leucine Rich Alpha-2-Glycoprotein 1 (LRG1) and analogues was observed.			

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/6 | Product datasheet for ABIN6730911 | 02/02/2024 | Copyright antibodies-online. All rights reserved.

Product Details

Sensitivity:	1.33 ng/mL
Components:	Pre-coated, ready to use 96-well strip plate, flat buttom
	Plate sealer for 96 wells
	Reference Standard
	Standard Diluent
	Detection Reagent A
	Detection Reagent B
	Assay Diluent A
	Assay Diluent B
	Reagent Diluent (if Detection Reagent is lyophilized)
	TMB Substrate
	Stop Solution
	• Wash Buffer (30 x concentrate)
	Instruction manual

Target Details

Target:	LRG1			
Alternative Name:	Leucine Rich Alpha-2-Glycoprotein 1 (LRG1) (LRG1 Products)			
UniProt:	P02750			
Pathways:	Brown Fat Cell Differentiation			
Application Details				
Comment:	Information on standard material:			
	The standard might be recombinant protein or natural protein, that will depend on the specific			
	kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin			
	300 in the standard as preservative.			
	Information on reagents:			
	The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash			
	solution is TBS. The standard diluent contains 0.02 % sodium azide, assay diluent A and assay			
	diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.			
	Information on antibodies:			
	The provided antibodies and their host vary in different kits.			
Sample Volume:	100 µL			

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/6 | Product datasheet for ABIN6730911 | 02/02/2024 | Copyright antibodies-online. All rights reserved.

Application Details

Assay Time:	3 h			
Plate:	Pre-coated			
Protocol:	1. Prepare all reagents, samples and standards,			
	2. Add 100 μ L standard or sample to each well. Incubate 1 hours at 37 °C,			
	3. Aspirate and add 100 μ L prepared Detection Reagent A. Incubate 1 hour at 37 °C,			
	4. Aspirate and wash 3 times,			
	5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,			
	6. Aspirate and wash 5 times,			
	7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,			
	8. Add 50µL Stop Solution. Read at 450nm immediately.			
Reagent Preparation:	1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit			
	will not be used up in one time, please only take out strips and reagents for present			
	experiment, and leave the remaining strips and reagents in required condition.			
	2. Standard - Reconstitute the Standard with 1.0mL of Standard Diluent, keep for 10 minutes at			
	room temperature, shake gently (not to foam). The concentration of the standard in the stock			
	solution is 200ng/mL. Prepare 7 tubes containing 0.5mL Standard Diluent and produce a			
	double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of			
	diluted standard such as 200ng/mL			
	,100ng/mL,50ng/mL,25ng/mL,12.5ng/mL,6.25ng/mL,3.12ng/mL,and the last			
	EP tube with Standard Diluent is the blank as 0 ng/mL.			
	3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection			
	Reagent A with 150 μ L of Reagent Diluent, keep for 10 minutes at room temperature, shake			
	gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before			
	use. Dilute them to the working concentration 100-fold with Assay Diluent A and B,			
	respectively.			
	4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized			
	or distilled water to prepare 600 mL of Wash Solution (1x).			
	5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not			
	dump the residual solution into the vial again.			
	Note:			
	1. Making serial dilution in the wells directly is not permitted.			
	2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.			
	3. Please carefully reconstitute Standards or working Detection Reagent A and B according to			
	the instruction, and avoid foaming and mix gently until the crystals are completely dissolved.			
	To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors			
	are calibrated. It is recommended to suck more than 10μ L for one pipetting.			
	4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only			
	once.			
	5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature			

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 3/6 | Product datasheet for ABIN6730911 | 02/02/2024 | Copyright antibodies-online. All rights reserved.


	and mix gently until the crystals are completely dissolved.				
	6. Contaminated water or container for reagent preparation will influence the detection result.				
Sample Preparation:	 It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 - 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates. If the sample type is not specified in the instructions, a preliminary test is necessary to determinecompatibility with the kit. If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a possibility of causing a deviation due to the introduced chemical substance. The recommended dilution factor is for reference only. Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2). 				
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level of target were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level of				
	target were tested on 3 different plates, 8 replicates in each plate.				
	CV(%) = SD/meanX100				
	Intra-Assay: CV < 10%				
	Inter-Assay: CV < 12%				
Restrictions:	For Research Use only				
Handling					
Precaution of Use:	The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material.				
Storage:	4 °C/-20 °C				
Storage Comment:	 For unopened kit: All reagents should be stored according to the labels on the vials. The Standard, Detection Reagent A, Detection Reagent B, and 96-well Strip Plate should be stored at -20 °C upon receipt, while the other reagents should be stored at 4 °C. For opened kits: the remaining reagents must be stored according to the above storage conditions. In addition, please return the unused wells to the foil pouch containing the desiccant and seal the foil pouch with the zipper. 				
Expiry Date:	6 months				

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/6 | Product datasheet for ABIN6730911 | 02/02/2024 | Copyright antibodies-online. All rights reserved. Product cited in:

Zhang, Huang, Tang, Yu, Huang, Chen, Wang, Wang: "Leucine-rich alpha-2-glycoprotein-1 is upregulated in colorectal cancer and is a tumor promoter." in: **OncoTargets and therapy**, Vol. 11, pp. 2745-2752, (2018) (PubMed).


Suresh, Saha, Kaur, Kumar, Dubey, Basak, Tanwar, Bhardwaj, Sengupta, Batra, Upadhyay: " Differentially expressed urinary biomarkers in children with idiopathic nephrotic syndrome." in: **Clinical and experimental nephrology**, Vol. 20, Issue 2, pp. 273-83, (2016) (PubMed).

Validation report #104288 for Cleavage Under Targets and Release Using Nuclease (CUT&RUN)



ELISA

Image 1. Typical standard curve



SDS-PAGE

Image 2. SDS-PAGE of Protein Standard from the Kit (Highly purified E. coli-expressed recombinant human LRG1).

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 5/6 | Product datasheet for ABIN6730911 | 02/02/2024 | Copyright antibodies-online. All rights reserved. Validation report #104288 for Cleavage Under Targets and Release Using Nuclease (CUT&RUN)




Image 3. WB of Protein Standard: different control antibodies against Highly purified E. coli-expressed	Western Blotting						
	Image 3.	WB of Pr	otein Standard	: different control			
no secolo de la contra de DO1	antibodies	against H	ighly purified	E. coli-expressed			
recombinant human LRG1.							

Please check the product details page for more images. Overall 5 images are available for ABIN6730911.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 6/6 | Product datasheet for ABIN6730911 | 02/02/2024 | Copyright antibodies-online. All rights reserved.