antibodies - online.com







LGALS3BP ELISA Kit





Publications



/ //	10	K / /	\sim 1	A 1
1 11	$^{\prime}$	I \/ I	-	ΛI
Ο١	$^{\prime}$	1 V I	-	/ V

Quantity:	96 tests	
Target:	LGALS3BP	
Reactivity:	Human	
Method Type:	Sandwich ELISA	
Detection Range:	1.56 ng/mL - 100 ng/mL	
Minimum Detection Limit:	1.56 ng/mL	
Application:	ELISA	
Product Details		
Purpose:	The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of	
	LGALS3BP in human serum, plasma, tissue homogenates, cell lysates, cell culture supernates.	
	We offer validation data (WB) for the kit components. So you can be sure to order a reliable	
	ELISA kit product composed of high quality reagents.	
Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This assay has high sensitivity and excellent specificity for detection of Lectin Galactoside	
	Binding, Soluble 3 Binding Protein (LGALS3BP)	
Cross-Reactivity (Details):	No significant cross-reactivity or interference between Lectin Galactoside Binding, Soluble 3	
	Binding Protein (LGALS3BP) and analogues was observed.	

Product Details

Sensitivity:	0.63 ng/mL
Sensitivity: Components:	 O.63 ng/mL Pre-coated, ready to use 96-well strip plate, flat buttom Plate sealer for 96 wells Reference Standard Standard Diluent Detection Reagent A Detection Reagent B Assay Diluent A Assay Diluent B Reagent Diluent (if Detection Reagent is lyophilized) TMB Substrate Stop Solution Wash Buffer (30 x concentrate)
	Instruction manual

Target Details

Assay Time:

3 h

Target:	LGALS3BP	
Alternative Name:	Lectin Galactoside Binding, Soluble 3 Binding Protein (LGALS3BP) (LGALS3BP Products)	
UniProt:	Q08380	
Application Details		
Comment:	Information on standard material:	
	The standard might be recombinant protein or natural protein, that will depend on the specific	
	kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin	
	300 in the standard as preservative.	
	Information on reagents:	
	The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash	
	solution is TBS. The standard diluent contains 0.02 % sodium azide, assay diluent A and assay	
	diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.	
	Information on antibodies:	
	The provided antibodies and their host vary in different kits.	
Sample Volume:	100 μL	

Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards,
	2. Add 100µL standard or sample to each well. Incubate 1 hours at 37 °C,
	3. Aspirate and add 100µL prepared Detection Reagent A. Incubate 1 hour at 37 °C,
	4. Aspirate and wash 3 times,
	5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,
	6. Aspirate and wash 5 times,
	7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,
	8. Add 50µL Stop Solution. Read at 450nm immediately.
Reagent Preparation:	1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit
	will not be used up in one time, please only take out strips and reagents for present
	experiment, and leave the remaining strips and reagents in required condition.
	2. Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at
	room temperature, shake gently (not to foam). The concentration of the standard in the stock
	solution is 400 ng/mL. Firstly dilute the stock solution to 100 ng/mL and the diluted standard
	serves as the highest standard (100 ng/mL). Then prepare 7 tubes containing 0.5 mL
	Standard Diluent and use the diluted standard to produce a double dilution series. Mix each
	tube thoroughly before the next transfer. Set up 7 points of diluted standard such as
	100 ng/mL, 50 ng/mL, 25 ng/mL, 12.5 ng/mL, 6.25 ng/mL, 3.12 ng/mL, 1.56 ng/mL, and the
	last microcentrifuge tube with Standard Diluent is the blank as 0 ng/mL.
	3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection
	Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake
	gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before
	use. Dilute them to the working concentration 100-fold with Assay Diluent A and B,
	respectively.
	4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized
	or distilled water to prepare 600 mL of Wash Solution (1x).
	5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not
	dump the residual solution into the vial again.
	Note:
	1. Making serial dilution in the wells directly is not permitted.
	2. Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.
	3. Please carefully reconstitute Standards or working Detection Reagent A and B according to
	the instruction, and avoid foaming and mix gently until the crystals are completely dissolved.
	To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors

and mix gently until the crystals are completely dissolved.

once.

are calibrated. It is recommended to suck more than 10µL for one pipetting.

4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only

5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature

	6. Contaminated water or container for reagent preparation will influence the detection result.		
Sample Preparation:	 It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 - 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates. If the sample type is not specified in the instructions, a preliminary test is necessary to determinecompatibility with the kit. If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a possibility of causing a deviation due to the introduced chemical substance. The recommended dilution factor is for reference only. Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2). 		
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level of		
	target were tested 20 times on one plate, respectively.		
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level of		
	target were tested on 3 different plates, 8 replicates in each plate.		
	CV(%) = SD/meanX100		
	Intra-Assay: CV < 10%		
	Inter-Assay: CV < 12%		
Restrictions:	For Research Use only		
Handling			
Precaution of Use:	The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and		
	clothing protection when using this material.		
Storage:	4 °C/-20 °C		
Storage Comment:	 For unopened kit: All reagents should be stored according to the labels on the vials. The Standard, Detection Reagent A, Detection Reagent B, and 96-well Strip Plate should be stored at -20 °C upon receipt, while the other reagents should be stored at 4 °C. For opened kits: the remaining reagents must be stored according to the above storage conditions. In addition, please return the unused wells to the foil pouch containing the desiccant and seal the foil pouch with the zipper. 		
Expiry Date:	6 months		


Product cited in:

Mokart, Saillard, Zemmour, Bisbal, Sannini, Chow-Chine, Brun, Faucher, Boher, Toiron, Chabannon, Borg, Gonçalves, Camoin: "Early prognostic factors in septic shock cancer patients: a prospective study with a proteomic approach." in: **Acta anaesthesiologica Scandinavica**, Vol. 62, Issue 4, pp. 493-503, (2018) (PubMed).

Ferrín, Ranchal, Llamoza, Rodríguez-Perálvarez, Romero-Ruiz, Aguilar-Melero, López-Cillero, Briceño, Muntané, Montero-Álvarez, De la Mata: "Identification of candidate biomarkers for hepatocellular carcinoma in plasma of HCV-infected cirrhotic patients by 2-D DIGE." in: **Liver international : official journal of the International Association for the Study of the Liver**, Vol. 34, Issue 3, pp. 438-46, (2015) (PubMed).


Freour, Com, Barriere, Bouchot, Jean, Masson, Pineau: "Comparative proteomic analysis coupled with conventional protein assay as a strategy to identify predictors of successful testicular sperm extraction in patients with non-obstructive azoospermia." in: **Andrology**, Vol. 1, Issue 3, pp. 414-20, (2013) (PubMed).

Images



ELISA

Image 1. Typical standard curve



Western Blotting

Image 2. Mouse Capture antibody from the kit in WB with Positive Control: Human HepG2 cell lysate.

Western Blotting

Image 3. Mouse Detection antibody from the kit in WB with Positive Control: Sample Human 293T cell lysate.

Please check the product details page for more images. Overall 5 images are available for ABIN6730946.