

Datasheet for ABIN6810505

IL18BP ELISA Kit



Overview

Quantity:	96 tests
Target:	IL18BP
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	22 pg/mL - 5000 pg/mL
Minimum Detection Limit:	22 pg/mL
Application:	ELISA

Product Details

Product Details		
Purpose:	Mouse IL-18 BPc ELISA Kit for Serum, Plasma, and Cell Culture Supernatants.	
Sample Type:	Cell Culture Supernatant, Plasma, Serum	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This ELISA antibody pair recognizes Mouse IL-18 BPc.	
Characteristics:	 Strip plates and additional reagents allow for use in multiple experiments Quantitative protein detection Establishes normal range The best products for confirmation of antibody array data 	
Components:	Pre-Coated 96-well Strip MicroplateWash BufferStop Solution	

Product Details

- Assay Diluent(s)
- · Lyophilized Standard
- · Biotinylated Detection Antibody
- · Streptavidin-Conjugated HRP
- · TMB One-Step Substrate

Material not included:

- · Distilled or deionized water
- Precision pipettes to deliver 2 μl to 1 μl volumes
- Adjustable 1-25 µl pipettes for reagent preparation
- 100 µl and 1 liter graduated cylinders
- · Tubes to prepare standard and sample dilutions
- · Absorbent paper

IL18BP

- Microplate reader capable of measuring absorbance at 450nm
- · Log-log graph paper or computer and software for ELISA data analysis

Target Details

Target:

Alternative Name:	IL-18 BPc (IL18BP Products)		
UniProt:	Q9Z0M9		
Application Details			
Application Notes:	Optimal working dilution should be determined by the investigator.		
Protocol:	1. Prepare all reagents, samples and standards as instructed in the manual.		
	2. Add 100 μl of standard or sample to each well.		
	3. Incubate 2.5 h at RT or O/N at 4°C.		
	4. Add 100 μl of prepared biotin antibody to each well.		
	5. Incubate 1 h at RT.		
	6. Add 100 μl of prepared Streptavidin solution to each well.		
	7. Incubate 45 min at RT.		
	8. Add 100 μl of TMB One-Step Substrate Reagent to each well.		
	9. Incubate 30 min at RT.		
	10. Add 50 μl of Stop Solution to each well.		
	11. Read at 450 nm immediately.		
Reagent Preparation:	Recommended Dilution for serum and plasma samples4 fold		
Restrictions:	For Research Use only		

1 1		
\blacksquare	$\alpha \alpha \alpha$	lina
	land	1111(
	I G I G	11119

Expiry Date:

6 months