



[Go to Product page](#)

Datasheet for ABIN6938813

Folic Acid ELISA Kit

Overview

Quantity: 96 tests

Target: Folic Acid (FA)

Reactivity: Chemical

Application: ELISA

Product Details

Purpose: The Folic Acid ELISA Kit is a competitive enzyme immunoassay developed for rapid detection and quantitation of folic acid in serum, cell or tissue samples.

Sample Type: Cell Samples, Plasma, Serum, Tissue Samples

Analytical Method: Quantitative

Detection Method: Colorimetric

Sensitivity: 24 pg/mL

Characteristics: The Folic Acid ELISA Kit is a competitive enzyme immunoassay developed for rapid detection and quantitation of folic acid in serum, cell or tissue samples. The quantity of folic acid in unknown samples is determined by comparing its absorbance with that of a known folic acid standard curve. The kit has detection sensitivity limit of 24 pg/mL folic acid. Each Folic Acid ELISA Kit provides sufficient reagents to perform up to 96 assays, including standard curve and unknown samples.

Components:

1. 96-well Protein Binding Plate : One strip well 96-well plate.
2. Anti-Folic Acid Antibody (500X) : One 10 µL vial.
3. Secondary Antibody, HRP Conjugate : One 20 µL vial.
4. Assay Diluent : One 50 mL bottle.

Product Details

5. 10X Wash Buffer : One 100 mL bottle.
6. Substrate Solution : One 12 mL amber bottle.
7. Stop Solution: One 12 mL bottle.
8. Folic Acid Standard : One 100 µL amber vial of 10 µg/mL Folic Acid in water.

Target Details

Target:	Folic Acid (FA)																		
Alternative Name:	Folic Acid (FA Products)																		
Target Type:	Chemical																		
Background:	<p>Folic acid is a B vitamin also known as Vitamin B9. Since humans don't synthesize folic acid, it is required from the diet and is therefore considered to be an essential vitamin. In cells, folic acid is required for amino acid metabolism as well as to carry one-carbon groups used for methylation reactions and synthesis of nucleic acids (such as thymine and purine bases). Therefore, deficiency in folic acid disrupts DNA synthesis and cell division, affecting mostly hematopoietic cells and abnormal tissue growths because of their higher rate of cell division. Folic acid is used to supplement folic acid deficiency. This deficiency can cause anemia. Symptoms of anemia can include fatigue, heart palpitations, difficulty breathing, open sores observed on the tongue, and color changes of the hair or skin. Deficiency can occur in children after only a month of consuming a folic acid deficient diet. In adults, normal total body folic acid levels are between 10,000-30,000 micrograms (µg) with plasma levels of greater than 7 nM (3 ng/mL) (Table 1). Women also take supplemental folic acid during pregnancy to prevent fetal neural tube defects (NTDs). Insufficient levels of dietary folic acid in early pregnancy are thought to be the cause of over half of babies born with neural tube defects. As a result, over 50 countries add folic acid to certain foods as a way to decrease NTD incidents in the population.</p> <table><thead><tr><th>Concentration (ng/mL)</th><th>Concentration (nM)</th><th>Adults</th><th>3-20</th><th>7-45.3</th><th>Children</th><th>5-21</th><th>11.3-47.6</th><th>Infants</th></tr></thead><tbody><tr><td>14-51</td><td>31.7-115.5</td><td></td><td></td><td></td><td></td><td></td><td></td><td></td></tr></tbody></table> <p>Reference ranges for folic acid in human plasma.</p>	Concentration (ng/mL)	Concentration (nM)	Adults	3-20	7-45.3	Children	5-21	11.3-47.6	Infants	14-51	31.7-115.5							
Concentration (ng/mL)	Concentration (nM)	Adults	3-20	7-45.3	Children	5-21	11.3-47.6	Infants											
14-51	31.7-115.5																		

Application Details

Application Notes:	Optimal working dilution should be determined by the investigator.
Reagent Preparation:	<ul style="list-style-type: none">Folic Acid Conjugate Coated Plate: Dilute the proper amount of 100X Folic Acid Conjugate 1:100 into 1X PBS. Add 100 µL of the diluted 1X Folic Acid Conjugate to each well and incubate at 37 °C for two hours or overnight at 4 °C. Remove the coating solution and wash twice with 200 µL 3 of 1X PBS. Blot plate on paper towels to remove excess fluid. Add 200 µL

of Assay Diluent to each well and block for 1 hr at room temperature. Transfer the plate to 4 °C and remove the Assay Diluent immediately before use. Note: The Folic Acid Conjugate-coated wells are not stable and should be used within 24 hrs after coating. Only coat the number of wells to be used immediately.

- 1X Wash Buffer: Dilute the 10X Wash Buffer to 1X with deionized water. Stir to homogeneity.
 - Anti-Folic Acid Antibody and Secondary Antibody: Immediately before use dilute the Anti-Folic Acid Antibody 1:500 and Secondary Antibody 1:1000 with Assay Diluent. Do not store diluted solutions. Preparation of Standard Curve 1. Use the provided stock Folic Acid Standard 10 µg/mL solution to prepare a series of the remaining standards according to Table 1 below.
- | Folic Acid Standard (µg/mL) | Folic Acid Assay Diluent (µL) | Folic Acid Standard Tubes | Standard (µL) | (µL) | (ng/mL) | (nM) |
|-----------------------------|-------------------------------|---------------------------|---------------|------------|------------|------------|
| 10 | 990 | 100 | 227 | 2 | 100 | of Tube #1 |
| 300 | 25 | 56.75 | 3 | 100 | of Tube #2 | 300 |
| 6.25 | 14.19 | 4 | 100 | of Tube #3 | 300 | 1.56 |
| 3.55 | 5 | 100 | of Tube #4 | 300 | 0.391 | 0.887 |
| 6 | 100 | of Tube #5 | 300 | 0.098 | 0.222 | 7 |
| 100 | of Tube #6 | 300 | 0.024 | 0.055 | 8 | 0 |
| 300 | 0 | 0 | 300 | 0 | 0 | 0 |
- Table 1. Preparation of Folic Acid Standards.

Sample Preparation:

- Serum: Avoid hemolyzed and lipemic blood samples. Collect blood in a tube with no anticoagulant. Allow the blood to clot at room temperature for 30 minutes. Centrifuge at 2500 x g for 20 minutes. Remove the yellow serum supernatant without disturbing the white buffy layer. Aliquot samples for testing and store at -80 °C. Perform dilutions in Assay Diluent or PBS containing 0.1 % BSA as necessary.
- Plasma: Avoid hemolyzed and lipemic blood samples. Collect blood with heparin or citrate and centrifuge at 2000 x g and 4 °C for 10 minutes. Remove the plasma layer and store on ice. Avoid disturbing the white buffy layer. Aliquot samples for testing and store at -80 °C. Perform dilutions in Assay Diluent or PBS containing 0.1 % BSA as necessary. Note: This assay is not compatible with rabbit serum or plasma due to high levels of rabbit IgG that will cross react with the secondary antibody.
- Cells or tissues: Homogenize 50-200 mg of the cell pellet or tissue in 0.5-2 mL of ice cold PBS using a mortar and pestle or by dounce homogenization. Incubate the homogenate at 4 °C for 20 4 minutes. Transfer the homogenate to a centrifuge tube and centrifuge at 12000 x g for 20 minutes. Recover the supernatant and transfer to a fresh tube. Store resuspended sample at -20 °C or colder until ready to test by ELISA. Perform dilutions in Assay Diluent or PBS containing 0.1 % BSA as necessary.

Assay Procedure:

1. Prepare and mix all reagents thoroughly before use. Each folic acid sample including unknown and standard should be assayed in duplicate.
2. Add 50 µL of unknown sample or Folic Acid standards to the wells of the Folic Acid Conjugate coated plate. Incubate at room temperature for 10 minutes on an orbital shaker.
3. Add 50 µL of the diluted Anti-Folic Acid antibody to each well, incubate at room temperature for 1 hour on an orbital shaker.
4. Wash microwell strips 3 times with 250 µL 1X Wash Buffer per well with thorough aspiration between each wash. After the last wash, empty wells and tap microwell strips on absorbent pad or paper towel to remove excess 1X Wash Buffer.
5. Add 100 µL of the diluted Secondary Antibody-HRP Enzyme Conjugate to all wells.
6. Incubate at room temperature for 1 hour on an orbital shaker.

Application Details

7. Wash microwell strips 3 times according to step 4 above. Proceed immediately to the next step.
8. Warm Substrate Solution to room temperature. Add 100 μ L of Substrate Solution to each well, including the blank wells. Incubate at room temperature on an orbital shaker. Actual incubation time may vary from 2-30 minutes. Note: Watch plate carefully, if color changes rapidly, the reaction may need to be stopped sooner to prevent saturation.
9. Stop the enzyme reaction by adding 100 μ L of Stop Solution into each well, including the blank wells. Results should be read immediately (color will fade over time).
10. Read absorbance of each microwell on a spectrophotometer using 450 nm as the primary wave length.

Restrictions: For Research Use only

Handling

Handling Advice: Avoid multiple freeze/thaw cycles.

Storage: 4 °C, -20 °C

Storage Comment: Upon receipt, aliquot and store 100X Folic Acid Conjugate at -20°C and avoid multiple freeze/thaw cycles. Store all other components at 4°C. The 100X Folic Acid Conjugate and Folic Acid Standard are light sensitive and must be stored accordingly.