



[Go to Product page](#)

Datasheet for ABIN6951261  
**SCLY ELISA Kit**

### Overview

Quantity: 96 tests

Target: SCLY

Reactivity: Human

Method Type: Sandwich ELISA

Application: ELISA

### Product Details

Purpose: Human SCLY ELISA Kit.

Sample Type: Cell Culture Supernatant, Cell Samples, Plasma, Serum, Tissue Lysate

Analytical Method: Quantitative

Detection Method: Colorimetric

Characteristics:

- Strip plates and additional reagents allow for use in multiple experiments
- Quantitative protein detection
- Establishes normal range
- The best products for confirmation of antibody array data

Components:

- Pre-Coated 96-well Strip Microplate
- Wash Buffer
- Stop Solution
- Assay Diluent(s)
- Lyophilized Standard
- Biotinylated Detection Antibody
- Streptavidin-Conjugated HRP
- TMB One-Step Substrate

## Product Details

---

Material not included:	<ul style="list-style-type: none"><li>• Distilled or deionized water</li><li>• Precision pipettes to deliver 2 µl to 1 µl volumes</li><li>• Adjustable 1-25 µl pipettes for reagent preparation</li><li>• 100 µl and 1 liter graduated cylinders</li><li>• Tubes to prepare standard and sample dilutions</li><li>• Absorbent paper</li><li>• Microplate reader capable of measuring absorbance at 450nm</li><li>• Log-log graph paper or computer and software for ELISA data analysis</li></ul>
------------------------	---

## Target Details

---

Target:	SCLY
Alternative Name:	SCLY ( <a href="#">SCLY Products</a> )
Gene ID:	50880
UniProt:	<a href="#">Q9JL16</a>
Pathways:	<a href="#">Stem Cell Maintenance</a>

## Application Details

---

Application Notes:	Optimal working dilution should be determined by the investigator.
Plate:	Pre-coated
Protocol:	<ol style="list-style-type: none"><li>1. Prepare all reagents, samples and standards as instructed in the manual.</li><li>2. Add 100 µl of standard or sample to each well.</li><li>3. Incubate 2.5 h at RT or O/N at 4°C.</li><li>4. Add 100 µl of prepared biotin antibody to each well.</li><li>5. Incubate 1 h at RT.</li><li>6. Add 100 µl of prepared Streptavidin solution to each well.</li><li>7. Incubate 45 min at RT.</li><li>8. Add 100 µl of TMB One-Step Substrate Reagent to each well.</li><li>9. Incubate 30 min at RT.</li><li>10. Add 50 µl of Stop Solution to each well.</li><li>11. Read at 450 nm immediately.</li></ol>
Restrictions:	For Research Use only

## Handling

---

Expiry Date:	6 months
--------------	----------