

[Go to Product page](#)

Datasheet for ABIN6951739 **HIST1H3A ELISA Kit**

Overview

Quantity:	96 tests
Target:	HIST1H3A
Reactivity:	Human
Method Type:	Sandwich ELISA
Application:	ELISA

Product Details

Purpose:	Human HIST1H3A ELISA Kit.
Sample Type:	Cell Culture Supernatant, Cell Samples, Plasma, Serum, Tissue Lysate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Characteristics:	<ul style="list-style-type: none">• Strip plates and additional reagents allow for use in multiple experiments• Quantitative protein detection• Establishes normal range• The best products for confirmation of antibody array data
Components:	<ul style="list-style-type: none">• Pre-Coated 96-well Strip Microplate• Wash Buffer• Stop Solution• Assay Diluent(s)• Lyophilized Standard• Biotinylated Detection Antibody• Streptavidin-Conjugated HRP• TMB One-Step Substrate

Product Details

Material not included:	<ul style="list-style-type: none">• Distilled or deionized water• Precision pipettes to deliver 2 µl to 1 µl volumes• Adjustable 1-25 µl pipettes for reagent preparation• 100 µl and 1 liter graduated cylinders• Tubes to prepare standard and sample dilutions• Absorbent paper• Microplate reader capable of measuring absorbance at 450nm• Log-log graph paper or computer and software for ELISA data analysis
------------------------	---

Target Details

Target:	HIST1H3A
Alternative Name:	HIST1H3A (HIST1H3A Products)
UniProt:	P68431

Application Details

Application Notes:	Optimal working dilution should be determined by the investigator.
Plate:	Pre-coated
Protocol:	<ol style="list-style-type: none">1. Prepare all reagents, samples and standards as instructed in the manual.2. Add 100 µl of standard or sample to each well.3. Incubate 2.5 h at RT or O/N at 4°C.4. Add 100 µl of prepared biotin antibody to each well.5. Incubate 1 h at RT.6. Add 100 µl of prepared Streptavidin solution to each well.7. Incubate 45 min at RT.8. Add 100 µl of TMB One-Step Substrate Reagent to each well.9. Incubate 30 min at RT.10. Add 50 µl of Stop Solution to each well.11. Read at 450 nm immediately.
Restrictions:	For Research Use only

Handling

Expiry Date:	6 months
--------------	----------