.-online.com antibodies

Datasheet for ABIN6954594 CCL2 ELISA Kit

1 Image

6 Publications



Overview

Quantity:	96 tests
Target:	CCL2
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	15.6 pg/mL - 1000 pg/mL
Minimum Detection Limit:	15.6 pg/mL
Application:	ELISA

Product Details

The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of MCP1 in human serum, plasma, tissue homogenates, cell lysates, cell culture supernates.
Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate
Quantitative
Colorimetric
This assay has high sensitivity and excellent specificity for detection of Monocyte Chemotactic Protein 1 (MCP1)
6.4 pg/mL
 Pre-coated, ready to use 96-well strip plate, flat buttom Plate sealer for 96 wells Reference Standard

• Standard Diluent

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/5 | Product datasheet for ABIN6954594 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

- Detection Reagent A
- Detection Reagent B
- Assay Diluent A
- Assay Diluent B
- Reagent Diluent (if Detection Reagent is lyophilized)
- TMB Substrate
- Stop Solution
- Wash Buffer (30 x concentrate)
- Instruction manual

Target Details

Target:	CCL2
Alternative Name:	Monocyte Chemotactic Protein 1 (MCP1) (CCL2 Products)
Pathways:	Cellular Response to Molecule of Bacterial Origin, Positive Regulation of Immune Effector Process, ER-Nucleus Signaling, Unfolded Protein Response, The Global Phosphorylation Landscape of SARS-CoV-2 Infection
Application Details	

Comment:	Information on standard material:
	The standard might be recombinant protein or natural protein, that will depend on the specific
	kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin
	300 in the standard as preservative.
	Information on reagents:
	The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash
	solution is TBS. The standard diluent contains 0.02 % sodium azide, assay diluent A and assay
	diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.
	Information on antibodies:
	The provided antibodies and their host vary in different kits.
Sample Volume:	100 μL
Assay Time:	3 h
Plate:	Pre-coated

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/5 | Product datasheet for ABIN6954594 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

Application Details

Protocol:	1. Prepare all reagents, samples and standards,
	2. Add 100µL standard or sample to each well. Incubate 1 hours at 37 °C,
	3. Aspirate and add 100 μ L prepared Detection Reagent A. Incubate 1 hour at 37 °C,
	4. Aspirate and wash 3 times,
	5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,
	6. Aspirate and wash 5 times,
	7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,
	8. Add 50µL Stop Solution. Read at 450nm immediately.
Reagent Preparation:	1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit
	will not be used up in one time, please only take out strips and reagents for present
	experiment, and leave the remaining strips and reagents in required condition.
	2. Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at
	room temperature, shake gently (not to foam). The concentration of the standard in the stock
	solution is 2,000pg/mL. Firstly dilute the stock solution to 1,000pg/mL and the diluted
	standard serves as the highest standard (1,000pg/mL). Then prepare 7 tubes containing
	0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix
	each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as
	1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the
	last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.
	3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection
	Reagent A with 150 μ L of Reagent Diluent, keep for 10 minutes at room temperature, shake
	gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before
	use. Dilute them to the working concentration 100-fold with Assay Diluent A and B, respectively.
	4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized
	or distilled water to prepare 600 mL of Wash Solution (1x).
	5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not
	dump the residual solution into the vial again.
	Note:
	1. Making serial dilution in the wells directly is not permitted.
	 Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.
	3. Please carefully reconstitute Standards or working Detection Reagent A and B according to
	the instruction, and avoid foaming and mix gently until the crystals are completely dissolved.
	To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors
	are calibrated. It is recommended to suck more than 10μ L for one pipetting.
	4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.
	5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature
	and mix gently until the crystals are completely dissolved.
	6. Contaminated water or container for reagent preparation will influence the detection result.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 3/5 | Product datasheet for ABIN6954594 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

ot in therange of the standard curve, the optimal sample dilution for the particular iment has to be determined.Samples should then be diluted with PBS (pH =7.0-7.2). say Precision (Precision within an assay): 3 samples with low, middle and high level of vere tested 20 times on one plate, respectively. say Precision (Precision between assays): 3 samples with low, middle and high level of vere tested on 3 different plates, 8 replicates in each plate. SD/meanX100 say: CV < 10% say: CV < 12% earch Use only
iment has to be determined.Samples should then be diluted with PBS (pH =7.0-7.2). say Precision (Precision within an assay): 3 samples with low, middle and high level of vere tested 20 times on one plate, respectively. say Precision (Precision between assays): 3 samples with low, middle and high level of vere tested on 3 different plates, 8 replicates in each plate. SD/meanX100 say: CV < 10%
iment has to be determined.Samples should then be diluted with PBS (pH =7.0-7.2). say Precision (Precision within an assay): 3 samples with low, middle and high level of vere tested 20 times on one plate, respectively. say Precision (Precision between assays): 3 samples with low, middle and high level of vere tested on 3 different plates, 8 replicates in each plate. SD/meanX100
iment has to be determined.Samples should then be diluted with PBS (pH =7.0-7.2). say Precision (Precision within an assay): 3 samples with low, middle and high level of vere tested 20 times on one plate, respectively. say Precision (Precision between assays): 3 samples with low, middle and high level of vere tested on 3 different plates, 8 replicates in each plate.
iment has to be determined.Samples should then be diluted with PBS (pH =7.0-7.2). say Precision (Precision within an assay): 3 samples with low, middle and high level of vere tested 20 times on one plate, respectively. say Precision (Precision between assays): 3 samples with low, middle and high level of
iment has to be determined.Samples should then be diluted with PBS (pH =7.0-7.2). say Precision (Precision within an assay): 3 samples with low, middle and high level of vere tested 20 times on one plate, respectively.
iment has to be determined.Samples should then be diluted with PBS (pH =7.0-7.2). say Precision (Precision within an assay): 3 samples with low, middle and high level of
ot in therange of the standard curve, the optimal sample dilution for the particular
e estimate the concentration of the samples before performing the test. If the values
bilityof causing a deviation due to the introduced chemical substance.The nmended dilution factor is for reference only.
sis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a
minecompatibility with the kit.
sample type is not specified in the instructions, a preliminary test is necessary to
ples should be slowly thawed and centrifuged toremove precipitates.
fore be stored for a short periodat 2 - 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ nths). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen
lenaturationmay occur in these samples, leading to false results. Samples should
f r l

Precaution of Use:	The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and
	clothing protection when using this material.
Storage:	4 °C/-20 °C
Storage Comment:	 For unopened kit: All reagents should be stored according to the labels on the vials. The Standard, Detection Reagent A, Detection Reagent B, and 96-well Strip Plate should be stored at -20 °C upon receipt, while the other reagents should be stored at 4 °C. For opened kits: the remaining reagents must be stored according to the above storage conditions. In addition, please return the unused wells to the foil pouch containing the desiccant and seal the foil pouch with the zipper.
Expiry Date:	6 months

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/5 | Product datasheet for ABIN6954594 | 09/10/2023 | Copyright antibodies-online. All rights reserved. Product cited in:

Shen, Zhang, Feng, Chen, Li, Iqbal, Liao, Zhao, Zhai: "iTRAQ-Based Proteomic Analysis Reveals Protein Profile in Plasma from Children with Autism." in: **Proteomics. Clinical applications**, Vol. 12, Issue 3, pp. e1700085, (2019) (PubMed).


Klupczynska, Swiatly, Hajduk, Matysiak, Dyszkiewicz, Pawlak, Kokot: "Identification of Serum Peptidome Signatures of Non-Small Cell Lung Cancer." in: **International journal of molecular sciences**, Vol. 17, Issue 4, pp. 410, (2016) (PubMed).

Ma, Cui, Liao, Zhao: "Complement activation contributes to the injury and outcome of kidney in human anti-glomerular basement membrane disease." in: **Journal of clinical immunology**, Vol. 33, Issue 1, pp. 172-8, (2013) (PubMed).

Burillo, Mateo-Gallego, Cenarro, Fiddyment, Bea, Jorge, Vázquez, Civeira: "Beneficial effects of omega-3 fatty acids in the proteome of high-density lipoprotein proteome." in: **Lipids in health and disease**, Vol. 11, Issue 1, pp. 116, (2012) (PubMed).

Oberbach, Blüher, Wirth, Till, Kovacs, Kullnick, Schlichting, Tomm, Rolle-Kampczyk, Murugaiyan, Binder, Dietrich, von Bergen: "Combined proteomic and metabolomic profiling of serum reveals association of the complement system with obesity and identifies novel markers of body fat mass changes." in: **Journal of proteome research**, Vol. 10, Issue 10, pp. 4769-88, (2011) (PubMed).

There are more publications referencing this product on: Product page



Images

ELISA

Image 1. Typical standard curve

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 5/5 | Product datasheet for ABIN6954594 | 09/10/2023 | Copyright antibodies-online. All rights reserved.