.-online.com antibodies

Datasheet for ABIN6954858 F2RL1 ELISA Kit

1 Image

6 Publications



Overview

Quantity:	96 tests
Target:	F2RL1
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.15 ng/mL - 10 ng/mL
Minimum Detection Limit:	0.15 ng/mL
Application:	ELISA

Product Details

human tissSample Type:Cell LysateAnalytical Method:QuantitativeDetection Method:Colorimetre	a sandwich enzyme immunoassay for in vitro quantitative measurement of PAR2 in sue homogenates, cell lysates.
Analytical Method: Quantitative Detection Method: Colorimetre	
Detection Method: Colorimetr	e, Tissue Homogenate
	ve
Specificity: This assay	ric
Receptor 2	y has high sensitivity and excellent specificity for detection of Protease Activated 2 (PAR2)
Sensitivity: 0.063 ng/r	mL
Plate se	ted, ready to use 96-well strip plate, flat buttom ealer for 96 wells ce Standard

• Standard Diluent

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/5 | Product datasheet for ABIN6954858 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

- Detection Reagent A
- Detection Reagent B
- Assay Diluent A
- Assay Diluent B
- Reagent Diluent (if Detection Reagent is lyophilized)
- TMB Substrate
- Stop Solution
- Wash Buffer (30 x concentrate)
- Instruction manual

Target Details

Target:	F2RL1
Alternative Name:	Protease Activated Receptor 2 (PAR2) (F2RL1 Products)
Pathways:	Regulation of Leukocyte Mediated Immunity, Positive Regulation of Immune Effector Process,
	Production of Molecular Mediator of Immune Response, SARS-CoV-2 Protein Interactome

Application Details

Comment:	Information on standard material:
	The standard might be recombinant protein or natural protein, that will depend on the specific
	kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin
	300 in the standard as preservative.
	Information on reagents:
	The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash
	solution is TBS. The standard diluent contains 0.02 % sodium azide, assay diluent A and assay
	diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.
	Information on antibodies:
	The provided antibodies and their host vary in different kits.
Sample Volume:	100 µL
Assay Time:	3 h
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards,
	2. Add 100 μL standard or sample to each well. Incubate 1 hours at 37 °C,

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/5 | Product datasheet for ABIN6954858 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

3. Aspirate and add 100µL	prepared Detection Reagent A.	Incubate 1 hour at 37 °C	,
---------------------------	-------------------------------	--------------------------	---

- 4. Aspirate and wash 3 times,
- 5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,
- 6. Aspirate and wash 5 times,
- 7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,
- 8. Add 50µL Stop Solution. Read at 450nm immediately.

Reagent Preparation: 1. Bring all kit components and samples to room temperature (18-25 °C) before use.

- 2. Standard Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 20 ng/mL. Firstly dilute the stock solution to 10 ng/mL and the diluted standard serves as the highest standard (10 ng/mL). Then prepare 7 tubes containing 0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 10 ng/mL, 5 ng/mL, 2.5 ng/mL, 1.25 ng/mL, 0.625 ng/mL, 0.312 ng/mL, 0.156 ng/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0 ng/mL.
 - 3. Detection Reagent A and Detection Reagent B Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute to the working concentration with Assay Diluent A and B, respectively (1:100).
 - 4. Wash Solution Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).
 - 5. TMB substrate Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.

Note:

- 1. Making serial dilution in the wells directly is not permitted.
- 2. Prepare standard within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.
- 3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for once pipetting.
- 4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.
- 5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature and mix gently until the crystals are completely dissolved.
- 6. Contaminated water or container for reagent preparation will influence the detection result.

Sample Preparation:

 It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 - 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates.

If the sample type is not specified in the instructions, a preliminary test is necessary to

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 3/5 | Product datasheet for ABIN6954858 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

	Kouzaki, Matsumoto, Kato, Tojima, Shimizu, Shimizu: "Epithelial Cell-Derived Cytokines Contribute to the Pathophysiology of Eosinophilic Chronic Rhinosinusitis." in: Journal of interferon & cytokine research : the official journal of the International Society for Interferon
Publications Product cited in:	Kouzoli Motoumoto Koto Toiinoo Chinaizu Chinaizu "Faithalial Call Darius d Catalinas
Expiry Date:	6 months
	conditions. In addition, please return the unused wells to the foil pouch containing the desiccant and seal the foil pouch with the zipper.
Storage Comment:	 For unopened kit: All reagents should be stored according to the labels on the vials. The Standard, Detection Reagent A, Detection Reagent B, and 96-well Strip Plate should be stored at -20 °C upon receipt, while the other reagents should be stored at 4 °C. For opened kits: the remaining reagents must be stored according to the above storage
Storage:	4 °C/-20 °C
Precaution of Use:	The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material.
Handling	
Restrictions:	For Research Use only
	Inter-Assay: CV < 12%
	CV(%) = SD/meanX100 Intra-Assay: CV < 10%
	target were tested on 3 different plates, 8 replicates in each plate. CV(%) = SD(magnV100)
	Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level of
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level of target were tested 20 times on one plate, respectively.
	 determinecompatibility with the kit. If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a possibility of causing a deviation due to the introduced chemical substance. The recommended dilution factor is for reference only. Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2).

and Cytokine Research, Vol. 36, Issue 3, pp. 169-79, (2016) (PubMed).

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/5 | Product datasheet for ABIN6954858 | 09/10/2023 | Copyright antibodies-online. All rights reserved. Rohan, Smetakova, Kukal, Rusina, Matej: "Proteinase-activated receptor 2 and disease biomarkers in cerebrospinal fluid in cases with autopsy-confirmed prion diseases and other neurodegenerative diseases." in: **BMC neurology**, Vol. 15, pp. 50, (2015) (PubMed).


Viodé, Lejeune, Turlier, Rouquier, Casas, Mengeaud, Redoulès, Schmitt: "Cathepsin S, a new pruritus biomarker in clinical dandruff/seborrhoeic dermatitis evaluation." in: **Experimental dermatology**, Vol. 23, Issue 4, pp. 274-5, (2014) (PubMed).

Mat?j, Sm?táková, Vašáková, Nováková, Sterclová, Kukal, Olejár: "PAR-2, IL-4R, TGF-? and TNF-? in bronchoalveolar lavage distinguishes extrinsic allergic alveolitis from sarcoidosis." in: **Experimental and therapeutic medicine**, Vol. 8, Issue 2, pp. 533-538, (2014) (PubMed).

Vasakova, Sterclova, Stranska, Mandakova, Skibova, Matej: "Biomarkers of fibroproliferative healing in fibrosing idiopathic interstitial pneumonias." in: **The open respiratory medicine journal**, Vol. 6, pp. 160-4, (2013) (PubMed).

There are more publications referencing this product on: Product page

Images



ELISA

Image 1. Typical standard curve

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 5/5 | Product datasheet for ABIN6954858 | 09/10/2023 | Copyright antibodies-online. All rights reserved.