# antibodies -online.com







# **IDO2 ELISA Kit**





1//(			

OVERVIEW	
Quantity:	96 tests
Target:	IDO2
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	31.25 pg/mL - 2000 pg/mL
Minimum Detection Limit:	31.25 pg/mL
Application:	ELISA
Product Details	
Purpose:	The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of IDO2 in
	mouse serum, plasma, tissue homogenates.
Sample Type:	Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Indoleamine-2,3-Diox
	ygenase 2 (IDO2)
Sensitivity:	12.9 pg/mL
Components:	Pre-coated, ready to use 96-well strip plate, flat buttom
	Plate sealer for 96 wells  Deference Standard
	<ul><li>Reference Standard</li><li>Standard Diluent</li></ul>
	· Standard Diluent

- · Detection Reagent A
- · Detection Reagent B
- · Assay Diluent A
- · Assay Diluent B
- Reagent Diluent (if Detection Reagent is lyophilized)
- · TMB Substrate
- · Stop Solution
- Wash Buffer (30 x concentrate)
- · Instruction manual

# **Target Details**

Target:	IDO2		
Alternative Name:	Indoleamine-2,3-Dioxygenase 2 (IDO2) (IDO2 Products)		
Application Details			
Comment:	Information on standard material:		
	The standard might be recombinant protein or natural protein, that will depend on the specific		
	kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin		
	300 in the standard as preservative.		
	Information on reagents:		
	The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash		
	solution is TBS. The standard diluent contains 0.02 % sodium azide, assay diluent A and assay		
	diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.		
	Information on antihodica:		
	Information on antibodies:		
	The provided antibodies and their host vary in different kits.		
Sample Volume:	100 μL		
Assay Time:	3 h		
Plate:	Pre-coated		
Protocol:	1. Prepare all reagents, samples and standards,		
	2. Add 100µL standard or sample to each well. Incubate 1 hours at 37 °C,		
	3. Aspirate and add 100µL prepared Detection Reagent A. Incubate 1 hour at 37 °C,		
	4. Aspirate and wash 3 times,		
	5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,		

- 6. Aspirate and wash 5 times,
- 7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,
- 8. Add 50µL Stop Solution. Read at 450nm immediately.

### Reagent Preparation:

- 1. Bring all kit components and samples to room temperature (18-25 °C) before use.
- 2. Standard Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes at room temperature, shake gently (not to foam). The concentration of the standard in the stock solution is 8,000pg/mL. Firstly dilute the stock solution to 2,000pg/mL and the diluted standard serves as the highest standard (2,000pg/mL). Then prepare 7 tubes containing 0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as 2,000pg/mL, 1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, and the last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.
- 3. Detection Reagent A and Detection Reagent B Briefly spin or centrifuge the stock Detection A and Detection B before use. Dilute to the working concentration with Assay Diluent A and B, respectively (1:100).
- 4. Wash Solution Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized or distilled water to prepare 600 mL of Wash Solution (1x).
- 5. TMB substrate Aspirate the needed dosage of the solution with sterilized tips and do not dump the residual solution into the vial again.


### Note:

- 1. Making serial dilution in the wells directly is not permitted.
- 2. Prepare standard within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.
- 3. Please carefully reconstitute Standards or working Detection Reagent A and B according to the instruction, and avoid foaming and mix gently until the crystals are completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10µL for once pipetting.
- 4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only once.
- 5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature and mix gently until the crystals are completely dissolved.
- 6. Contaminated water or container for reagent preparation will influence the detection result.

### Sample Preparation:

- It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates.
- If the sample type is not specified in the instructions, a preliminary test is necessary to determine compatibility with the kit.
- If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a possibility of causing a deviation due to the introduced chemical substance. The

recommended dilution factor is for reference only.  • Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2).	
Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level of	
target were tested 20 times on one plate, respectively.	
Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level of	
target were tested on 3 different plates, 8 replicates in each plate.	
CV(%) = SD/meanX100	
Intra-Assay: CV < 10%	
Inter-Assay: CV < 12%	
For Research Use only	
The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and	
clothing protection when using this material.	
4 °C/-20 °C	
1. For unopened kit: All reagents should be stored according to the labels on the vials. The Standard, Detection Reagent A, Detection Reagent B, and 96-well Strip Plate should be stored at -20 °C upon receipt, while the other reagents should be stored at 4 °C.	
<ol><li>For opened kits: the remaining reagents must be stored according to the above storage conditions. In addition, please return the unused wells to the foil pouch containing the desiccant and seal the foil pouch with the zipper.</li></ol>	
6 months	



## **ELISA**

Image 1. Typical standard curve