.-online.com antibodies

Datasheet for ABIN6957017 IL12B ELISA Kit

1 Image

1 Publication



Overview

Quantity:	96 tests
Target:	IL12B
Reactivity:	Rat
Method Type:	Sandwich ELISA
Detection Range:	15.6 pg/mL - 1000 pg/mL
Minimum Detection Limit:	15.6 pg/mL
Application:	ELISA
Product Details	
Purpose:	The kit is a sandwich enzyme immunoassay for in vitro quantitative measurement of IL12B in rat serum, plasma, tissue homogenates, cell lysates, cell culture supernates.
Sample Type:	Cell Culture Supernatant, Cell Lysate, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of Interleukin 12B (IL12B)
Sensitivity:	5.6 pg/mL
Components:	 Pre-coated, ready to use 96-well strip plate, flat buttom Plate sealer for 96 wells Reference Standard Standard Diluent

• Detection Reagent A

- Detection Reagent B
- Assay Diluent A
- Assay Diluent B
- Reagent Diluent (if Detection Reagent is lyophilized)
- TMB Substrate
- Stop Solution
- Wash Buffer (30 x concentrate)
- Instruction manual

Target Details

IL12B
Interleukin 12B (IL12B) (IL12B Products)
JAK-STAT Signaling, Cellular Response to Molecule of Bacterial Origin, Regulation of Leukocyte Mediated Immunity, Positive Regulation of Immune Effector Process, Activated T Cell Proliferation
Information on standard material:
The standard might be recombinant protein or natural protein, that will depend on the specific
kit. Moreover, the expression system is E.coli or yeast or mammal cell. There is 0.05% proclin
300 in the standard as preservative.
Information on reagents:
The stop solution used in the kit is sulfuric acid with concentration of 1 mol/L. And the wash
solution is TBS. The standard diluent contains 0.02 % sodium azide, assay diluent A and assay
diluent B contain 0.01% sodium azide. Some kits can contain is BSA in them.
Information on antibodies:
The provided antibodies and their host vary in different kits.
100 μL
3 h
Pre-coated

Protocol: 1. Prepare all reagents, samples and standards,

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/5 | Product datasheet for ABIN6957017 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

	2. Add 100 μL standard or sample to each well. Incubate 1 hours at 37 °C,
	3. Aspirate and add 100 μ L prepared Detection Reagent A. Incubate 1 hour at 37 °C,
	4. Aspirate and wash 3 times,
	5. Add 100µL prepared Detection Reagent B. Incubate 30 minutes at 37 °C,
	6. Aspirate and wash 5 times,
	7. Add 90µL Substrate Solution. Incubate 10-20 minutes at 37 °C,
	8. Add 50µL Stop Solution. Read at 450nm immediately.
Reagent Preparation:	1. Bring all kit components and samples to room temperature (18-25 °C) before use. If the kit
	will not be used up in one time, please only take out strips and reagents for present
	experiment, and leave the remaining strips and reagents in required condition.
	2. Standard - Reconstitute the Standard with 1.0 mL of Standard Diluent, keep for 10 minutes a
	room temperature, shake gently (not to foam). The concentration of the standard in the stocl
	solution is 2,000pg/mL. Firstly dilute the stock solution to 1,000pg/mL and the diluted
	standard serves as the highest standard (1,000pg/mL). Then prepare 7 tubes containing
	0.5 mL Standard Diluent and use the diluted standard to produce a double dilution series. Mix
	each tube thoroughly before the next transfer. Set up 7 points of diluted standard such as
	1,000pg/mL, 500pg/mL, 250pg/mL, 125pg/mL, 62.5pg/mL, 31.2pg/mL, 15.6pg/mL, and the
	last microcentrifuge tube with Standard Diluent is the blank as 0pg/mL.
	3. Detection Reagent A and Detection Reagent B - If lyophilized reconstitute the Detection
	Reagent A with 150µL of Reagent Diluent, keep for 10 minutes at room temperature, shake
	gently (not to foam). Briefly spin or centrifuge the stock Detection A and Detection B before
	use. Dilute them to the working concentration 100-fold with Assay Diluent A and B,
	respectively.
	4. Wash Solution - Dilute 20 mL of Wash Solution concentrate (30x) with 580 mL of deionized
	or distilled water to prepare 600 mL of Wash Solution (1x).
	5. TMB substrate - Aspirate the needed dosage of the solution with sterilized tips and do not
	dump the residual solution into the vial again.
	Note:
	1. Making serial dilution in the wells directly is not permitted.
	 Prepare standards within 15 minutes before assay. Please do not dissolve the reagents at 37 °C directly.
	3. Please carefully reconstitute Standards or working Detection Reagent A and B according to
	the instruction, and avoid foaming and mix gently until the crystals are completely dissolved.
	To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors
	are calibrated. It is recommended to suck more than $10\mu L$ for one pipetting.
	4. The reconstituted Standards, Detection Reagent A and Detection Reagent B can be used only
	once.
	5. If crystals have formed in the Wash Solution concentrate (30x), warm to room temperature
	and mix gently until the crystals are completely dissolved.
	6. Contaminated water or container for reagent preparation will influence the detection result.
Sample Preparation:	• It is recommended to use fresh samples without long storage, otherwise protein degradation


Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 3/5 | Product datasheet for ABIN6957017 | 09/10/2023 | Copyright antibodies-online. All rights reserved.

Product cited in:	Klaver, Coffey, Aasly, Loeffler: "CSF lamp2 concentrations are decreased in female Parkinson's
Publications	
Expiry Date:	6 months
	 For opened kits: the remaining reagents must be stored according to the above storage conditions. In addition, please return the unused wells to the foil pouch containing the desiccant and seal the foil pouch with the zipper.
Storage Comment:	 For unopened kit: All reagents should be stored according to the labels on the vials. The Standard, Detection Reagent A, Detection Reagent B, and 96-well Strip Plate should be stored at -20 °C upon receipt, while the other reagents should be stored at 4 °C.
Storage:	4 °C/-20 °C
Precaution of Use:	The Stop Solution suggested for use with this kit is an acid solution. Wear eye, hand, face, and clothing protection when using this material.
Handling	
Restrictions:	For Research Use only
	Inter-Assay: CV < 12%
	Intra-Assay: CV < 10%
	CV(%) = SD/meanX100
	target were tested on 3 different plates, 8 replicates in each plate.
	target were tested 20 times on one plate, respectively. Inter-assay Precision (Precision between assays): 3 samples with low, middle and high level of
Assay Precision:	Intra-assay Precision (Precision within an assay): 3 samples with low, middle and high level of
	 therefore be stored for a short periodat 2 - 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates. If the sample type is not specified in the instructions, a preliminary test is necessary to determinecompatibility with the kit. If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a possibility of causing a deviation due to the introduced chemical substance. The recommended dilution factor is for reference only. Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2).

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/5 | Product datasheet for ABIN6957017 | 09/10/2023 | Copyright antibodies-online. All rights reserved. disease patients with LRRK2 mutations." in: **Brain research**, Vol. 1683, pp. 12-16, (2018) (PubMed).

Loeffler, Klaver, Coffey, Aasly: "Cerebrospinal Fluid Concentration of Key Autophagy Protein Lamp2 Changes Little During Normal Aging." in: **Frontiers in aging neuroscience**, Vol. 10, pp. 130, (2018) (PubMed).

Images



ELISA

Image 1. Typical standard curve

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 5/5 | Product datasheet for ABIN6957017 | 09/10/2023 | Copyright antibodies-online. All rights reserved.