antibodies - online.com





Datasheet for ABIN6963005

CD14 ELISA Kit





_			
	N/0	r\/I	ew
\sim	' v C	1 V I	C V V

Quantity:	96 tests	
Target:	CD14	
Reactivity:	Mouse	
Method Type:	Sandwich ELISA	
Detection Range:	1.56 ng/mL - 100 ng/mL	
Minimum Detection Limit:	1.56 ng/mL	
Application:	ELISA	
Product Details		
Purpose:	The kit is a sandwich enzyme immunoassay technique for the in vitro quantitative	
	measurement in various sample types.	
Sample Type:	Cell Culture Supernatant, Plasma, Serum	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This kit recognizes Mouse sCD14 in samples. No Significant cross-reactivity or interference	
	between Mouse sCD14 and analogues was observed.	
Sensitivity:	0.94 ng/mL	
Components:	Pre-coated, ready to use 96-well strip plate, flat buttom	
	Plate sealer for 96 wells	
	Reference Standard	

• Reference Standard & Sample Diluent

• Biotinylated Detection Antibody (100 x concentrate)

Soluble Cluster of Differentiation 14 (CD14 Products)

- HRP Conjugate (100 x concentrate)
- · Biotinylated Detection Antibody Diluent
- HRP Conjugate Diluent
- · Substrate Reagent
- · Stop Solution

CD14

- Wash Buffer (25 x concentrate)
- · Instruction manual

Target Details

Alternative Name:

Target:

Background:	CD14 Antigen, Monocyte Differentiation Antigen CD14		
Pathways:	TLR Signaling, Activation of Innate immune Response, Cellular Response to Molecule of		
	Bacterial Origin, Toll-Like Receptors Cascades		
Application Details			
Sample Volume:	100 μL		
Assay Time:	3.5 h		
Plate:	Pre-coated		
Protocol:	1. Add 100 µL standard or sample to each well. Incubate for 90 min at 37 °C.		
	2. Remove the liquid. Add 100 μL Biotinylated Detection Antibody. Incubate for 1 hour at 37 $^{\circ}C$		
	3. Aspirate and wash 3 times.		
	4. Add 100 μL HRP Conjugate. Incubate for 30 min at 37 °C.		
	5. Aspirate and wash 5 times.		
	6. Add 90 μ L Substrate Reagent. Incubate for 15 min at 37 °C.		
	7. Add 50 µL Stop Solution. Read at 450 nm immediately.		
	8. Calculation of results.		
Reagent Preparation:	1. Bring all reagents to room temperature (18~25 °C) before use. Follow the Microplate reade manual for set-up and preheat it for 15 min before OD measurement.		
	2. Wash Buffer: Dilute 30 mL of Concentrated Wash Buffer with 720 mL of deionized or distilled water to prepare 750 mL of Wash Buffer. Note: if crystals have formed in the concentrate, warm it in a 40 °C water bath and mix it gently until the crystals have completely dissolved		
	3. Standard working solution: Centrifuge the standard at 10,000xg for 1 min. Add 1.0 mL of Reference Standard &Sample Diluent, let it stand for 10 min and invert it gently several time.		

After it dissolves fully, mix it thoroughly with a pipette. This reconstitution produces a working solution of 100 ng/mL. Then make serial dilutions as needed. The recommended dilution gradient is as follows: 100, 50, 25, 12.5, 6.25, 3.13, 1.56, 0 ng/mL. Dilution method: Take 7 EP tubes, add 500 μ Lof Reference Standard & Sample Diluent to each tube. Pipette 500 μ Lof the 100 ng/mL working solution to the first tube and mix up to produce a 50 ng/mL working solution. Pipette 500 μ Lof the solution from the former tube into the latter one according to these steps. The illustration below is for reference. Note: the last tube is regarded as a blank. Don't pipette solution into it from the former tube.

- 4. Biotinylated Detection Antibody working solution: Calculate the required amount before the experiment (100 µL/well). In preparation, slightly more than calculated should be prepared. Centrifuge the stock tube before use, dilute the 100x Concentrated Biotinylated Detection Antibody to 1xworking solution with Biotinylated Detection Antibody Diluent.
- 5. Concentrated HRP Conjugate working solution: Calculate the required amount before the experiment (100 μL/well). In preparation, slightly more than calculated should be prepared. Dilute the 100x Concentrated HRP Conjugate to 1x working solution with Concentrated HRP Conjugate Diluent.

Sample Preparation:

- It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates.
- If the sample type is not specified in the instructions, a preliminary test is necessary to determine compatibility with the kit.
- If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a
 possibility of causing a deviation due to the introduced chemical substance. The
 recommended dilution factor is for reference only.
- Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2).

Assay Precision:

Intra-assay Precision (Precision within an assay): 3 samples with low, mid range and high level Mouse sCD14 were tested 20 times on one plate, respectively.

Inter-assay Precision (Precision between assays): 3 samples with low, mid range and high level Mouse sCD14 were tested on 3 different plates, 20 replicates in each plate.

Both intra-CV and inter-CV are < 10 %.

Restrictions:

For Research Use only

Handling

Storage: 4 °C,-20 °C

Handling


Storage Comment:

- 1. For unopened kit: All reagents should be stored according to the labels on the vials, so they are stable up to 6 months after receipt of the kit. The Reference Standard, Biotinylated Detection Antibody, HRP Conjugate and the 96-well stripe plate should be stored at -20 °C upon receipt while the other reagents should be stored at 4 °C.
- 2. For used kit: When the kit is used, the remaining reagents need to be stored according to the above storage condition. Besides, please return the unused wells to the foil pouch containing the desiccant pack, and zip-seal the foil pouch.

Expiry Date:

6 months

Images



ELISA

Image 1. Typical standard curve