antibodies -online.com







PTGDS ELISA Kit





Overview	
Quantity:	96 tests
Target:	PTGDS
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	0.78 ng/mL - 50 ng/mL
Minimum Detection Limit:	0.78 ng/mL
Application:	ELISA
Product Details	
Purpose:	For the quantitative determination of human prostaglandin-H2 D-isomerase (PTGDS)
	concentrations in corum, placema, cell cultura supernates, tiesus homogenates

Purpose:	For the quantitative determination of human prostaglandin-H2 D-isomerase (PTGDS) concentrations in serum, plasma, cell culture supernates, tissue homogenates.
Sample Type:	Cell Culture Supernatant, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of human PTGDS. No significant cross-reactivity or interference between human PTGDS and analogues was observed. Note: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between human PTGDS and all the analogues, therefore, cross reaction may still exist.
Sensitivity:	0.2 ng/mL

Product Details

Components:

- Assay plate
- Standard
- HRP-avidin (100 x concentrate)
- Biotin-antibody (100 x concentrate)
- · Sample Diluent
- · HRP-avidin Diluent
- · Biotin-antibody Diluent
- Wash Buffer (25 x concentrate)
- · TMB Substrate
- · Stop Solution
- · Adhesive Strip

Target Details

Target:	PTGDS
Alternative Name:	prostaglandin D2 synthase 21kDa (brain) (PTGDS Products)
Background:	Abbreviation: PTGDS Alias: RP11-229P13.6, L-PGDS, LPGDS, PDS, PGD2, PGDS, PGDS2, PGD2 synthase beta-trace protein g utathione-independent PGD synthase lipocalin-type prostaglandin D
UniProt:	synthase prostaglandin D synthase prostagla P41222

Application Details

Application Notes:	Optimal working dilution should be determined by the investigator.
Sample Volume:	100 μL
Assay Time:	1 - 4.5 h
Plate:	Pre-coated
Protocol:	1. Prepare reagents, samples and standards as instructed.
	2. Add 100 μ L standard or sample to each well. Incubate 2 hours at 37 °C.
	3. Remove the liquid of each well, don't wash.
	4. Add 100 µL Biotin-antibody (1x) to each well. Incubate 1 hour at 37 °C.
	5. Aspirate and wash 3 times.
	6. Add 100 µL HRP-avidin (1x) to each well. Incubate 1 hour at 37 °C
	7. Aspirate and wash 5 times.
	8. Add 90 µL of TMB Substrate to each well. Incubate for 15-30 minutes at 37 °C. Protect from
	light.

9. Add 50 µL Stop Solution to each well. Read at 450 nm within 5 minutes.

Reagent Preparation:

- 1. Biotin-antibody (1x) Centrifuge the vial before opening. Biotin-antibody requires a 100-fold dilution. A suggested 100-fold dilution is 10 μ L of Biotin-antibody + 990 μ L of Biotin-antibody Diluent.
- 2. HRP-avidin (1x) Centrifuge the vial before opening. HRP-avidin requires a 100-fold dilution. A suggested 100-fold dilution is 10 μ L of HRP-avidin + 990 μ L of HRP-avidin Diluent.
- 3. Wash Buffer (1x) If crystals have formed in the concentrate, warm up to room temperature and mix gently until the crystals have completely dissolved. Dilute 20 mL of Wash Buffer Concentrate (25 x) into deionized or distilled water to prepare 500 mL of Wash Buffer (1 x).
- 4. Standard Centrifuge the standard vial at 6000-10000rpm for 30s. Reconstitute the Standard with 1.0 mL of Sample Diluent. Do not substitute other diluents. This reconstitution produces a stock solution of 50 ng/mL. Mix the standard to ensure complete reconstitution and allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 μL of Sample Diluent into each tube (S0-S6). Use the stock solution to produce a 2-fold dilution series (below). Mix each tube thoroughly before the next transfer. The undiluted Standard serves as the high standard (50 ng/mL). Sample Diluent serves as the zero standard (0 ng/mL).


Note:

- Kindly use graduated containers to prepare the reagent. Please don't prepare the reagent directly in the Diluent vials provided in the kit.
- Bring all reagents to room temperature (18-25 °C) before use for 30 min.
- Prepare fresh standard for each assay. Use within 4 hours and discard after use.
- · Making serial dilution in the wells directly is not permitted.
- Please carefully reconstitute Standards according to the instruction, and avoid foaming and mix gently until the crystals have completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10 µL for once pipetting.
- Distilled water is recommended to be used to make the preparation for reagents.
 Contaminated water or container for reagent preparation will influence the detection result.

Sample Preparation:

- It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates.
- If the sample type is not specified in the instructions, a preliminary test is necessary to determine compatibility with the kit.
- If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a
 possibility of causing a deviation due to the introduced chemical substance. The
 recommended dilution factor is for reference only.
- Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular

experiment has to be determined. Samples should then be diluted with PBS Note: Recommend to dilute the serum or plasma samples with Sample Diluent (1:1) The suggested 100-fold dilution can be achieved by adding 10 µL sample to 4 Diluent. Complete the 100-fold dilution by adding 15 µL of this solution to 285 Diluent. The recommended dilution factor is for reference only. The optimal dishould be determined by users according to their particular experiments. Assay Precision: Intra-assay Precision (Precision within an assay): CV%<8% Three samples of a concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of a concentration were tested in twenty assays to assess. Restrictions: For Research Use only Handling Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. Mup to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Biotim make recent use, better keep it store at HRP-avidin -20°C. Biotin-antibody Diluent.	
Recommend to dilute the serum or plasma samples with Sample Diluent (1:19). The suggested 100-fold dilution can be achieved by adding 10 µL sample to 4 Diluent. Complete the 100-fold dilution by adding 15 µL of this solution to 285 Diluent. The recommended dilution factor is for reference only. The optimal dishould be determined by users according to their particular experiments. Assay Precision: Intra-assay Precision (Precision within an assay): CV%<8% Three samples of Concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of Concentration were tested in twenty assays to assess. Restrictions: For Research Use only Handling Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. Now up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Bioting and the damp is the samples of the scale of the samples	S (pH =7.0-7.2).
The suggested 100-fold dilution can be achieved by adding 10 µL sample to 4 Diluent. Complete the 100-fold dilution by adding 15 µL of this solution to 285 Diluent. The recommended dilution factor is for reference only. The optimal dishould be determined by users according to their particular experiments. Assay Precision: Intra-assay Precision (Precision within an assay): CV%<8% Three samples of 1 concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of concentration were tested in twenty assays to assess. Restrictions: For Research Use only Handling Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. Mup to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Biotin	
Diluent. Complete the 100-fold dilution by adding 15 µL of this solution to 285 Diluent. The recommended dilution factor is for reference only. The optimal di should be determined by users according to their particular experiments. Assay Precision: Intra-assay Precision (Precision within an assay): CV%<8% Three samples of 1 concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of concentration were tested in twenty assays to assess. Restrictions: For Research Use only Handling Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. M up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Biotin	100) before test.
Diluent. The recommended dilution factor is for reference only. The optimal dishould be determined by users according to their particular experiments. Assay Precision: Intra-assay Precision (Precision within an assay): CV%<8% Three samples of I concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of concentration were tested in twenty assays to assess. Restrictions: For Research Use only Handling Storage: 4 °C,-20 °C Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. Mup to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Biotin	40 μL of Sample
should be determined by users according to their particular experiments. Assay Precision: Intra-assay Precision (Precision within an assay): CV%<8% Three samples of a concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of a concentration were tested in twenty assays to assess. Restrictions: For Research Use only Handling Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. No up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Biotin	5 μL of Sample
Assay Precision: Intra-assay Precision (Precision within an assay): CV%<8% Three samples of I concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of concentration were tested in twenty assays to assess. Restrictions: For Research Use only Handling Storage: 4 °C,-20 °C Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. Mup to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Biotin	dilution factor
concentration were tested twenty times on one plate to assess. Inter-assay Precision (Precision between assays): CV%<10% Three samples of concentration were tested in twenty assays to assess. Restrictions: For Research Use only Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. Mup to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Biotin	
Inter-assay Precision (Precision between assays): CV%<10% Three samples of concentration were tested in twenty assays to assess. Restrictions: For Research Use only Handling Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. Mup to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Biotin	f known
concentration were tested in twenty assays to assess. Restrictions: For Research Use only Handling Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8 °C. Do not use the kit beyond the expiration date. No up to 1 month at 2 - 8 °C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8 °C. If Biotin	
Restrictions: For Research Use only Handling Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. No up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Bioting	of known
Handling Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8 °C. Do not use the kit beyond the expiration date. Mup to 1 month at 2 - 8 °C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8 °C. If Biotir	
Storage: 4 °C,-20 °C Storage Comment: Unopened kit Store at 2 - 8 °C. Do not use the kit beyond the expiration date. Mup to 1 month at 2 - 8 °C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8 °C. If Biotir	
Storage Comment: Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. No up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If Biotir	
up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foi avoid the damp. Standard May be stored for up to 1 month at 2 - 8° C. If Biotir	
avoid the damp. Standard May be stored for up to 1 month at 2 - 8° C. If Biotir	May be stored for
	oil bag, plate and
make recent use, better keep it store at HRP-avidin -20°C. Biotin-antibody Dilu	in-antibody don't
	uent Opened kit
HRP-avidin Diluent Sample May be stored for up to 1 month at 2 - 8°C. Diluent	nt Wash Buffer
TMB Substrate Stop Solution *Provided this is within the expiration date of the	ne kit.
Expiry Date: 6 months	



ELISA

Image 1. Typical Standard Curve