# antibodies - online.com







# Datasheet for ABIN6974451

# **Folic Acid ELISA Kit**



Image



( )	ve	K\ /		A .
	$\cup$	1 V/	Щ.	V۷

Quantity:	96 tests	
Target:	Folic Acid (FA)	
Reactivity:	Human	
Method Type:	Competition ELISA	
Detection Range:	12.5 pg/mL - 200 pg/mL	
Minimum Detection Limit:	12.5 pg/mL	
Application:	ELISA	
Product Details		
Purpose:	For the quantitative determination of endogenic human folic acid (FA) concentrations in serum, plasma.	
Sample Type:	Plasma, Serum	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This assay has high sensitivity and excellent specificity for detection of human FA. No significant cross-reactivity or interference between human FA and analogues was observed.  Note: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between human FA and all the analogues, therefore, cross reaction may still exist.	
Sensitivity:	6.25 pg/mL	

### **Product Details**

#### Components:

- Assay plate
- Standard
- HRP-conjugate (100 x concentrate)
- · Sample Diluent
- HRP-conjugate Diluent
- Wash Buffer (25 x concentrate)
- · TMB Substrate
- · Stop Solution
- · Adhesive Strip
- · Stop Solution
- Adhesive Strip

Folic Acid (FA)

Folic acid(FA) (FA Products)

## **Target Details**

Alternative Name:

Target:

Target Type:	Chemical	
Background:	Abbreviation: FA	
	Alias: N/A	
Application Details		
Application Notes:	Optimal working dilution should be determined by the investigator.	
Sample Volume:	50 μL	
Assay Time:	1 - 4.5 h	
Plate:	Pre-coated	
Protocol:	Prepare reagents, samples and standards as instructed.	
	2. Set a Blank well without any solution.	
	3. Add 50 µL standard or sample to each well.	
	4. Add 50 μL HRP-conjugate (1x) to each well (Not to Blank well).	
	5. Incubate 1 hour at 37 °C	
	6. Aspirate and wash 5 times.	
	7. Add 90 µL of TMB Substrate to each well. Incubate for 20 minutes at 37 °C. Protect from	
	light.	
	8. Add 50 µL Stop Solution to each well. Read at 450 nm within 5 minutes.	
Reagent Preparation:	1. HRP-conjugate (1x) - Centrifuge the vial before opening. HRP-conjugate requires a 100-fold	

- dilution. A suggested 100-fold dilution is 10  $\mu$ L of HRP-conjugate + 990  $\mu$ L of HRP-conjugate Diluent.
- 2. Wash Buffer(1x)- If crystals have formed in the concentrate, warm up to room temperature and mix gently until the crystals have completely dissolved. Dilute 20 mL of Wash Buffer Concentrate (25 x) into deionized or distilled water to prepare 500 mL of Wash Buffer (1 x).
- 3. Standard Centrifuge the standard vial at 6000-10000rpm for 30s before opening. Dilute the Standard(10x) with Sample Diluent. A suggested 10-fold dilution is 30  $\mu$ L of Standard(10x) + 270  $\mu$ L of Sample Diluent. This diluted Standard (S5) serves as the high standard (200 pg/mL). Do not substitute other diluents. Mix the standard to ensure complete dilution and allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 150  $\mu$ L of Sample Diluent into each tube (S0-S4). Use the diluted Standard (S5) solution to produce a 2-fold dilution series (below). Mix each tube thoroughly before the next transfer. Sample Diluent serves as the zero standard (0 pg/mL).

#### Note:

- Kindly use graduated containers to prepare the reagent. Please don't prepare the reagent directly in the Diluent vials provided in the kit.
- Bring all reagents to room temperature (18-25 °C) before use for 30 min.
- Prepare fresh standard for each assay. Use within 4 hours and discard after use.
- · Making serial dilution in the wells directly is not permitted.
- Distilled water is recommended to be used to make the preparation for reagents.
   Contaminated water or container for reagent preparation will influence the detection result.

#### Sample Preparation:

- It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates.
- If the sample type is not specified in the instructions, a preliminary test is necessary to determine compatibility with the kit.
- If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a
  possibility of causing a deviation due to the introduced chemical substance. The
  recommended dilution factor is for reference only.
- Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2).

#### Note:

Recommend to dilute the serum or plasma samples with Sample Diluent(1:100) before test. The suggested 100-fold dilution can be achieved by adding 10  $\mu$ L sample to 40  $\mu$ L of Sample Diluent. Complete the 100-fold dilution by adding 15  $\mu$ L of this solution to 285  $\mu$ L of Sample Diluent. The recommended dilution factor is for reference only. The optimal dilution factor should be determined by users according to their particular experiments. 6


# **Application Details**

Assay Precision:	Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known
	concentration were tested twenty times on one plate to assess.
	Inter-assay Precision (Precision between assays): CV%<10% Three samples of known
	concentration were tested in twenty assays to assess.
Restrictions:	For Research Use only

# Handling

Storage:	4 °C,-20 °C
Storage Comment:	Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. May be stored for up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foil bag, plate and avoid the damp. Standard May be stored for up to 1 month at 2 - 8°C. If don't make recent use, better keep it store at HRP-conjugate -20°C. Opened kit HRP-conjugate Diluent Sample Diluent May be stored for up to 1 month at 2 - 8°C. Wash Buffer TMB Substrate Stop Solution *Provided this is within the expiration date of the kit.
Expiry Date:	6 months

## **Images**



## **ELISA**

Image 1. Typical Standard Curve