# .-online.com antibodies

# Datasheet for ABIN6975229 Estrogen Receptor alpha ELISA Kit

1 Image

1 Publication



#### Overview

Quantity:	96 tests
Target:	Estrogen Receptor alpha (ESR1)
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	3.12 pg/mL - 200 pg/mL
Minimum Detection Limit:	3.12 pg/mL
Application:	ELISA

### Product Details

Purpose:	For the quantitative determination of human estradiol receptor (ER) concentrations in serum, plasma, tissue homogenates, cell lysates.
Sample Type:	Cell Lysate, Plasma, Serum, Tissue Homogenate
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	This assay has high sensitivity and excellent specificity for detection of human ER. No significant cross-reactivity or interference between human ER and analogues was observed. Note: Limited by current skills and knowledge, it is impossible for us to complete the cross-reactivity detection between human ER and all the analogues, therefore, cross reaction may still exist.
Sensitivity:	0.78 pg/mL

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/5 | Product datasheet for ABIN6975229 | 01/16/2024 | Copyright antibodies-online. All rights reserved.

#### Product Details

#### Components:

- Assay plate
- Standard
- HRP-avidin (100 x concentrate)
- Biotin-antibody (100 x concentrate)
- Sample Diluent
- HRP-avidin Diluent
- Biotin-antibody Diluent
- Wash Buffer (25 x concentrate)
- TMB Substrate
- Stop Solution
- Adhesive Strip

## Target Details

5	
Target:	Estrogen Receptor alpha (ESR1)
Alternative Name:	estrogen receptor 1 (ESR1 Products)
Background:	Abbreviation: ESR1
	Alias: RP1-130E4.1, DKFZp686N23123, ER, ESR, ESRA, Era, NR3A1, estrogen receptor
	alpha estrogen receptor alpha delta 3*,4,5,6,7*/819-2 isoform estrogen receptor alpha delta 4
	+49 isoform estrogen receptor
UniProt:	P03372
Pathways:	Nuclear Receptor Transcription Pathway, EGFR Signaling Pathway, Retinoic Acid Receptor
	Signaling Pathway, Intracellular Steroid Hormone Receptor Signaling Pathway, Steroid
	Hormone Mediated Signaling Pathway, Ribonucleoprotein Complex Subunit Organization,
	Ribosome Assembly

#### Application Details

Application Notes:	Optimal working dilution should be determined by the investigator.
Sample Volume:	100 µL
Assay Time:	1 - 4.5 h
Plate:	Pre-coated
Protocol:	<ol> <li>Prepare reagents, samples and standards as instructed.</li> <li>Add 100 μL standard or sample to each well. Incubate 2 hours at 37 °C.</li> <li>Remove the liquid of each well, don't wash.</li> <li>Add 100 μL Biotin-antibody (1x) to each well. Incubate 1 hour at 37 °C.</li> </ol>

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/5 | Product datasheet for ABIN6975229 | 01/16/2024 | Copyright antibodies-online. All rights reserved.


	5. Aspirate and wash 3 times.
	6. Add 100 $\mu$ L HRP-avidin (1x) to each well. Incubate 1 hour at 37 °C
	7. Aspirate and wash 5 times.
	8. Add 90 $\mu L$ of TMB Substrate to each well. Incubate for 15-30 minutes at 37 °C. Protect from
	light.
	9. Add 50 $\mu L$ Stop Solution to each well. Read at 450 nm within 5 minutes.
Reagent Preparation:	<ol> <li>Biotin-antibody (1x) - Centrifuge the vial before opening. Biotin-antibody requires a 100-fold dilution. A suggested 100-fold dilution is 10 μL of Biotin-antibody + 990 μL of Biotin-antibody Diluent.</li> </ol>
	2. HRP-avidin (1x) - Centrifuge the vial before opening. HRP-avidin requires a 100-fold dilution. A suggested 100-fold dilution is 10 μL of HRP-avidin + 990 μL of HRP-avidin Diluent.
	3. Wash Buffer (1x) - If crystals have formed in the concentrate, warm up to room temperature and mix gently until the crystals have completely dissolved. Dilute 20 mL of Wash Buffer Concentrate (25 x) into deionized or distilled water to prepare 500 mL of Wash Buffer (1 x).
	4. Standard Centrifuge the standard vial at 6000-10000rpm for 30s. Reconstitute the Standard with 1.0 mL of Sample Diluent. Do not substitute other diluents. This reconstitution produces a stock solution of 200 pg/mL. Mix the standard to ensure complete reconstitution and allow the standard to sit for a minimum of 15 minutes with gentle agitation prior to making dilutions. Pipette 250 µL of Sample Diluent into each tube (S0-S6). Use the stock solution to produce a 2-fold dilution series (below). Mix each tube thoroughly before the next transfer. The undiluted Standard serves as the high standard (200 pg/mL). Sample Diluent serves as the zero standard (0 pg/mL).
	Note:
	<ul> <li>Kindly use graduated containers to prepare the reagent. Please don't prepare the reagent directly in the Diluent vials provided in the kit.</li> </ul>
	<ul> <li>Bring all reagents to room temperature (18-25 °C) before use for 30 min.</li> </ul>
	<ul> <li>Prepare fresh standard for each assay. Use within 4 hours and discard after use.</li> </ul>
	<ul> <li>Making serial dilution in the wells directly is not permitted.</li> </ul>
	<ul> <li>Please carefully reconstitute Standards according to the instruction, and avoid foaming and mix gently until the crystals have completely dissolved. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors are calibrated. It is recommended to suck more than 10 µL for once pipetting.</li> </ul>
	• Distilled water is recommended to be used to make the preparation for reagents or samples.
	Contaminated water or container for reagent preparation will influence the detection result.
Sample Preparation:	<ul> <li>It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturationmay occur in these samples, leading to false results. Samples should therefore be stored for a short periodat 2 - 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thawcycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged toremove precipitates.</li> </ul>
	<ul> <li>If the sample type is not specified in the instructions, a preliminary test is necessary to</li> </ul>

• If the sample type is not specified in the instructions, a preliminary test is necessary to determinecompatibility with the kit.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 3/5 | Product datasheet for ABIN6975229 | 01/16/2024 | Copyright antibodies-online. All rights reserved.

	• If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is possibility of causing a deviation due to the introduced chemical substance. The recommended dilution factor is for reference only.
	• Please estimate the concentration of the samples before performing the test. If the values are not in therange of the standard curve, the optimal sample dilution for the particular experiment has to be determined. Samples should then be diluted with PBS (pH =7.0-7.2).
Assay Precision:	Intra-assay Precision (Precision within an assay): CV%<8% Three samples of known
	concentration were tested twenty times on one plate to assess.
	Inter-assay Precision (Precision between assays): CV%<10% Three samples of known
	concentration were tested in twenty assays to assess.
Restrictions:	For Research Use only
Handling	
-	4 °C,-20 °C
Storage:	4 °C,-20 °C Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. May be stored fo
Storage:	Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. May be stored fo
Storage:	
Storage:	Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. May be stored fo up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foil bag, plate and
Storage:	Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. May be stored fo up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foil bag, plate and avoid the damp. Standard May be stored for up to 1 month at 2 - 8° C. If Biotin-antibody don't
Storage:	Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. May be stored for up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foil bag, plate and avoid the damp. Standard May be stored for up to 1 month at 2 - 8° C. If Biotin-antibody don't make recent use, better keep it store at HRP-avidin -20°C. Biotin-antibody Diluent Opened kit
Handling Storage: Storage Comment: Expiry Date:	Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. May be stored fo up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foil bag, plate and avoid the damp. Standard May be stored for up to 1 month at 2 - 8° C. If Biotin-antibody don't make recent use, better keep it store at HRP-avidin -20°C. Biotin-antibody Diluent Opened kit HRP-avidin Diluent Sample May be stored for up to 1 month at 2 - 8°C. Diluent Wash Buffer
Storage: Storage Comment: Expiry Date:	Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. May be stored fo up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foil bag, plate and avoid the damp. Standard May be stored for up to 1 month at 2 - 8° C. If Biotin-antibody don't make recent use, better keep it store at HRP-avidin -20°C. Biotin-antibody Diluent Opened kit HRP-avidin Diluent Sample May be stored for up to 1 month at 2 - 8°C. Diluent Wash Buffer TMB Substrate Stop Solution *Provided this is within the expiration date of the kit.
Storage: Storage Comment:	Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. May be stored fo up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foil bag, plate and avoid the damp. Standard May be stored for up to 1 month at 2 - 8° C. If Biotin-antibody don't make recent use, better keep it store at HRP-avidin -20°C. Biotin-antibody Diluent Opened kit HRP-avidin Diluent Sample May be stored for up to 1 month at 2 - 8°C. Diluent Wash Buffer TMB Substrate Stop Solution *Provided this is within the expiration date of the kit.
Storage: Storage Comment: Expiry Date: Publications	Unopened kit Store at 2 - 8°C. Do not use the kit beyond the expiration date. May be stored for up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foil bag, plate and avoid the damp. Standard May be stored for up to 1 month at 2 - 8° C. If Biotin-antibody don't make recent use, better keep it store at HRP-avidin -20°C. Biotin-antibody Diluent Opened kit HRP-avidin Diluent Sample May be stored for up to 1 month at 2 - 8°C. Diluent Wash Buffer TMB Substrate Stop Solution *Provided this is within the expiration date of the kit. 6 months

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/5 | Product datasheet for ABIN6975229 | 01/16/2024 | Copyright antibodies-online. All rights reserved.



ELISA

Image 1. Typical Standard Curve

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 5/5 | Product datasheet for ABIN6975229 | 01/16/2024 | Copyright antibodies-online. All rights reserved.