

# Datasheet for ABIN6975275

# **Rubisco ELISA Kit**





## Overview

Quantity:	96 tests	
Target:	Rubisco (RuBisCO)	
Reactivity:	Arabidopsis	
Method Type:	Competition ELISA	
Detection Range:	3.12 μg/mL - 800 μg/mL	
Minimum Detection Limit:	3.12 μg/mL	
Application:	ELISA	
Product Details		
Purpose:	For the quantitative determination of endogenic ribulose-1,5-bisphosphate	
	carboxylase/oxygenase (RuBisCO) concentrations in plant tissues.	
Sample Type:	Plant Tissue	
Analytical Method:	Quantitative	
Detection Method:	Colorimetric	
Specificity:	This assay has high sensitivity and excellent specificity for detection of plant RuBisCO. No	
	significant cross-reactivity or interference between plant RuBisCO and analogues was	
	observed. Note: Limited by current skills and knowledge, it is impossible for us to complete the	
	cross-reactivity detection between plant RuBisCO and all the analogues, therefore, cross	
	reaction may still exist.	
Components:	Assay plate	
	Standard	

- HRP-conjugate (100 x concentrate)
- · Sample Diluent
- · HRP-conjugate Diluent
- Wash Buffer (25 x concentrate)
- TMB Substrate
- · Stop Solution
- · Adhesive Strip
- Stop Solution
- · Adhesive Strip

Rubisco (RuBisCO)

## Target Details

Target:

Alternative Name:	Ribulose-1,5-bisphosphate carboxylase/oxygenase(RuBisCO) (RuBisCO Products)		
Background:	RuBisCO		
Application Details			
Application Notes:	Optimal working dilution should be determined by the investigator.		
Sample Volume:	50 μL		
Assay Time:	1 - 4.5 h		
Plate:	Pre-coated		
Protocol:	1. Prepare reagents, samples and standards as instructed.		
	2. Set a Blank well without any solution.		
	3. Add 50 $\mu$ L standard or sample to each well.		
	4. Add 50 μL HRP-conjugate (1x) to each well (Not to Blank well).		
	5. Incubate 1 hour at 37 °C		
	6. Aspirate and wash 5 times.		
	7. Add 90 μL of TMB Substrate to each well. Incubate for 20 minutes at 37 °C. Protect from light.		
	8. Add 50 µL Stop Solution to each well. Read at 450 nm within 5 minutes.		
Reagent Preparation:	Kindly use graduated containers to prepare the reagent.		
	2. Bring all reagents to room temperature (18-25 °C) before use for 30 min.		
	3. Prepare fresh standard for each assay. Use within 4 hours and discard after use.		
	4. Making serial dilution in the wells directly is not permitted.		
	5. To minimize imprecision caused by pipetting, use small volumes and ensure that pipettors		
	are calibrated. It is recommended to suck more than 10 $\mu L$ for once pipetting.		

6. Distilled water is recommended to be used to make the preparation for reagents. Contaminated water or container for reagent preparation will influence the detection result. y Antibody (1x) - Centrifuge the vial before opening. Antibody requires a 100-fold dilution. A suggested 100-fold dilution is 10 µL of Antibody + 990 µL of Antibody Diluent. y HRPconjugate (1x) - Centrifuge the vial before opening. HRP- conjugate requires a 100-fold dilution. A suggested 100-fold dilution is 10 µL of HRP- conjugate + 990 µL of HRP- conjugate Diluent. y Sample Extraction Buffer(1x)- If crystals have formed in the concentrate, warm up to room temperature and mix gently until the crystals have completely dissolved. Dilute 20 mL of Sample Extraction Buffer Concentrate (25 x) into deionized or distilled water to prepare 500 mL of Sample Extraction Buffer(1x). y Wash Buffer(1x)- If crystals have formed in the concentrate, warm up to room temperature and mix gently until the crystals have completely dissolved. Dilute 20 mL of Wash Buffer Concentrate (25 x) into deionized or distilled water to prepare 500 mL of Wash Buffer (1 x). 7 y Standard Centrifuge the standard vial at 6000-10000rpm for 30s. Allow the stock standard solution of 800 µg/mL to sit for a minimum of 30 minutes to warm to room temperature prior to making dilutions. Pipette 150 µL of Sample Diluent into each tube (S0-S4). Use the stock solution(S5) to produce a 4-fold dilution series (below). Mix each tube thoroughly before the next transfer. The undiluted Standard serves as the high standard (800 µg/mL). Sample Diluent serves as the zero standard (0 µg/mL).

#### Note:

#### Sample Preparation:

- It is recommended to use fresh samples without long storage, otherwise protein degradation and denaturation may occur in these samples, leading to false results. Samples should therefore be stored for a short period at 2 8 °C or aliquoted at -20 °C (≤1 month) or -80 °C (≤ 3 months). Repeated freeze-thaw cycles should be avoided. Prior to assay, the frozen samples should be slowly thawed and centrifuged to remove precipitates.
- If the sample type is not specified in the instructions, a preliminary test is necessary to determine compatibility with the kit.
- If a lysis buffer is used to prepare tissue homogenates or cell culture supernatant, there is a
  possibility of causing a deviation due to the introduced chemical substance. The
  recommended dilution factor is for reference only.
- Please estimate the concentration of the samples before performing the test. If the values
  are not in the range of the standard curve, the optimal sample dilution for the particular
  experiment has to be determined.

#### Note:

Samples require a 5-fold dilution into Sample Diluent. The recommended dilution factor is for reference only. The optimal dilution factor should be determined by users according to their particular experiments.

#### Assay Precision:

Intra-assay Precision (Precision within an assay): CV%<10% Three samples of known concentration were tested twenty times on one plate to assess.

Inter-assay Precision (Precision between assays): CV%<20% Three samples of known

## **Application Details**

	concentration were tested in twenty assays to assess.	
Restrictions:	For Research Use only	

## Handling


Storage:	4 °C,-20 °C	
Storage.	4 0, 20 0	

Storage Comment:

Unopened Store at 2 - 8°C. Do not use the kit beyond the expiration date. kit May be stored for up to 1 month at 2 - 8°C. Coated assay Try to keep it in a sealed aluminum foil bag, plate and avoid the damp. HRP-conjugate May be stored for up to 1 month at 2 - 8°C. If don't make recent use, better keep it store Antibody at -20°C. Standard Antibody Diluent Opened kit HRP-conjugate Diluent Sample Diluent May be stored for up to 1 month at 2 - 8°C. Sample Extraction Buffer Wash Buffer TMB Substrate Stop Solution \*Provided this is within the expiration date of the kit.

Expiry Date: 6 months

## **Images**



### **ELISA**

Image 1. Typical Standard Curve