

Datasheet for ABIN7525466

DBN1 IQ-ELISA Kit



_					
	W	0	rv	10	W

samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added to				
Reactivity: Human Method Type: Sandwich ELISA Application: ELISA Product Details Purpose: Human Immunoquantitative (PCR-Based) Drebrin 1 ELISA Kit for cell culture supernatants, plasma, and serum samples. Sample Type: Cell Culture Supernatant, Plasma, Serum Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added to the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Quantity:	5 x 96 tests		
Method Type: Sandwich ELISA Application: ELISA Product Details Purpose: Human Immunoquantitative (PCR-Based) Drebrin 1 ELISA Kit for cell culture supernatants, plasma, and serum samples. Sample Type: Cell Culture Supernatant, Plasma, Serum Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Target:	DBN1		
Application: ELISA Product Details Purpose: Human Immunoquantitative (PCR-Based) Drebrin 1 ELISA Kit for cell culture supernatants, plasma, and serum samples. Sample Type: Cell Culture Supernatant, Plasma, Serum Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Reactivity:	Human		
Purpose: Human Immunoquantitative (PCR-Based) Drebrin 1 ELISA Kit for cell culture supernatants, plasma, and serum samples. Sample Type: Cell Culture Supernatant, Plasma, Serum Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Method Type:	Sandwich ELISA		
Purpose: Human Immunoquantitative (PCR-Based) Drebrin 1 ELISA Kit for cell culture supernatants, plasma, and serum samples. Sample Type: Cell Culture Supernatant, Plasma, Serum Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added to the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Application:	ELISA		
plasma, and serum samples. Cell Culture Supernatant, Plasma, Serum Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Product Details			
Sample Type: Cell Culture Supernatant, Plasma, Serum Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added to the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Purpose:	Human Immunoquantitative (PCR-Based) Drebrin 1 ELISA Kit for cell culture supernatants,		
Analytical Method: Semi-Quantitative Detection Method: qPCR Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added to the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.		plasma, and serum samples.		
Detection Method: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added to the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Sample Type:	Cell Culture Supernatant, Plasma, Serum		
Characteristics: IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Analytical Method:	Semi-Quantitative		
samples are pipetted into the wells, the target protein in the standards and samples binds to the immobilized antibody. The wells are washed and the detection affinity reagent is added to the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Detection Method:	qPCR		
the immobilized antibody. The wells are washed and the detection affinity reagent is added to the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.	Characteristics:	IQ-ELISAs employ specific capture antibodies coated on a 96-well PCR plate. Standards and		
the wells where it binds to any captured antigen. The wells are washed, and primers and PCI master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.		samples are pipetted into the wells, the target protein in the standards and samples binds to		
master mix are added to each well. The plate is placed into a real time PCR instrument for cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.		the immobilized antibody. The wells are washed and the detection affinity reagent is added to		
cycling and measurement of DNA amplification. The cycle number where amplification is detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.		the wells where it binds to any captured antigen. The wells are washed, and primers and PCR		
detected is proportional to the amount of affinity detection reagent that bound to captured antigen in each well.		master mix are added to each well. The plate is placed into a real time PCR instrument for		
antigen in each well.		cycling and measurement of DNA amplification. The cycle number where amplification is		
		detected is proportional to the amount of affinity detection reagent that bound to captured		
Components: • Drebrin 1 Microplate (Item A): 96 well PCR plate coated with anti-Human Drebrin 1		antigen in each well.		
	Components:	Drebrin 1 Microplate (Item A): 96 well PCR plate coated with anti-Human Drebrin 1		

- Wash Buffer I Concentrate (20x) (Item B): 25 ml of 20x concentrated solution
- · Standards (Item C): 2 vials of recombinant Human Drebrin 1
- Assay Diluent A (Item D): 30 ml diluent buffer, 0.09% sodium azide as preservative. For Standard/Sample (serum/plasma) diluent
- Assay Diluent B (Item E): 15 ml of 5x concentrated buffer. For Standard/Sample (cell culture medium/urine) diluent
- Detection Affinity Reagent for Drebrin 1 (Item F): 2 vials of a 4x concentrated solution of anti-Human Drebrin 1 affinity reagent
- IQELISA Detection Reagent (Item G): 1.4ml of a 10x concentrated stock
- Primer Solution (Item I): 1.7 ml vial
- PCR Master Mix (Item J): 1.2 ml vial
- PCR Preparation buffer (Item K): 1ml vial of 10x concentrated buffer
- · Final Wash Buffer (Item L): 10ml vial of 10x concentrated buffer

Material not included:

- · Real-time PCR instrument, Bio-Rad recommended
- Precision pipettes to deliver 2 µL to 1 mL volumes
- · Adjustable 1-25 mL pipettes for reagent preparation
- · 100 mL and 1 liter graduated cylinders
- · Absorbent paper
- · Distilled or deionized water
- · Log-log graph paper or computer and software for data analysis
- · Tubes to prepare standard or sample dilutions
- Heating block or water bath capable of 80°C

Target Details

Target:	DBN1	
Alternative Name:	Drebrin 1 (DBN1 Products)	
Gene ID:	1627	
UniProt:	Q16643	
Pathways:	Maintenance of Protein Location	

Application Details	
Application Notes:	The Immuno-Quantitative ELISA (IQELISA) kits are an innovative assay platform that combines
	the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. This results in
	an assay that is simultaneously familiar and cutting edge and enables the use of only 1/10th
	the sample volume while also providing 10x more sensitivity than a traditional ELISA.
Comment:	The Immuno-Quantitative ELISA (IQELISA™) kits are an innovative assay platform that

Application Details

	combines the specificity and ease of use of an ELISA with the sensitivity of real-time PCR. Also
	called immuno-PCR, this detection platform results in an assay that is simultaneously familiar
	and cutting edge. Compared to traditional ELISA, IQELISA™ enables the use of only 1/10th the
	sample volume while also providing 10x more sensitivity.
Sample Volume:	25 μL
Plate:	Pre-coated
Protocol:	1. Prepare all reagents, samples and standards as instructed
	2. Add 25 μL standard or sample to each well. Incubate for 2.5 hours at room temperature or overnight at 4 $^{\circ}C$
	3. Add 25 µL detection affinity reagent to each well. Incubate 1 hour at room temperature
	4. Add 25µL of IQELISA Detection Reagent to each well. Incubate 1 hour
	5. Add 15µL Primer solution 10µL of PCR master mix to each well
	6. Run real-time PCR
Restrictions:	For Research Use only
Handling	
Storage:	4 °C,-20 °C,-80 °C
Storage Comment:	May be stored for up to 6 months at 2° to 8°C from the date of shipment. Standard
	(recombinant protein) should be stored at -20°C or -80°C (recommended at -80°C) after
	reconstitution. Opened PCR plate or reagents may be stored for up to 1 month at 2° to 8°C.
	Note: the kit can be used within one year if the whole kit is stored at -20°C. Avoid repeated
	freeze-thaw cycles.
Expiry Date:	6 months