antibodies -online.com





TNFRSF11A ELISA Kit





Overview

Quantity:	96 tests
Target:	TNFRSF11A
Binding Specificity:	AA 30-213
Reactivity:	Mouse
Method Type:	Sandwich ELISA
Detection Range:	62.5-4000 pg/mL
Minimum Detection Limit:	62.5 pg/mL
Application:	ELISA

Product Details

Purpose:	Sandwich High Sensitivity ELISA kit for Quantitative Detection of Mouse RANK
Brand:	PicoKine™
Sample Type:	Cell Culture Supernatant, Serum
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Immunogen:	Expression system for standard: NSO
	Immunogen sequence: Q30-P213
Specificity:	Expression system for standard: NSO
	Immunogen sequence: Q30-P213
Cross-Reactivity (Details):	There is no detectable cross-reactivity with other relevant proteins.

Product Details

Sensitivity:	<2pg/mL
Material not included:	Microplate reader in standard size. Automated plate washer. Adjustable pipettes and pipette
	tips. Multichannel pipettes are recommended in the condition of large amount of samples in the
	detection. Clean tubes and Eppendorf tubes. Washing buffer (neutral PBS or TBS). Preparation
	of 0.01M TBS: Add 1.2g Tris, 8.5g Nacl
Target Details	
Target:	TNFRSF11A
Alternative Name:	RANK (TNFRSF11A Products)
Background:	Protein Function: Receptor for TNFSF11/RANKL/TRANCE/OPGL, essential for RANKL-mediated
	osteoclastogenesis. Involved in the regulation of interactions between T-cells and dendritic
	cells
	Background: Receptor Activator of Nuclear Factor kappa B(RANK), also known as TRANCE
	Receptor, is a type I membrane protein that is expressed on the surface of osteoclasts and is
	involved in their activation upon ligand binding. RANK is a recently described TNF receptor
	family member, and its ligand, RANKL, promote survival of dendritic cells and differentiation of
	osteoclasts. RANK contains 383 amino acids in its intracellular domain(residues 234-616),
	which contain three putative TRAF-binding domains(termed I, II, and III). RANK interacts with
	various TRAFs through distinct motifs and activates NF-kappaB via a novel TRAF6 interaction
	motif, which then activates NIK, thus leading to NF-kappaB activation, whereas RANK most
	likely activates JNK through a TRAF2-interacting region in RANK.
	Synonyms: Tumor necrosis factor receptor superfamily member 11A,Osteoclast differentiation
	factor receptor,ODFR,Receptor activator of NF-KB,CD265,Tnfrsf11a,Rank,
	Full Gene Name: Tumor necrosis factor receptor superfamily member 11A
	Cellular Localisation: Cell membrane, Single-pass type I membrane protein.
Gene ID:	21934
JniProt:	035305
Pathways:	NF-kappaB Signaling
Application Details	
Application Notes:	Before using Kit, spin tubes and bring down all components to bottom of tube. Duplicate well
	assay was recommended for both standard and sample testing.

Application Details

Comment:	Tissue Specificity: Ubiquitous expression with high levels in trabecular bone, thymus, small
	intestine, lung, brain and kidney. Weakly expressed in spleen and bone marrow.
Plate:	Pre-coated
Protocol:	mouse RANK ELISA Kit was based on standard sandwich enzyme-linked immune-sorbent
	assay technology. A monoclonal antibody from rat specific for RANK has been precoated onto
	96-well plates. Standards(NSO, Q30-P213) and test samples are added to the wells, a
	biotinylated detection polyclonal antibody from goat specific for RANK is added subsequently
	and then followed by washing with PBS or TBS buffer. Avidin-Biotin-Peroxidase Complex was
	added and unbound conjugates were washed away with PBS or TBS buffer. HRP substrate
	TMB was used to visualize HRP enzymatic reaction. TMB was catalyzed by HRP to produce a
	blue color product that changed into yellow after adding acidic stop solution. The density of
	yellow is proportional to the mouse RANK amount of sample captured in plate.
Assay Procedure:	Aliquot 0.1 mL per well of the 4000pg/mL, 2000pg/mL, 1000pg/mL, 500pg/mL, 250pg/mL,
,	125pg/mL, 62.5pg/mL mouse RANK standard solutions into the precoated 96-well plate. Add
	0.1 mL of the sample diluent buffer into the control well (Zero well). Add 0.1 mL of each
	properly diluted sample of mouse cell culture supernates or serum to each empty well. See
	"Sample Dilution Guideline" above for details. It is recommended that each mouse RANK
	standard solution and each sample be measured in duplicate.
Assay Precision:	 Sample 1: n=16, Mean(pg/ml): 735, Standard deviation: 38.22, CV(%): 5.2
	• Sample 2: n=16, Mean(pg/ml): 1428, Standard deviation: 87.11, CV(%): 6.1
	 Sample 3: n=16, Mean(pg/ml): 2873, Standard deviation: 140.8, CV(%): 4.9, Sample 1: n=24, Mean(pg/ml): 873, Standard deviation: 57.62, CV(%): 6.6
	Sample 2: n=24, Mean(pg/ml): 1740, Standard deviation: 125.28, CV(%): 7.2
	• Sample 3: n=24, Mean(pg/ml): 3226, Standard deviation: 170.98, CV(%): 5.3
Restrictions:	For Research Use only
Handling	
Handling Advice:	Avoid multiple freeze-thaw cycles.
Storage:	-20 °C,4 °C
Storage Comment:	Store at 4°C for 6 months, at -20°C for 12 months. Avoid multiple freeze-thaw cycles
Expiry Date:	12 months

ELISA

Image 1. Mouse RANK PicoKine ELISA Kit standard curve