.-online.com antibodies

Datasheet for ABIN924857 IL23 ELISA Kit

Image



Overview

Quantity:	96 tests
Target:	IL23
Reactivity:	Human
Method Type:	Sandwich ELISA
Detection Range:	63-8000 pg/mL
Minimum Detection Limit:	63 pg/mL
Application:	ELISA

Product Details

Purpose:

The OmniKine? Human IL-23 ELISA Kit contains the components necessary for quantitative determination of natural or recombinant Human IL-23 concentrations within any experimental sample including cell lysates, serum and plasma. This particular immunoassay utilizes the quantitative technique of a "Sandwich" Enzyme-Linked Immunosorbent Assay (ELISA) where the target protein (antigen) is bound in a "sandwich" format by the primary capture antibodies coated to each well-bottom and the secondary detection antibodies added subsequently by the investigator. The capture antibodies coated to the bottom of each well are specific for a particular epitope on Human IL-23 while the user-added detection antibodies bind to epitopes on the captured target protein. Amid each step of the procedure, a series of wash steps must be performed to ensure the elimination of non- specific binding between proteins to other proteins or to the solid phase. After incubation and "sandwiching" of the target antigen, a peroxidase enzyme is conjugated to the constant heavy chain of the secondary antibody (either covalently or via Avidin/Streptavidin-Biotin interactions), allowing for a colorimetric reaction to ensue upon substrate addition. When the substrate TMB (3, 3', 5, 5'-Tetramethylbenzidine) is

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/7 | Product datasheet for ABIN924857 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Product Details

	added, the reaction catalyzed by peroxidase yields a blue color that is representative of the antigen concentration. Upon sufficient color development, the reaction can be terminated through addition of Stop Solution (2 N Sulfuric Acid) where the color of the solution will turn yellow. The absorbance of each well can then be read by a spectrophotometer, allowing for generation of a standard curve and subsequent determination of protein concentration.
Brand:	OmniKine™
Sample Type:	Cell Lysate, Serum, Plasma
Analytical Method:	Quantitative
Detection Method:	Colorimetric
Specificity:	The Human IL-23 ELISA Kit allows for the detection and quantification of endogenous levels of natural and/or recombinant Human IL-23 proteins.
Cross-Reactivity (Details):	The Human IL-23 ELISA is capable of recognizing both recombinant and naturally produced Human IL-23 proteins. The antigens listed below were tested at 50 ng/mL and did not exhibit significant cross reactivity or interference. Human: IL-6, IL-11, IL-12, IL-12 p40, IL-12 Rβ2, IL-23 p19 Murine: IL-23, IL-23 p19
Characteristics:	The Human IL-23 ELISA Kit allows for the detection and quantification of endogenous levels of natural and/or recombinant Human IL-23 proteins within the range of 63-8000 pg/mL.
Components:	 Microstrips Coated w / Capture Antibody: 12 x 8-Well Microstrips Protein Standard: Lyophilized (100 ng), Red container Biotinylated Detection Antibody: Lyophilized, Yellow container 400x Streptavidin-HRP: 30 µL, Blue container Wash Buffer (10x): 50 mL, Clear containter Assay Diluent: 50 mL, Clear container Ready-to-Use Substrate: 12 mL, Brown container Stop Solution: 12 mL, Clear container Adhesive Plate Sealers: 4 Sheets Technical Manual 1 Manual
Material not included:	The following materials and/or equipment are NOT provided in this kit but are necessary to successfully conduct the experiment: Microplate reader able to measure absorbance at 450 nm (with correction wavelength set to 540 nm or 570 nm) Micropipettes with capability of measuring volumes ranging from 1 µl to 1 mL Deionized or sterile water Squirt bottle, manifold dispenser, multichannel pipette reservoir or automated microplate

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/7 | Product datasheet for ABIN924857 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

washer

Graph paper or computer software capable of generating or displaying logarithmic functions Absorbent paper or vacuum aspirator Test tubes or microfuge tubes capable of storing ≥1 mL Bench top centrifuge (optional) Bench top vortex (optional) Orbital shaker (optional)

Target Details

Target:	IL23
Alternative Name:	IL-23 (IL23 Products)
Background:	This gene encodes a subunit of the heterodimeric cytokine interleukin 23 (IL23). IL23 is
	composed of this protein and the p40 subunit of interleukin 12 (IL12B). The receptor of IL23 is
	formed by the beta 1 subunit of IL12 (IL12RB1) and an IL23 specific subunit, IL23R. Both IL23
	and IL12 can activate the transcription activator STAT4, and stimulate the production of
	interferon-gamma (IFNG). In contrast to IL12, which acts mainly on naive CD4(+) T cells, IL23
	preferentially acts on memory CD4(+) T cells. IL-23 associates with IL12B to form the IL-23
	interleukin, a heterodimeric cytokine that functions in innate and adaptive immunity. IL-23 may
	constitute with IL-17, an acute response to infection in peripheral tissues. IL-23 binds to a
	heterodimeric receptor complex composed of IL12RB1 and IL23R, activates the Jak-Stat
	signaling cascade, stimulates memory rather than naive T-cells, and promotes production of
	proinflammatory cytokines. IL-23 induces autoimmune inflammation and thus may be
	responsible for autoimmune inflammatory diseases and may be important for tumorigenesis.
	IL-23 forms disulfide-linked heterodimer with IL12B. The heterodimer is known as interleukin IL
	23 and is secreted by activated dendritic and phagocytic cells and keratinocytes. Interleukin IL-
	23 is expressed by dermal Langerhans cells (at protein level), up-regulated by a wide array of
	pathogens and pathogen-products together with self-signals for danger or injury, and up-
	regulated in psoriatic dermal tissues, in dendritic cells of multiple sclerosis patients and in
	tumors. Source: Entrez Gene, Swiss-Prot
Gene ID:	51561, 3593
Liu: Durata	

UniProt:

Q9NPF7, P29460

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 3/7 | Product datasheet for ABIN924857 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Plate:	Pre-coated
Protocol:	This particular immunoassay utilizes the quantitative technique of a Sandwich Enzyme-Linked
	Immunosorbent Assay (ELISA) where the target protein (antigen) is bound in a sandwich
	format by the primary capture antibodies coated to each well-bottom and the secondary
	detection antibodies added subsequently by the investigator. The capture antibodies coated to
	the bottom of each well are specific for a particular epitope on the Human IL-23 cytokine while
	the user-added detection antibodies bind to epitopes on the captured target protein. Amid each
	step of the procedure, a series of wash steps must be performed to ensure the elimination of
	non-specific binding between proteins to other proteins or to the solid phase. After incubation
	and sandwiching of the target antigen, a peroxidase enzyme is conjugated to the constant
	heavy chain of the secondary antibody (either covalently or via Avidin/Streptavidin-Biotin
	interactions), allowing for a colorimetric reaction to ensue upon substrate addition. When the
	substrate TMB (3, 3', 5, 5'- Tetramethylbenzidine) is added, the reaction catalyzed by peroxidas
	yields a blue color that is representative of the antigen concentration. Upon sufficient color
	development, the reaction can be terminated through addition of Stop Solution (2 N Sulfuric
	Acid) where the color of the solution will turn yellow. The absorbance of each well can then be
	read by a spectrophotometer, allowing for generation of a standard curve and subsequent
	determination of protein concentration.
Sample Preparation:	If samples are to be used within 24 hours, aliquot and store at 4 °C. If samples are to be used
	over a long period of time, aliquot and store between -20 °C and -80 °C, depending on the
	duration of storage.
	Note: Samples containing a visible precipitate or pellet must be clarified prior to use in the
	assay.
	Caution: Avoid repeated freeze/thaw cycles to prevent loss of biological activity of proteins in
	experimental samples.
	Cell Lysate and Supernatants:
	Remove large cell components via centrifugation and perform the assay. Cell lysates and
	supernatants require a dilution using Assay Diluent. A serial dilution may be performed to
	determine a suitable dilution factor for the sample. For future use of the sample, follow the
	sample storage guidelines stated above.Serum:
	Allow samples to clot in a serum separator tube (SST) for 30 minutes. After sufficient
	clotting, centrifuge at 1000 x g for 15 minutes and remove serum from SST in preparation fo
	the assay. Serum samples require at least a 1:50 dilution using Assay Diluent. For future use
	of the sample, follow the storage guidelines above.
	Plasma: Use henorin, citrate or EDTA as an anticeographent to gather plasma from original biological
	Use heparin, citrate or EDTA as an anticoagulant to gather plasma from original biological


Application Details

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/7 | Product datasheet for ABIN924857 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

	sample. After collection of the plasma, centrifuge for 15 minutes at 1000 x g. This step must be performed within 30 minutes of plasma collection. Plasma samples require at least a 1:50 dilution using Assay Diluent. Afterwards, perform the assay or for future use of the sample, follow the storage guidelines stated above.
Assay Procedure:	Note: If possible, all incubation steps should be performed on an orbital shaker to equilibrate solutions when added to the microplate wells. Also, all provided solutions should be at ambient temperature prior to use. Note: Avoid adding solutions into wells at an angle, always keep pipette tip perpendicular to
	plate bottom.
	Reconstitution of Provided Materials:
	 Reconstitute the Biotin-Conjugated Detection Antibody in 67 μL of ddHIO for a concentration of 180 μg/ml. Reconstitute the Protein Standard in 100 μL of ddHIO for a concentration of 340 ng/ml. Dilute the 50 mL of 10x Wash Buffer in 450 mL of ddH2O for 500 mL of 1x Wash Buffer.
	Addition of Known Standard and Unknown Sample to Immunoassay:
	The OmniKine™ Human CD163 ELISA Kit allows for the detection and quantification of endogenous levels of natural and/or recombinant Human CD163 proteins
Calculation of Results:	Generation of Standard Curve and Interpretation of Data
	1. Average the duplicate or triplicate readings for each standard, control and sample and
	subtract the average zero standard optical density.
	2. Generate a standard curve by using Microsoft Excel or other computer software capable of
	establishing a 4- Parameter Logistic (4-PL) curve fit. If using Excel or an alternative graphing
	tool, plot the average optical density values in absorbance units (y-axis) against the known
	standard concentrations in pg/ml (x-axis). Note: Only use the values in which a noticeable
	gradient can be established. Afterwards, generate a best fit curve or trend-line through the
	plotted points via regression analysis.
Restrictions:	For Research Use only
Handling	
Precaution of Use:	Reagents provided in this kit may be harmful if ingested, inhaled or absorbed through the skin.
	Please carefully review the MSDS for each reagent before conducting the experiment.
	Stop Solution contains 2 N Sulfuric Acid (H2SO4) and is an extremely corrosive agent. Please
	wear proper eye, hand and face protection when handling this material. When the experiment is

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 5/7 | Product datasheet for ABIN924857 | 09/12/2023 | Copyright antibodies-online. All rights reserved.

Handling	
	finished, be sure to rinse the plate with copious amounts of running water to dilute the Stop Solution prior to disposing the plate.
Handling Advice:	This ELISA kit is intended for research purposes only, NOT diagnostic or clinical procedures of any kind. Materials included in this kit should NOT be used past the expiration date on the kit label. Reagents or substrates included in this kit should NOT be mixed or substituted with reagents or
	substrates from any other kits. Variations in pipetting technique, washing technique, operator laboratory technique, kit age, incubation time or temperature may cause differences in binding affinity of the materials
	provided. The assay is designed to eliminate interference and background by other cellular macromolecules or factors present within any biological samples. However, the possibility of
	background noise cannot be fully excluded until all factors have been tested using the assay kit. Reagents provided in this kit may be harmful if ingested, inhaled or absorbed through the skin. Please carefully review the MSDS for each reagent before conducting the experiment. Stop Solution contains 2 N Sulfuric Acid (H2SO4) and is an extremely corrosive agent. Please wear proper eye, hand and face protection when handling this material. When the experiment is finished, be sure to rinse the plate with copious amounts of running water to dilute the Stop
Storage:	Solution prior to disposing the plate. 4 °C
Storage Comment:	 Note: If used frequently, reagents may be stored at 4 °C. Unopened Kits: Store at 4 °C for 6 months. Microstrips Coated w/ Capture Antibody, 400x Streptavidin-HRP Wash Buffer (10x), Assay Diluent Ready-to-Use Substrate, Stop Solution: 6 Months at 4 °C Protein Standard, Biotinylated Detection Antibody: Lyophilized: 6 Months (if Reconstituted: 1 Month) at 4 °C



ELISA

Image 1. This is an example of what a typical standard curve will look like. You must make your own standard curve. Do not use this example as your own standard curve.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 7/7 | Product datasheet for ABIN924857 | 09/12/2023 | Copyright antibodies-online. All rights reserved.