antibodies -online.com





anti-CD20 antibody (FITC)

3 Images

4

Publications



Go to Product page

()	11/0	r\ /1	$\triangle 1 $
	$\lor \lor \vdash$	$I \vee I$	ew

Quantity:	100 tests
Target:	CD20 (MS4A1)
Reactivity:	Human
Host:	Mouse
Clonality:	Monoclonal
Conjugate:	This CD20 antibody is conjugated to FITC
Application:	Flow Cytometry (FACS)

Product Details

Immunogen:	Normal human lymphocytes from lymph node.
Clone:	LT20
Isotype:	lgG2a
Specificity:	The antibody LT20 reacts with an extracellular epitope of CD20 (Bp35), a 33-37 kDa non-glycosylated membrane receptor with four transmembrane domains, expressed on B lymphocytes (it is lost on plasma cells), follicular dendritic cells, and at low levels on peripheral blood T lymphocytes.
Cross-Reactivity (Details):	Human
Purification:	Purified antibody is conjugated with fluorescein isothiocyanate (FITC) under optimum conditions and unconjugated antibody and free fluorochrome are removed by size-exclusion chromatography.

Target Details

Target:	CD20 (MS4A1)	
Alternative Name:	CD20 (MS4A1 Products)	
Background:	MS4A1,CD20 is a cell surface 33-37 (depending on the degree of phosphorylation) kDa non-	
	glycosylated surface phosphoprotein expressed on mature and most malignant B cells, but not	
	stem cells or plasma cells (low number of the CD20 has been also detected on a subpopulation	
	of T lymphocytes and it can be expressed on follicular dendritic cells). Its expression on B cells	
	is synchronous with the expression of surface IgM. CD20 regulates transmembrane calcium	
	conductance (probably functioning as a component of store-operated calcium channel), cell	
	cycle progression and B-cell proliferation. It is associated with lipid rafts, but the intensity of this	
	association depends on extracellular triggering, employing CD20 conformational change and/or	
	BCR (B cell antigen receptor) aggregation. After the receptor ligation, BCR and CD20 colocalize	
	and then rapidly dissociate before BCR endocytosis, whereas CD20 remains at the cell surface.	
	CD20 serves as a useful target for antibody-mediated therapeutic depletion of B cells, as it is	
	expressed at high levels on most B-cell malignancies, but does not become internalized or shed	
	from the plasma membrane following mAb treatment.,B1, S7, MS4A, Bp35, CVID5, LEU-16	
Gene ID:	931	
UniProt:	P11836	
	P11836	
UniProt: Application Details	P11836	
	P11836 Flow cytometry: The reagent is designed for analysis of human blood cells using 20 μL reagent	
Application Details		
Application Details	Flow cytometry: The reagent is designed for analysis of human blood cells using 20 µL reagent	
Application Details	Flow cytometry: The reagent is designed for analysis of human blood cells using 20 µL reagent / 100 µL of whole blood or 10 ⁶ cells in a suspension. The content of a vial (2 ml) is sufficient for	
Application Details Application Notes:	Flow cytometry: The reagent is designed for analysis of human blood cells using 20 μ L reagent / 100 μ L of whole blood or 10 ⁶ cells in a suspension. The content of a vial (2 ml) is sufficient for 100 tests.	
Application Details Application Notes:	Flow cytometry: The reagent is designed for analysis of human blood cells using 20 µL reagent / 100 µL of whole blood or 10 ⁶ cells in a suspension. The content of a vial (2 ml) is sufficient for 100 tests. The purified antibody is conjugated with Fluorescein isothiocyanate (FITC) under optimum	
Application Details Application Notes:	Flow cytometry: The reagent is designed for analysis of human blood cells using 20 µL reagent / 100 µL of whole blood or 10 ⁶ cells in a suspension. The content of a vial (2 ml) is sufficient for 100 tests. The purified antibody is conjugated with Fluorescein isothiocyanate (FITC) under optimum conditions. The reagent is free of unconjugated FITC and adjusted for direct use. No	
Application Details Application Notes: Comment:	Flow cytometry: The reagent is designed for analysis of human blood cells using 20 µL reagent / 100 µL of whole blood or 10 ⁶ cells in a suspension. The content of a vial (2 ml) is sufficient for 100 tests. The purified antibody is conjugated with Fluorescein isothiocyanate (FITC) under optimum conditions. The reagent is free of unconjugated FITC and adjusted for direct use. No reconstitution is necessary.	
Application Details Application Notes: Comment: Restrictions: Handling	Flow cytometry: The reagent is designed for analysis of human blood cells using 20 µL reagent / 100 µL of whole blood or 10 ⁶ cells in a suspension. The content of a vial (2 ml) is sufficient for 100 tests. The purified antibody is conjugated with Fluorescein isothiocyanate (FITC) under optimum conditions. The reagent is free of unconjugated FITC and adjusted for direct use. No reconstitution is necessary. For Research Use only	
Application Details Application Notes: Comment: Restrictions: Handling Reconstitution:	Flow cytometry: The reagent is designed for analysis of human blood cells using 20 µL reagent / 100 µL of whole blood or 10 ⁶ cells in a suspension. The content of a vial (2 ml) is sufficient for 100 tests. The purified antibody is conjugated with Fluorescein isothiocyanate (FITC) under optimum conditions. The reagent is free of unconjugated FITC and adjusted for direct use. No reconstitution is necessary. For Research Use only	
Application Details Application Notes: Comment: Restrictions: Handling	Flow cytometry: The reagent is designed for analysis of human blood cells using 20 µL reagent / 100 µL of whole blood or 10 ⁶ cells in a suspension. The content of a vial (2 ml) is sufficient for 100 tests. The purified antibody is conjugated with Fluorescein isothiocyanate (FITC) under optimum conditions. The reagent is free of unconjugated FITC and adjusted for direct use. No reconstitution is necessary. For Research Use only	
Application Details Application Notes: Comment: Restrictions: Handling Reconstitution:	Flow cytometry: The reagent is designed for analysis of human blood cells using 20 µL reagent / 100 µL of whole blood or 10 ⁶ cells in a suspension. The content of a vial (2 ml) is sufficient for 100 tests. The purified antibody is conjugated with Fluorescein isothiocyanate (FITC) under optimum conditions. The reagent is free of unconjugated FITC and adjusted for direct use. No reconstitution is necessary. For Research Use only	

Handling


	should be handled by trained staff only.	
Handling Advice:	Do not freeze.	
	Avoid prolonged exposure to light.	
Storage:	4 °C	
Storage Comment:	Store at 2-8°C. Protect from prolonged exposure to light. Do not freeze.	
Publications		
Product cited in:	Všianská, Říhová, Varmužová, Suská, Kryukov, Mikulášová, Kupská, Penka, Pour, Adam, Hájek: "	

Všianská, Říhová, Varmužová, Suská, Kryukov, Mikulášová, Kupská, Penka, Pour, Adam, Hájek: "Analysis of B-cell subpopulations in monoclonal gammopathies." in: **Clinical lymphoma, myeloma & leukemia**, Vol. 15, Issue 4, pp. e61-71, (2015) (PubMed).

Filatov, Krotov, Zgoda, Volkov: "Fluorescent immunoprecipitation analysis of cell surface proteins: a methodology compatible with mass-spectrometry." in: **Journal of immunological methods**, Vol. 319, Issue 1-2, pp. 21-33, (2007) (PubMed).


Teeling, Mackus, Wiegman, van den Brakel, Beers, French, van Meerten, Ebeling, Vink, Slootstra, Parren, Glennie, van de Winkel: "The biological activity of human CD20 monoclonal antibodies is linked to unique epitopes on CD20." in: **Journal of immunology (Baltimore, Md.: 1950)**, Vol. 177, Issue 1, pp. 362-71, (2006) (PubMed).

Chan, Hughes, French, Tutt, Walshe, Teeling, Glennie, Cragg: "CD20-induced lymphoma cell death is independent of both caspases and its redistribution into triton X-100 insoluble membrane rafts." in: **Cancer research**, Vol. 63, Issue 17, pp. 5480-9, (2003) (PubMed).



Flow Cytometry

Image 1. Flow cytometry surface staining pattern of human peripheral whole blood stained using anti-human CD20 (LT20) FITC antibody (20 μ L reagent / 100 μ L of peripheral whole blood).



Flow Cytometry

Image 2. Surface staining of human peripheral blood cells with anti-human CD20 (LT20) FITC.

Relative Cell Count

Flow Cytometry

Image 3. Separation of human CD20 positive lymphocytes (red-filled) from neutrophil granulocytes (black-dashed) in flow cytometry analysis (surface staining) of human peripheral whole blood stained using anti-human CD20 (LT20) FITC antibody (20 μ L reagent / 100 μ L of peripheral whole blood).