.-online.com antibodies

Datasheet for ABIN956151 TNNC1 ELISA Kit



Overview

Quantity:	96 tests
Target:	TNNC1
Reactivity:	Pig
Method Type:	Sandwich ELISA
Application:	ELISA

Product Details

Serum
Quantitative
Colorimetric
The Pig Cardiac Troponin-I ELISA uses two different affinity purified cTnI specific antibodies.
One is used for solid phase immobilization (on the microtiter wells). The second is conjugated
to horseradish peroxidase (HRP). The serum sample is allowed to react simultaneously with the
two antibodies, resulting in cTnI being sandwiched between the solid phase and HRP-
conjugated antibodies. After one hour incubation at room temperature on a plate shaker, the
wells are washed to remove unbound HRP-conjugated antibodies. A solution of TMB
(Tetramethylbenzidine), an HRP substrate, is then added and incubated for 20 minutes,
resulting in the development of a blue color. The color development is stopped with the addition
of 1N HCl changing the color to yellow. The concentration of cTnl is proportional to the
absorbance at 450 nm.
Anti-cTnl-coated microtiter wells, 96 wells
Pig cTnl Calibrator (lyophilized), reconstitute with 0.40 mL H2O

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 1/4 | Product datasheet for ABIN956151 | 01/16/2024 | Copyright antibodies-online. All rights reserved.

	cTnl Diluent, 12 mL
	cTnI HRP Conjugate, 11 mL
	Wash Solution (20X), 50 mL
	TMB Reagent, 11 mL
	Stop Solution (1N HCl), 11 mL.
Material not included:	Pipettes: P-10, P-200 & P-1000 or equivalent
	Disposable pipette tips
	Distilled or de-ionized water
	Vortex mixer
	Absorbent paper
	Graph paper or appropriate PC graphing software
	Polypropylene microcentrifuge tubes (1.5 mL)
	Microtiter plate reader capable of reading 0 to 4 OD at 450 nm.

Target Details

Target:	TNNC1
Alternative Name:	Cardiac Troponin-1 (TNNC1 Products)
Background:	Troponin is the inhibitory or contractile regulating protein complex of striated muscle. It is
	located periodically along the thin filament of the muscle and consists of three distinct proteins:
	troponin I, troponin C, and troponin T. The troponin I subunit exists in three isoforms, two in
	fast-twitch and slow-twitch skeletal muscle fibers, and one in cardiac muscle. At the sequence
	level cardiac troponin-I (cTnI) is significantly different from the skeletal isoforms and antibodies
	can be prepared that specifically recognize cTnl. The unique isoform and tissue specificity of
	cTnl are the basis for its use as a marker of cardiac muscle damage.

Application Details

Plate:	Pre-coated
Sample Preparation:	Serum should be prepared as quickly as possible after blood collection and stored at 4°C. All samples should be similarly processed (i.e., storage times and temperatures should be the same). If serum samples cannot be assayed immediately they should be frozen at -70°C and thawed only once prior to use.
Assay Procedure:	 Secure the desired number of coated wells in the holder. Dispense 100 μL of cTnI HRP Conjugate into each well.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 2/4 | Product datasheet for ABIN956151 | 01/16/2024 | Copyright antibodies-online. All rights reserved.

	3. Dispense 100 μL of calibrators and samples into the appropriate wells.
	4. Incubate at room temperature (18-25°C) on a plate shaker (150 rpm) for one hour.
	5. Remove the incubation mixture using a plate washer or by flicking plate contents into a bio-
	waste container.
	6. Wash and empty the microtiter wells 5 times with 1X wash solution. This may be performed
	using either a plate washer (400 $\mu\text{L/well})$ or with a squirt bottle. The entire wash procedure
	should be performed as quickly as possible. (Use of a plate washer gives optimal results).
	7. Strike the wells sharply onto adsorbent paper or paper towels to remove all residual droplets.
	8. Dispense 100 μL of TMB Reagent solution into each well. Gently mix for 5 seconds.
	9. Incubate on a plate shaker (150 rpm) at room temperature for 20 minutes.
	10. Stop the reaction by adding 100 μL of Stop Solution to each well.
	11. Gently mix until all the blue color changes to yellow.
	12. Read absorbance at 450 nm with a microtiter plate reader within 15 minutes. Please note:
	Due to plate reader differences, the high calibrator absorbance values may be out of range
	occasionally. If this occurs, absorbance values may be determined at 405 nm instead.
	13. If absorbance values exceed the high calibrator, the samples should be appropriately diluted
	with cTnI diluent and re-determined. Samples with absorbance values below those of the
	lowest calibrator should be assigned a zero troponin-l value.
Calculation of Results:	1. Calculate the mean absorbance values (A450) for each set of reference calibrators and
	samples.
	samples. 2. Construct a calibration curve by plotting the mean absorbance obtained from each reference
	2. Construct a calibration curve by plotting the mean absorbance obtained from each reference
	2. Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on graph paper, with absorbance values on the
	2. Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on graph paper, with absorbance values on the vertical or Y-axis and concentration on the horizontal or X-axis.
	 Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on graph paper, with absorbance values on the vertical or Y-axis and concentration on the horizontal or X-axis. Using the mean absorbance value for each sample, determine the corresponding
	 Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on graph paper, with absorbance values on the vertical or Y-axis and concentration on the horizontal or X-axis. Using the mean absorbance value for each sample, determine the corresponding concentration of cTnl in ng/mL from the calibration curve.
	 Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on graph paper, with absorbance values on the vertical or Y-axis and concentration on the horizontal or X-axis. Using the mean absorbance value for each sample, determine the corresponding concentration of cTnl in ng/mL from the calibration curve. If available, graphing software may be used to analyze the data. Depending on the range of
	 Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on graph paper, with absorbance values on the vertical or Y-axis and concentration on the horizontal or X-axis. Using the mean absorbance value for each sample, determine the corresponding concentration of cTnl in ng/mL from the calibration curve. If available, graphing software may be used to analyze the data. Depending on the range of the calibration curve used, we find that good fits of the data may be obtained with linear
Restrictions:	 Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on graph paper, with absorbance values on the vertical or Y-axis and concentration on the horizontal or X-axis. Using the mean absorbance value for each sample, determine the corresponding concentration of cTnl in ng/mL from the calibration curve. If available, graphing software may be used to analyze the data. Depending on the range of the calibration curve used, we find that good fits of the data may be obtained with linear regression analysis or using a two-site binding model. Alternatively, calibration curves may be
	 Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on graph paper, with absorbance values on the vertical or Y-axis and concentration on the horizontal or X-axis. Using the mean absorbance value for each sample, determine the corresponding concentration of cTnl in ng/mL from the calibration curve. If available, graphing software may be used to analyze the data. Depending on the range of the calibration curve used, we find that good fits of the data may be obtained with linear regression analysis or using a two-site binding model. Alternatively, calibration curves may be generated using a point-to-point fit.
Handling	 2. Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on graph paper, with absorbance values on the vertical or Y-axis and concentration on the horizontal or X-axis. 3. Using the mean absorbance value for each sample, determine the corresponding concentration of cTnl in ng/mL from the calibration curve. 4. If available, graphing software may be used to analyze the data. Depending on the range of the calibration curve used, we find that good fits of the data may be obtained with linear regression analysis or using a two-site binding model. Alternatively, calibration curves may be generated using a point-to-point fit. For Research Use only
	 Construct a calibration curve by plotting the mean absorbance obtained from each reference calibrator against its concentration in ng/mL on graph paper, with absorbance values on the vertical or Y-axis and concentration on the horizontal or X-axis. Using the mean absorbance value for each sample, determine the corresponding concentration of cTnl in ng/mL from the calibration curve. If available, graphing software may be used to analyze the data. Depending on the range of the calibration curve used, we find that good fits of the data may be obtained with linear regression analysis or using a two-site binding model. Alternatively, calibration curves may be generated using a point-to-point fit.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 3/4 | Product datasheet for ABIN956151 | 01/16/2024 | Copyright antibodies-online. All rights reserved. to damp air. The expiration date of the kit is indicated on the box label.

Expiry Date:

The expiry date is stated on the label.

Order at www.antibodies-online.com | www.antikoerper-online.de | www.anticorps-enligne.fr | www.antibodies-online.cn International: +49 (0)241 95 163 153 | USA & Canada: +1 877 302 8632 | support@antibodies-online.com Page 4/4 | Product datasheet for ABIN956151 | 01/16/2024 | Copyright antibodies-online. All rights reserved.