



[Go to Product page](#)

Datasheet for ABIN996984
EBV VCA IgG ELISA Kit

Overview

Quantity:	96 tests
Target:	EBV VCA IgG
Reactivity:	Epstein-Barr Virus (EBV)
Method Type:	Competition ELISA
Application:	ELISA

Product Details

Sample Type:	Serum
Analytical Method:	Qualitative
Detection Method:	Colorimetric
Specificity:	98.1 %

Material not included:	<ol style="list-style-type: none">1. ELISA microwell reader capable of reading at a wavelength of 450 nm.2. Pipettes capable of accurately delivering 10 to 200 μL.3. Multichannel pipette capable of accurately delivering (50-200 μK)4. Reagent reservoirs for multichannel pipettes.5. Wash bottle or microwell washing system.6. Distilled or deionized water.7. One liter graduated cylinder.8. Serological pipettes.9. Disposable pipette tips.10. Paper towels.11. Laboratory timer to monitor incubation steps.
------------------------	---

Product Details

12. Disposal basin and disinfectant. (example: 10 % household bleach, 0.5 % sodium hypochlorite.)

Target Details

Target:	EBV VCA IgG
Abstract:	EBV VCA IgG Products
Target Type:	Antibody, Antibody

Application Details

Comment:	<p>Quality Control:</p> <ol style="list-style-type: none">1. Each time the assay is run the Calibrator must be run in triplicate. A reagent blank, Negative Control, and Positive Control must also be included in each assay.2. Calculate the mean of the three Calibrator wells. If any of the three values differ by more than 15 % from the mean, discard that value and calculate the mean using the remaining two wells.3. The mean OD value for the Calibrator and the OD values for the Positive and Negative Controls should fall within the following ranges: OD Range Negative Control <0.250 Calibrator > 0.300 Positive Control > 0.500 a. The OD of the Negative Control divided by the mean OD of the Calibrator should be < 0.9. b. The OD of the Positive Control divided by the mean OD of the Calibrator should be > 1.25. c. If the above conditions are not met the test should be considered invalid and should be repeated.4. The Positive Control and Negative Control are intended to monitor for substantial reagent failure and will not ensure precision at the assay cutoff.5. Additional controls may be tested according to guidelines or requirements of local, state, and/or federal regulations or accrediting organizations.6. Refer to NCCLS document C24: Statistical Quality Control for Quantitative Measurements for guidance on appropriate QC practices. <p>Limitations of procedure:</p> <ol style="list-style-type: none">1. Most (80 %) of IM individuals have peak anti-VCA IgG titers before they consult a physician (4). Therefore, testing paired acute and convalescent sera for significant changes in antibody levels is not useful in most patients with IM (4).2. The antibody titer of a single serum specimen cannot be used to determine recent infection. Test results for anti-VCA should be interpreted in conjunction with the clinical evaluation and results of antibody tests for other EBV antigens (i.e., EBNA, EA, and IgM-VCA).
Sample Volume:	10 µL

Application Details

Assay Time: 1 h

Plate: Pre-coated

Sample Preparation:

1. It is recommended that specimen collection be carried out in accordance with NCCLS document M29: Protection of Laboratory Workers from Infectious Disease.
2. No known test method can offer complete assurance that human blood samples will not transmit infection. Therefore, all blood derivatives should be considered potentially infectious.
3. Only freshly drawn and properly refrigerated sera obtained by approved aseptic venipuncture procedures should be used in this assay (27, 28). No anticoagulants or preservatives should be added. Avoid using hemolyzed, lipemic, or bacterially contaminated sera.
4. Store sample at room temperature for no longer than 8 hours. If testing is not performed within 8 hours, sera may be stored between 2° and 8 °C for no longer than 48 hours. If delay in testing is anticipated, store test sera at -20 °C or lower. Avoid multiple freeze/thaw cycles that may cause loss of antibody activity and give erroneous results.

Assay Procedure:

1. Break off number of wells needed (four for calibrators plus number of samples) and place in strip holder.
2. Add 100 µL of each calibrator to wells 1-4, then add 100 µL of the diluted test samples to the remaining wells. Note: Standards are supplied prediluted. Do not dilute further.
3. Incubate at room temperature (15 to 2. °C) for 10 minutes.
4. Shake out contents and wash 3 times with the diluted wash buffer.
5. Add 100 µL of Enzyme Conjugate to each well.
6. Incubate at room temperature for 5 minutes.
7. Shake out contents and wash 3 times with wash buffer, then rinse once with DI water. Slap wells against paper towels to remove excess moisture.

Calculation of Results:

A. Calculations:

1. Correction Factor A cutoff OD value for positive samples has been determined by the manufacturer and correlated to the Calibrator. The correction factor (CF) will allow you to determine the cutoff value for positive samples and to correct for slight day-to-day variations in test results. The correction factor is determined for each lot of kit components and is printed on the Component List located in the kit box.
2. CutoffOD Value To obtain the cutoff OD value, multiply the CF by the mean OD of the Calibrator determined above. (CFx mean OD of Calibrator = cutoffOD value)

3. Index Values or OD Ratios Calculate the Index Value or OD Ratio for each specimen by dividing its OD value by the cutoff OD from step

2.2 Relative

1. Comparative Study A comparative study was conducted to compare the DAI ELISA EBV-VCA IgG Test System to another commercial ELISA for the detection of IgG antibodies against EBV-VCA. A total of 199 serum specimens were obtained from two plasma donor centers and a reference laboratory. Below is a results summary. Table 1: DAI ELISA EBV-VCA IgG Test System vs. Commercial EBV-VCA IgG ELISA DAI EBV-VCA IgG ELISA Test System Positive Negative Equivocal * Total Commercial EBV- IgG ELISA Positive 104 8* 3 115 Negative 10** 46 4 60 Equivocal * 11 12 1 24 Totals 125 66 8 199 Relative Specificity = 46/56 = 8

2.1 % Percent Agreement: 150/168 = 89.3 % *Equivocal samples not included in calculations ** Discrepant results. See Table 2 below for the results of the retesting of the discrepant samples using the DAI IFA EBV-VCA IgG Test System: Table 2: Analysis of Discrepant Results Sample ID DAI ELISA Commerical DAI IFA EBV- EBV- VCA IgG VCA- IgG VCA IgG Test Test System ELISA System 26 0.251

1.16 - 29 0.261

1.00 - 39 0.358

1.00 - 39 0.368

1.13 - 41 0.391

1.30 - 53 0.582

1.05 - 63 0.767

1.10 + 34 0.797

1.30 + 81

1.243 0.65 + 83

1.270 0.70 + 89

1.399 0.73 + 93

1.449 0.39 - 101

1.614 0.57 + 111

1.859 0.60 + 133

2.216 0.48 + 137

2.254 0.75 + 139

2.320 0.77 + 152

2.730 0.20 + Summary a. Eight samples were negative by DAI ELISA and positive by the commercial VCA IgG ELISA. DAI IFA confirmed six of these eight samples to be negative. b. Ten

samples were positive by DAI ELISA and negative by the commercial VCA IgG ELISA. DAI IFA confirmed nine of these ten samples to be positive. c. Recalculation of the relative sensitivity, relative specificity, and percent agreement, based upon the resolution of the discrepant samples by IFA, provided the results are shown in Table 3: Table 3 Recalculation of Relative

All immunocompetent persons infected with EBV produce antibodies to VCA (6). Both IgM and IgG antibodies to VCA appear rapidly following infection and reach peak titers within three to four weeks (4). IgG antibodies to VCA decline slowly after peaking but persist indefinitely (15). Indications of a primary acute EBV infection are the presence of IgG antibodies to VCA, coupled with anti-EA, and/or IgM anti-VCA antibodies, and the absence of antibodies to EBNA (4, 11). The presence of IgG anti-VCA and anti-EBNA indicates the infection was not recent (4, 11). The incidence of EBV infection varies with age and socioeconomic status (6). In the underdeveloped countries, most persons acquire EBV in early childhood and the infection is usually unapparent (2, 4). In one study in the United States, about 50 % of college freshmen were seropositive for EBV (30). In a study conducted by DAI Scientific technicians using 135 normal samples from the Southeastern United States, 134/135 samples were positive by both the DAI ELISA EBV-VCA IgG Test System and the DAI IFA EBV-VCA IgG Test System. In another study consisting of 32 normal pediatric samples, 32/32 samples were negative using both the DAI ELISA EBV-VCA IgG Test System, and the DAI IFA EBV-VCA IgG Test System. The number of individuals with IgG antibody to EBV-VCA varies with age and socioeconomic status. DAI Scientific recommends that each laboratory establish their expected values based upon the population type typically tested. cc

Restrictions: For Research Use only

Handling

Storage: 4 °C

Expiry Date: 14 months